



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 STATE PUBLIC HEALTH LABORATORY  
**ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT**

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 111759	NAME OF AGENCY CLAY COUNTY SHERIFF'S OFFICE	DATE OF INSPECTION 12/02/2021
LOCATION OF INSTRUMENT (STREET AND CITY) 12 S. WATER STREET, LIBERTY, MO 64068		TIME OF INSPECTION 8:40 pm

**CHECKLIST:** Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)

TEMPERATURE OF ALCO SENSOR (10°C - 40°C)

PRINTER WORKING PROPERLY

TIME AND DATE DISPLAYING PROPERLY

**BREATH ALCOHOL ACCURACY STANDARDS**

SIMULATOR SOLUTION  COMPRESSED ETHANOL-GAS MIXTURE

STANDARD SUPPLIER INTOXIMETERS \_\_\_\_\_ LOT # AG102503 \_\_\_\_\_ EXP. DATE 01/25/2023 \_\_\_\_\_

SIMULATOR TEMPERATURE (34°C ± 0.2°C) \_\_\_\_\_ SIM. SN \_\_\_\_\_ SIM. NIST EXP DATE \_\_\_\_\_

CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
- 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
- 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 ➡ .098	TEST 2 ➡ .098	TEST 3 ➡ .097
---------------	---------------	---------------

RFI DETECTOR OPERATING

**INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:  
 (DO NOT INCLUDE SELF-ADMINISTERED TESTS)**

REFUSALS	0	(0-.04)	0	(.05-.09)	0	(.10-.14)	1	(.15-.19)	1	(OVER .19)	1
----------	---	---------	---	-----------	---	-----------	---	-----------	---	------------	---

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

CHANGE CLOCK TO CENTRAL STANDARD TIME

**INSPECTING OFFICER**

SIGNATURE 	PRINT NAME JAMES BUSH
TYPE II PERMIT NUMBER/EXPIRATION DATE 200048 - 01/10/2022	TELEPHONE NUMBER (816) 407-3702

**Return completed report to the:** Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office by mail, fax, or email.

AS IV Serial no: 111759  
Version no: 532B

TEST RECORD 80891

Temp Date Time %/ 219L

Air Blank  
12/22/21 20:48 .888  
Calibration Check  
24 12/22/21 20:44 .898

Subject Name  
Monthly Maintenance  
Subject ID

Test 1  
Operator Name, I.D.  
J. Bush #8199

Location  
12 S. Water St.  
Liberty, MO 64068

AS IV Serial no: 111759  
Version no: 532B

TEST RECORD 80892

Temp Date Time %/ 219L

Air Blank  
12/22/21 20:41 .888  
Calibration Check  
24 12/22/21 20:41 .898

Subject Name  
Monthly Maintenance  
Subject ID

Test 2  
Operator Name, I.D.  
J. Bush #8199

Location  
12 S. Water St.,  
Liberty, MO 64068

AS IV Serial no: 111759  
Version no: 532B

TEST RECORD 80893

Temp Date Time %/ 219L

Air Blank  
12/22/21 20:43 .888  
Calibration Check  
24 12/22/21 20:38 .897

Subject Name  
Monthly Maintenance  
Subject ID

Test 3  
Operator Name, I.D.  
J. Bush #8199

Location  
12 S. Water St.,  
Liberty, MO 64068

AS IV Serial no: 111759  
Version no: 532B

TEST RECORD 80894

Temp Date Time %/ 219L

VOID: 301  
12 12/22/21 20:44

Subject Name  
Monthly Maintenance

RFI Test  
Operator Name, I.D.  
J. Bush #8199

Location  
12 S. Water St.,  
Liberty, MO 64068



Airgas USA LLC (LAB)  
 3500 Bernard Street  
 St. Louis, Mo. 63103  
 Ph: (314) 533-3100  
 Fax: (314) 533-7328

## Certificate of Analysis

**Customer Name**  
 Exclusive Supplier  
 Intoximeters, Inc.  
 2081 Craig Road  
 St. Louis, Mo 63146

**Test Date:** 27-Jan-2021

**Lot # AG102503 Model 108cadd**

<u>Exp. Date</u>	<u>Cyl. Type</u>	<u>Component</u>	<u>Certified Concentration</u>
25-Jan-2023	108	Ethanol Nitrogen	0.100 ± 2% BrAC (272 ppm) Balance

Certification Traceable to N.I.S.T. RGM and to CRM Ethanol Standards:

<u>RGM Serial No.</u>	<u>Concentration</u>	<u>RGM Serial No.</u>	<u>Concentration</u>
EB0010581	392.1 ppm	EB0010603	393.0 ppm
EB0010570	259.8 ppm	EB0010559	258.2 ppm
EB0010285	208.0 ppm	EB0010595	208.3 ppm
EB0010561	103.6 ppm	EB0010562	104.2 ppm
EB0010681	52.12 ppm	EB0010579	52.81 ppm

  

<u>CRM Serial No.</u>	<u>Concentration</u>	<u>CRM Serial No.</u>	<u>Concentration</u>
CC727481	800.0 ppm	CC727493	390.0 ppm
CC727496	253.0 ppm	CC727498	150.0 ppm

**Analytical Method:** NDIR

Digitally signed by Quality Control  
 Date: 2021.01.29 13:36:13 -06:00  
 Reason: Dry gas standard certification of analysis  
 Location: Airgas USA LLC (Lab)

Approved for Release:

\_\_\_\_\_  
 Rod Marsala

**ISO 17025:2005 A2LA accredited. Certificate Number 3082.06**  
**ISO 17034:2016 A2LA accredited. Certificate Number 3082.07**



STATE OF MISSOURI  
 DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 BREATH ALCOHOL PROGRAM

2

**PERMIT  
 TYPE II**

**JAMES S. BUSH**

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

**ALCO-SENSOR IV WITH PRINTER, INTOX EC/IR II**

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 1/10/2020

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

NUMBER 200048

EXPIRES 1/10/2022

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO-580-0771 (6-10)

LAB-4 (R6-10)

STATE OF MISSOURI  
 DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 BREATH ALCOHOL PROGRAM

**INSTRUMENT OPERATOR CARD**

*The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.*

Operator BUSH, JAMES  
 Permit No 200048  
 Date Issued 1/10/2020    Date Expires 1/10/2022