



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 STATE PUBLIC HEALTH LABORATORY  
**ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT**

**RECEIVED**  
 By Tracy Crews at 9:51 am, Dec 29, 2021

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 111770	NAME OF AGENCY MSHP	DATE OF INSPECTION 12/28/2021
LOCATION OF INSTRUMENT (STREET AND CITY) 891 TECHNOLOGY DRIVE, WELDON SPRING (MOBILE UNIT)		TIME OF INSPECTION 11:50 am

**CHECKLIST:** Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)
- TEMPERATURE OF ALCO SENSOR (10°C - 40°C)
- PRINTER WORKING PROPERLY
- TIME AND DATE DISPLAYING PROPERLY

**BREATH ALCOHOL ACCURACY STANDARDS**

- SIMULATOR SOLUTION  COMPRESSED ETHANOL-GAS MIXTURE
- STANDARD SUPPLIER REPCO MARKETING CO. LOT # 20001 EXP. DATE 10/07/2022
- SIMULATOR TEMPERATURE (34°C ± 0.2°C) 34.00 SIM. SN MP2459 SIM. NIST EXP DATE 03/01/2022

- CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)  
 Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)
  - 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
  - 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
  - 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1  .101	TEST 2  .101	TEST 3  .101
--------------	--------------	--------------

- RFI DETECTOR OPERATING

**INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)**

REFUSALS	0	(0-.04)	0	(.05-.09)	0	(.10-.14)	0	(.15-.19)	0	(OVER .19)	0
----------	---	---------	---	-----------	---	-----------	---	-----------	---	------------	---

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

**INSPECTING OFFICER**

SIGNATURE 	PRINT NAME A. MICHELS
TYPE II PERMIT NUMBER/EXPIRATION DATE PERMIT NUMBER 210321, EXPIRATION 12/16/2023	TELEPHONE NUMBER (636) 300-2800

**Return completed report to the:** Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office by mail, fax, or email.

AS IV Serial no: 111770  
Version no: 532B

TEST RECORD 00200

Temp Date Time <sup>s/</sup> 210L

Air Blank:  
12/28/21 11:29 .000  
Calibration Check:  
20 12/28/21 11:29 .101

Subject Name

*Maint*

Subject I.D.

Operator Name, I.D.

*A. Michels #210321*

Location

*Troop C HQ*

*Mobile*

AS IV Serial no: 111770  
Version no: 532B

TEST RECORD 00201

Temp Date Time <sup>s/</sup> 210L

Air Blank:  
12/28/21 11:31 .000  
Calibration Check:  
21 12/28/21 11:31 .101

Subject Name

*Maint*

Subject I.D.

Operator Name, I.D.

*A. Michels #210321*

Location

*Troop C HQ*

AS IV Serial no: 111770  
Version no: 532B

TEST RECORD 00202

Temp Date Time <sup>s/</sup> 210L

Air Blank:  
12/28/21 11:32 .000  
Calibration Check:  
21 12/28/21 11:32 .101

Subject Name

*Maint*

Subject I.D.

Operator Name, I.D.

*A. Michels #210321*

Location

*Troop C HQ*

AS IV Serial no: 111770  
Version no: 532B

TEST RECORD 00203

Temp Date Time <sup>s/</sup> 210L

VOID: RFI  
12 12/28/21 11:35

Subject Name

*Maint*

Subject I.D.

Operator Name, I.D.

*A. Michels #210321*

Location

*Troop C HQ*



## CERTIFICATE OF ANALYSIS

**MANUFACTURER AND SUPPLIER: RepCo Marketing Co.**

**LOT NUMBER: 20001**

**EXPIRATION DATE: October 7, 2022 at 11:59 p.m.**

RepCo Marketing Co. certifies the following:

RepCo Marketing Co. prepared, tested and supplied Lot Number 20001 of Alcohol Certified Solution for simulators. This solution was manufactured and tested by RepCo Marketing Co., with confirmation by ISO 17025 and ISO 17034 accredited institution Alcohol Countermeasure Systems, using NIST standards. Random samples were analyzed by Alcohol Countermeasure Systems utilizing a gas chromatograph and found to contain .1227 gms/dl +/- .003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of .100 +/- 3% gms/210L breath when heated to 34 Degrees Celsius +/- 0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is October 8, 2020 The expiration date for this lot number is October 7, 2022 at 11:59 p.m.

This document is a true representation of the original Certificate of Analysis.



Alma Palmer, Operations Manager  
RepCo Marketing Co.



STATE OF MISSOURI  
DEPARTMENT OF HEALTH AND SENIOR SERVICES  
BREATH ALCOHOL PROGRAM



**PERMIT**  
**TYPE II**  
**ADAM J. MICHELS**

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

**ALCO-SENSOR IV WITH PRINTER**

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE **12/16/2021**

NUMBER **210321**

EXPIRES **12/16/2023**

*Laura Q. Day*

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

*Donald A. Ramsey*

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES



**STATE OF MISSOURI**  
DEPARTMENT OF HEALTH AND SENIOR SERVICES  
BREATH ALCOHOL PROGRAM

**INSTRUMENT OPERATOR CARD**

*The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.*

**Operator** MICHELS, ADAM  
**Permit No** 210321  
**Date Issued** 12/16/2021    **Date Expires** 12/16/2023

