

AS IV Serial no: 111774
Version no: 532B

TEST RECORD 00269

Temp	Date	Time	s/ 210L
------	------	------	------------

Air Blank:
09/04/21 05:55 .000
Calibration Check:
21 09/04/21 05:55 .099

Subject Name
Test #1
Subject I.D.

Operator Name, I.D.
B.C. Woods, 782
Location

AS IV Serial no: 111774
Version no: 532B

TEST RECORD 00270

Temp	Date	Time	s/ 210L
------	------	------	------------

Air Blank:
09/04/21 05:58 .000
Calibration Check:
21 09/04/21 05:58 .100

Subject Name
Test #2
Subject I.D.

Operator Name, I.D.
B.C. Woods, 782
Location

AS IV Serial no: 111774
Version no: 532B

TEST RECORD 00271

Temp	Date	Time	s/ 210L
------	------	------	------------

Air Blank:
09/04/21 05:59 .000
Calibration Check:
22 09/04/21 05:59 .099

Subject Name
Test #3
Subject I.D.

Operator Name, I.D.
B.C. Woods, 782
Location

AS IV Serial no: 111774
Version no: 532B

TEST RECORD 00272

Temp	Date	Time	s/ 210L
------	------	------	------------

VOID: RFI
12 09/04/21 06:00

Subject Name
RFI Test
Subject I.D.

Operator Name, I.D.
B.C. Woods, 782
Location

AS IV Serial no: 111774
Version no: 532B

TEST RECORD 00273

Temp	Date	Time	s/ 210L
------	------	------	------------

Air Blank:
09/04/21 06:03 .000
Subject Test: Auto
23 09/04/21 06:03 .000

Subject Name
Blank Test
Subject I.D.

Operator Name, I.D.
B.C. Woods, 782
Location



GUTH LABORATORIES, INC.

590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number **21080** of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on **March 10, 2021**, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain **0.1214%** (w/vol) ethyl alcohol. The expiration date for this lot number is **March 8, 2023** at 11:59 PM.

When used in a calibrated Simulator, operating at $34^{\circ}\text{C} \pm .2^{\circ}\text{C}$, this solution will give a breath alcohol analysis instrument reading of **0.100 g/210L** \pm 3%.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number FN05052002 whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.

