

RECEIVED

By Tracy Crews at 8:22 am, Jan 22, 2021



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES  
STATE PUBLIC HEALTH LABORATORY

ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN <b>111776</b>	NAME OF AGENCY <b>Kansas City MO PD</b>	DATE OF INSPECTION <b>01/08/2021</b>
LOCATION OF INSTRUMENT (STREET AND CITY) <b>9701 Marion Park Drive, Kansas City MO 64137</b>		TIME OF INSPECTION <b>2100</b>

**CHECKLIST:** Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)
- TEMPERATURE OF ALCO SENSOR (10°C - 40°C)
- PRINTER WORKING PROPERLY
- TIME AND DATE DISPLAYING PROPERLY

**BREATH ALCOHOL ACCURACY STANDARDS**

- SIMULATOR SOLUTION  COMPRESSED ETHANOL-GAS MIXTURE
- STANDARD SUPPLIER Intoximeters LOT # AG010103 EXP. DATE 04/10/2022
- SIMULATOR TEMPERATURE (34°C ± 0.2°C) \_\_\_\_\_ SIM. SN \_\_\_\_\_ SIM. NIST EXP DATE \_\_\_\_\_

- CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)  
Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)
- 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
- 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
- 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 ←	<b>.081</b>	TEST 2 ←	<b>.080</b>	TEST 3 ←	<b>.080</b>
----------	-------------	----------	-------------	----------	-------------

- RFI DETECTOR OPERATING


INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:  
(DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	<b>1</b>	(0-.04)	<b>0</b>	(.05-.09)	<b>0</b>	(.10-.14)	<b>2</b>	(.15-.19)	<b>1</b>	(OVER .19)	<b>4</b>
----------	----------	---------	----------	-----------	----------	-----------	----------	-----------	----------	------------	----------

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

TESTED AND CERTIFIED

**INSPECTING OFFICER**

SIGNATURE  #5771 208231 / 08/20/2022	PRINT NAME <b>PO Jeremy White</b>
TYPE II PERMIT NUMBER/EXPIRATION DATE <b>200231 08/20/2022</b>	TELEPHONE NUMBER <b>( ) 816-482-8141</b>

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office by mail, fax, or email.

AS IV Serial no: 111776  
Version no: 532B

TEST RECORD 00202

Temp Date Time 210L  
9/

Air Blank:  
01/08/21 21:00 .000  
Calibration Check:  
20 01/08/21 21:00 .081

Subject Name  
Test #1  
Subject I.D.

Operator Name, I.D.  
White #5771  
Location 200231  
08/20/2022

AS IV Serial no: 111776  
Version no: 532B

TEST RECORD 00203

Temp Date Time 210L  
9/

Air Blank:  
01/08/21 21:02 .000  
Calibration Check:  
21 01/08/21 21:02 .089

Subject Name  
Test #2  
Subject I.D.

Operator Name, I.D.  
White #5771  
Location 200231  
08/20/2022

AS IV Serial no: 111776  
Version no: 532B

TEST RECORD 00204

Temp Date Time 210L  
9/

Air Blank:  
01/08/21 21:04 .000  
Calibration Check:  
22 01/08/21 21:04 .080

Subject Name  
Test #3  
Subject I.D.

Operator Name, I.D.  
White #5771  
Location 200231  
08/20/2022

AS IV Serial no: 111776  
Version no: 532B

TEST RECORD 00205

Temp Date Time 210L  
9/

VOID: RFI  
12 01/08/21 21:08

Subject Name  
RFI Test  
Subject I.D.

Operator Name, I.D.  
White #5771  
Location 200231  
08/20/2022



STATE OF MISSOURI  
DEPARTMENT OF HEALTH AND SENIOR SERVICES  
BREATH ALCOHOL PROGRAM

**PERMIT  
TYPE II**

**JEREMY A. WHITE**

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

**ALCO-SENSOR IV WITH PRINTER, INTOXILYZER 8000**

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 8/20/2020

NUMBER 200231

EXPIRES 8/20/2022

MO 084-971 (6-10)

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

LAB-4 (6-10)



**Airgas**

Airgas USA LLC (LAB)  
3500 Bernard Street  
St. Louis, Mo. 63103  
Ph: (314) 593-3100  
Fax: (314) 593-7328

**Certificate of Analysis**

Test Date: 13-Apr-2020

Customer Name  
Exclusive Supplier  
Inoximeters, Inc.  
2081 Craig Road  
St. Louis, Mo 63146

Lot # AG010103 Model 108CADC

Exp. Date  
10-Apr-2022

Cyl. Type  
108

Component  
Ethanol  
Nitrogen

Certified Concentration  
0.082 ± 0.002 B/A/C (223 ppm)  
Balance

Certification Traceable to N.I.S.T. RGM and to CRM Ethanol Standards:

RGM Serial No.	Concentration	RGM Serial No.	Concentration
EB0010581	392.1 ppm	EB0010583	393.0 ppm
EB0010570	259.8 ppm	EB0010599	258.2 ppm
EB0010285	208.0 ppm	EB0010595	208.3 ppm
EB0010561	103.6 ppm	EB0010582	104.2 ppm
EB0010681	52.12 ppm	EB0010579	52.81 ppm
CRM Serial No.	Concentration	CRM Serial No.	Concentration
CC434668	800.0 ppm	0056649	390.1 ppm
CC234503	253.0 ppm	0056662	150.2 ppm

Analytical Method: **NDIR**

Digitally signed by Quality Control  
Date: 2020.04.14 13:17:41 -0500  
Reason: I am the author of this message  
Location: Alphas USA LLC (Lab)

Approved for Release:

*Rod Marsala*  
Rod Marsala

STATE OF MISSOURI  
DEPARTMENT OF HEALTH AND SENIOR SERVICES  
BREATH ALCOHOL PROGRAM

**INSTRUMENT OPERATOR CARD**

The named contributor is authorized to operate an additional breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator: **WHITE, JEREMY**  
Permit No: **200231**  
Date Issued: **8/20/2020**  
Date Expires: **8/20/2022**

ISO 17025:2005 A2LA accredited, Certificate Number 3082.06  
ISO 17034:2016 A2LA accredited, Certificate Number 3082.07