

RECEIVED

By Tracy Crews at 11:47 am, Feb 17, 2021



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES  
STATE PUBLIC HEALTH LABORATORY

ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 111778	NAME OF AGENCY Oak Grove Police Department	DATE OF INSPECTION 02/14/2021
LOCATION OF INSTRUMENT (STREET AND CITY) 2110 S Broadway, Oak Grove, MO 64075		TIME OF INSPECTION 5:08 pm

**CHECKLIST:** Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)
- TEMPERATURE OF ALCO SENSOR (10°C - 40°C)
- PRINTER WORKING PROPERLY
- TIME AND DATE DISPLAYING PROPERLY

**BREATH ALCOHOL ACCURACY STANDARDS**

- SIMULATOR SOLUTION  COMPRESSED ETHANOL-GAS MIXTURE
- STANDARD SUPPLIER Guth LOT # 20190 EXP. DATE 02/14/2021
- SIMULATOR TEMPERATURE (34°C ± 0.2°C) 34.0 SIM. SN SD734 SIM. NIST EXP DATE 09/22/2021

- CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)  
Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)
- 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
- 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
- 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 <u>.101</u>	TEST 2 <u>.099</u>	TEST 3 <u>.099</u>
--------------------	--------------------	--------------------

- RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	0	(0-.04)	0	(.05-.09)	0	(.10-.14)	0	(.15-.19)	0	(OVER .19)	0
----------	---	---------	---	-----------	---	-----------	---	-----------	---	------------	---

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

**INSPECTING OFFICER**

SIGNATURE 	PRINT NAME Sgt. Brandon S. Grote #603
TYPE II PERMIT NUMBER/EXPIRATION DATE 200252 09-24-2022	TELEPHONE NUMBER (816) 690-3773

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office by mail, fax, or email.

Serial no: 1117  
Version no: 5323

TEST RECORD 00244

Date Time 2021

By: Blank  
02/14/21 17:58 .000  
Calibration Check:  
02/14/21 17:53 .100

Project Name  
Test 1

Project ID  
B, Gmk

Generator Name, (L.D.)  
200252

Location  
9-24-22

DGPD

Serial no: 1117  
Version no: 5323

TEST RECORD 00245

Date Time 2021

By: Blank  
02/14/21 18:00 .000  
Calibration Check:  
02/14/21 18:00 .000

Project Name  
Test 2

Project ID  
B Gmk

Generator Name, (L.D.)  
200252

Location  
9-24-22

DGPD

Serial no: 1117  
Version no: 5323

TEST RECORD 00246

Date Time 2021

By: Blank  
02/14/21 18:01 .000  
Calibration Check:  
02/14/21 18:01 .000

Project Name  
Test 3

Project ID  
B Gmk

Generator Name, (L.D.)  
200252

Location  
9-24-22

DGPD

Serial no: 1117  
Version no: 5323

TEST RECORD 00247

Date Time 2021

By: Blank  
02/14/21 18:03

Project Name  
RFI Test

Project ID  
B, Gmk

Generator Name, (L.D.)  
200252

Location  
9-24-22

DGPD



**GUTH LABORATORIES, INC.**

590 NORTH 6TH STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

## CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number 20190 of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on April 8, 2020, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain 0.1206% (w/vol) ethyl alcohol. The expiration date for this lot number is April 6, 2022 at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of 0.100 g/210L +/- 3%.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President  
GUTH LABORATORIES, INC.

*NIST Traceability:*

*Testing was conducted using Cerilliant Reference Standard lot number FN02271802 whose values are traceable to NIST.*

*All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.*



STATE OF MISSOURI  
 DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 BREATH ALCOHOL PROGRAM



**PERMIT**  
**TYPE II**

**BRANDON GROTE**

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

**ALCO-SENSOR IV WITH PRINTER**

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE: 9/24/2020

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

NUMBER 200252

EXPIRES 9/24/2022

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO-580-0771 (6-10)

LAB-4 (R6-10)



**STATE OF MISSOURI**  
 DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 BREATH ALCOHOL PROGRAM

**INSTRUMENT OPERATOR CARD**

*The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.*

Operator GROTE, BRANDON  
 Permit No 200252  
 Date Issued 9/24/2020    Date Expires 9/24/2022

