



AS IV Serial no: 111780  
Version no: 532B

TEST RECORD 00480

Temp Date Time <sup>a/</sup> 210L

Air Blank:  
06/29/21 22:12 .000  
Calibration Check:  
26 06/29/21 22:12 .079

Subject Name

Test # 1

Subject I.D.

Operator Name, I.D.

Davis 210059

Location

AS IV Serial no: 111780  
Version no: 532B

TEST RECORD 00482

Temp Date Time <sup>a/</sup> 210L

Air Blank:  
06/29/21 22:14 .000  
Calibration Check:  
27 06/29/21 22:14 .081

Subject Name

Test # 3

Subject I.D.

Operator Name, I.D.

Davis 210059

Location

AS IV Serial no: 111780  
Version no: 532B

TEST RECORD 00481

Temp Date Time <sup>a/</sup> 210L

Air Blank:  
06/29/21 22:13 .000  
Calibration Check:  
26 06/29/21 22:13 .080

Subject Name

Test # 2

Subject I.D.

Operator Name, I.D.

Davis 210059

Location

AS IV Serial no: 111780  
Version no: 532B

TEST RECORD 00483

Temp Date Time <sup>a/</sup> 210L

VDIB: RFI  
12 06/29/21 22:16

Subject Name

RFI Test

Subject I.D.

Operator Name, I.D.

Davis 210059

Location



STATE OF MISSOURI  
DEPARTMENT OF HEALTH AND SENIOR SERVICES  
BREATH ALCOHOL PROGRAM



**PERMIT  
TYPE II**

**SHAWN E. DAVIS**

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

**ALCO-SENSOR IV WITH PRINTER, INTOXILYZER 8000**

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 4/6/2021  
 NUMBER 210059  
 EXPIRES 4/6/2023  
 MO 580 0771 (6-10)

*Shawn E. Davis*  
 DIRECTOR OF STATE PUBLIC HEALTH LABORATORY  
 LAG-4 (06-10)

STATE OF MISSOURI  
DEPARTMENT OF HEALTH AND SENIOR SERVICES  
BREATH ALCOHOL PROGRAM  
**INSTRUMENT OPERATOR CARD**

The named cardholder is authorized to operate the breathalyzer instrument in accordance with the provisions of the Missouri Alcohol Program.

Operator: **DAVIS, SHAWN**  
 Permit No: **210059**  
 Date Issued: **4/6/2021** Date Expires: **4/6/2023**




Airgas USA LLC (LAB)  
3500 Barrard Street  
St. Louis, Mo. 63103  
Ph: (314) 533-3100  
Fax: (314) 533-7328

**Certificate of Analysis**

Customer Name  
Exclusive Supplier  
Intoximeters, Inc.  
2081 Craig Road  
St. Louis, Mo 63146

Test Date: 13-Apr-2020

Lot # AG010103 Model 108cacc

Exp. Date  
10-Apr-2022

Cvl. Type  
108

Component  
Ethanol  
Nitrogen

Certified Concentration  
0.082 ± 0.002 BrAC (223 ppm)  
Balance

Certification Traceable to N.I.S.T. RGM and to CRM Ethanol Standards

| RGM Serial No. | Concentration | RGM Serial No. | Concentration |
|----------------|---------------|----------------|---------------|
| EB0010581      | 392.1 ppm     | EB0010503      | 393.0 ppm     |
| EB0010570      | 259.8 ppm     | EB0010559      | 258.2 ppm     |
| EB0010285      | 208.0 ppm     | EB0010595      | 208.3 ppm     |
| EB0010561      | 103.6 ppm     | EB0010562      | 104.2 ppm     |
| EB0010681      | 52.12 ppm     | EB0010579      | 52.81 ppm     |
| CRM Serial No. | Concentration | CRM Serial No. | Concentration |
| CC434668       | 800.0 ppm     | 0056649        | 390.1 ppm     |
| CC234503       | 253.0 ppm     | 0056662        | 150.2 ppm     |

Analytical Method: NDIR

Digitally signed by Quibby Castel  
 Date: 2020.04.14 13:17:54 -0500  
 Location: Airgas USA LLC  
 Location: Airgas USA LLC (LAB)

Approved for Release:

*Rod Marsala*  
 Rod Marsala

ISO 17025:2005 A2LA accredited. Certificate Number 3082.06  
 ISO 17034:2016 A2LA accredited. Certificate Number 3082.07