



**ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT**

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN <b>111780</b>	NAME OF AGENCY <b>Kansas City Police Department</b>	DATE OF INSPECTION <b>12/10/2021</b>
------------------------------------	--	---

LOCATION OF INSTRUMENT (STREET AND CITY) <b>9701 MARION PARK DR, KANSAS CITY</b>	TIME OF INSPECTION <b>1705 hours</b>
---	---

**CHECKLIST:** Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)

TEMPERATURE OF ALCO SENSOR (10°C - 40°C)

PRINTER WORKING PROPERLY

TIME AND DATE DISPLAYING PROPERLY

**BREATH ALCOHOL ACCURACY STANDARDS**

SIMULATOR SOLUTION  COMPRESSED ETHANOL-GAS MIXTURE

STANDARD SUPPLIER INTOXIMETERS LOT # AG010103 EXP. DATE 04/10/2022

SIMULATOR TEMPERATURE (34°C ± 0.2°C) SIM. SN SIM. NIST EXP DATE

CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE

0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE

0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 <b>.080</b>	TEST 2 <b>.079</b>	TEST 3 <b>.079</b>
--------------------	--------------------	--------------------

RFI DETECTOR OPERATING

**INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)**

REFUSALS	<b>0</b>	(0-.04)	<b>1</b>	(.05-.09)	<b>1</b>	(.10-.14)	<b>1</b>	(.15-.19)	<b>3</b>	(OVER .19)	<b>5</b>
----------	----------	---------	----------	-----------	----------	-----------	----------	-----------	----------	------------	----------

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

Instrument meets all DOHSS standards and guidelines.

**INSPECTING OFFICER**

SIGNATURE P.O. [Signature] 5396

PRINT NAME **Shawn Davis**

TYPE II PERMIT NUMBER/EXPIRATION DATE **210059 / 04/06/2023**

TELEPHONE NUMBER **( ) 816-234-5000**

**Return completed report to the:** Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office by mail, fax, or email.

AS IV Serial no: 111780  
Version no: 532B

TEST RECORD 00572

Temp Date Time <sup>a/</sup> 210L

Air Blank:  
12/10/21 17:05 .000  
Calibration Check:  
16 12/10/21 17:05 .000

Subject Name

Test #1

Subject I.D.

Operator Name, I.D.

Davis 210059  
Location

AS IV Serial no: 111780  
Version no: 532B

TEST RECORD 00573

Temp Date Time <sup>a/</sup> 210L

Air Blank:  
12/10/21 17:06 .000  
Calibration Check:  
18 12/10/21 17:06 .079

Subject Name

Test #2

Subject I.D.

Operator Name, I.D.

Davis 210059  
Location

AS IV Serial no: 111780  
Version no: 532B

TEST RECORD 00574

Temp Date Time <sup>a/</sup> 210L

Air Blank:  
12/10/21 17:08 .000  
Calibration Check:  
19 12/10/21 17:08 .079

Subject Name

Test #3

Subject I.D.

Operator Name, I.D.

Davis 210059  
Location

AS IV Serial no: 111780  
Version no: 532B

TEST RECORD 00575

Temp Date Time <sup>a/</sup> 210L

VOID: RFI  
12 12/10/21 17:09

Subject Name

RFI Test

Subject I.D.

Operator Name, I.D.

Davis 210059  
Location



STATE OF MISSOURI  
DEPARTMENT OF HEALTH AND SENIOR SERVICES  
BREATH ALCOHOL PROGRAM



**PERMIT**  
TYPE II

**SHAWN E. DAVIS**

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

**ALCO-SENSOR IV WITH PRINTER, INTOXILYZER 8000**

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.113 RSMo.

DATE 4/6/2021  
NUMBER 210059  
EXPIRES 4/6/2023

*W. S. Shaw*  
DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

*Shawn E. Davis*  
DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 390 0771 (6-10) L&4 (R6-10)

STATE OF MISSOURI  
DEPARTMENT OF HEALTH AND SENIOR SERVICES  
BREATH ALCOHOL PROGRAM  
**INSTRUMENT OPERATOR CARD**

The holder of this instrument operator card is authorized to operate the instrument for the determination of the alcoholic content of a sample of expired air in Missouri.

Operator: **DAVIS, SHAWN**  
Date Issued: **4/6/2021** Date Expires: **4/6/2023**



Airgas USA LLC (LAB)  
3500 Bernard Street  
St. Louis, Mo. 63103  
Ph: (314) 533-3100  
Fax: (314) 533-7328

**Certificate of Analysis**

Customer Name  
**Exclusive Supplier**  
Intoximeters, Inc.  
2081 Craig Road  
St. Louis, Mo 63146

Test Date: 13-Apr-2020

Lot # AG010103 Model 108cacd

Exp. Date 10-Apr-2022  
Cvl. Type 108  
Component Ethanol Nitrogen  
Certified Concentration 0.082 ± 0.002 BrAC (223 ppm)  
Balance

Certification Traceable to N.I.S.T. RGM and to CRM Ethanol Standards:

RGM Serial No.	Concentration	RGM Serial No.	Concentration
EB0010581	392.1 ppm	EB0010603	393.0 ppm
EB0010570	259.8 ppm	EB0010559	258.2 ppm
EB0010285	208.0 ppm	EB0010595	208.3 ppm
EB0010561	103.6 ppm	EB0010562	104.2 ppm
EB0010681	52.12 ppm	EB0010579	52.81 ppm
CRM Serial No.	Concentration	CRM Serial No.	Concentration
CC434668	800.0 ppm	0056649	390.1 ppm
CC234503	253.0 ppm	0056662	150.2 ppm

Analytical Method: NDIR

Sublot obtained by Quality Control  
Date: 20-Apr-2020  
Reason: Dry gas extracted certification of analysis  
Location: Airgas USA LLC (Lab)

Approved for Release:

*Rod Marsala*  
Rod Marsala

ISO 17025:2005 A2LA accredited. Certificate Number 3082.06  
ISO 17034:2016 A2LA accredited. Certificate Number 3082.07