

RECEIVED

By Tracy Crews at 7:49 am, Nov 30, 2021



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES  
STATE PUBLIC HEALTH LABORATORY  
BREATH ALCOHOL PROGRAM

INTOX EC/IR II MAINTENANCE REPORT

REPORT #3

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 35 days). Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service. Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

INTOX EC/IR II SN 12841	NAME OF AGENCY GREENE COUNTY S/O	DATE OF INSPECTION 10/29/2021
----------------------------	-------------------------------------	----------------------------------

LOCATION OF INSTRUMENT (STREET AND CITY) 1010 N. BOONVILLE AVE SPRINGFIELD	TIME OF INSPECTION 10:23 CDT
---	---------------------------------

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or is operating within established limits. (Write in observed values where determined). Unmarked items must be corrected before using instrument.

<input checked="" type="checkbox"/> DIAGNOSTIC RECORD	
<input checked="" type="checkbox"/> BLANK CHECK	<input checked="" type="checkbox"/> CO2 CHECK
<input checked="" type="checkbox"/> FC 1 TEMP	<input checked="" type="checkbox"/> FLOW CHECK
<input checked="" type="checkbox"/> SRC TEMP	<input checked="" type="checkbox"/> FCB CHECK
<input checked="" type="checkbox"/> DET TEMP	<input checked="" type="checkbox"/> CRC COMP CHECK
<input checked="" type="checkbox"/> BT TEMP	<input checked="" type="checkbox"/> CRC CAL CHECK
<input checked="" type="checkbox"/> STD 2 TEMP	<input checked="" type="checkbox"/> PRINT TEST
<input checked="" type="checkbox"/> ETH CHECK	

BREATH ANALYZER ACCURACY STANDARDS

<input type="checkbox"/> SIMULATOR SOLUTION	<input checked="" type="checkbox"/> COMPRESSED ETHANOL-GAS MIXTURE				
<input checked="" type="checkbox"/> STANDARD SUPPLIER	INTOXIMETERS	LOT#	AG111803	EXP. DATE	04/28/2023
<input type="checkbox"/> SIMULATOR TEMP (34°C ±0.2°C)	SIM. SN	SIM. NIST EXP DATE			

CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within +5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used.

0.10% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE

0.08% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE

0.04% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

TEST 1 → 0.102 g/210L	TEST 2 → 0.101 g/210L	TEST 3 → 0.101 g/210L
-----------------------	-----------------------	-----------------------

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:

REFUSALS	11	0-.04	54	.05-.09	1	.10-.14	0	.15-.19	3	OVER .19	0
----------	----	-------	----	---------	---	---------	---	---------	---	----------	---

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY).

INSPECTING OFFICER	
SIGNATURE ▶ <i>Ronald Killingsworth</i>	PRINT FULL NAME KILLINGSWORTH/RONALD
TYPE II PERMIT NUMBER 210003	EXPIRATION DATE 01/07/2023
TELEPHONE NUMBER ( 417 ) 868-4040	

RETURN COMPLETED REPORT TO THE:  
Breath Alcohol Program, Missouri Department of Health and Senior Services,  
by mail, fax, or e-mail



Airgas USA LLC (LAB)  
 3500 Bernard Street  
 St. Louis, Mo. 63103  
 Ph: (314) 533-3100  
 Fax: (314) 533-7328

**Certificate of Analysis**

**Customer Name**  
 Exclusive Supplier  
 Intoximeters, Inc.  
 2081 Craig Road  
 St. Louis, Mo 63146

**Test Date:** 28-Apr-2021

**Lot # AG111803 Model 108caccd**

<b><u>Exp. Date</u></b> 28-Apr-2023	<b><u>Cyl. Type</u></b> 108	<b><u>Component</u></b> Ethanol Nitrogen	<b><u>Certified Concentration</u></b> 0.100 ± 2% BrAC (272 ppm) Balance
--	--------------------------------	--	---

**Certification Traceable to N.I.S.T. RGM and to CRM Ethanol Standards:**

<b><u>RGM Serial No.</u></b> EB0010581 EB0010570 EB0010285 EB0010561 EB0010681	<b><u>Concentration</u></b> 392.1 ppm 259.8 ppm 208.0 ppm 103.6 ppm 52.12 ppm	<b><u>RGM Serial No.</u></b> EB0010603 EB0010559 EB0010595 EB0010562 EB0010579	<b><u>Concentration</u></b> 393.0 ppm 258.2 ppm 208.3 ppm 104.2 ppm 52.81 ppm
<b><u>CRM Serial No.</u></b> CC434668 CC234503	<b><u>Concentration</u></b> 800.0 ppm 253.0 ppm	<b><u>CRM Serial No.</u></b> 0056649 0056662	<b><u>Concentration</u></b> 390.1 ppm 150.2 ppm

**Analytical Method:** NDIR

Digitally signed by Quality Control  
 Date: 2021.04.28 18:40:40 -05:00  
 Reason: Dry gas standard certification of analysis  
 Location: Airgas USA LLC (Lab)

Approved for Release:

Rod Marsala

ISO 17025:2017 A2LA accredited. Certificate Number 3082.06  
 ISO 17034:2016 A2LA accredited. Certificate Number 3082.07



STATE OF MISSOURI  
DEPARTMENT OF HEALTH AND SENIOR SERVICES  
BREATH ALCOHOL PROGRAM

2

**PERMIT  
TYPE II**

**RONALD L. KILLINGSWORTH**

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

**ALCO-SENSOR IV WITH PRINTER, INTOX EC/IR II**

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 1/7/2021

NUMBER 210003

EXPIRES 1/7/2023

MO 885-0771 (6-18)

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

LAB-4 (06-19)

STATE OF MISSOURI  
DEPARTMENT OF HEALTH AND SENIOR SERVICES  
BREATH ALCOHOL PROGRAM

**INSTRUMENT OPERATOR CARD**

The named operator is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath from of expired air in Missouri.

Operator **KILLINGSWORTH, RONALD**  
Permit No **210003**  
Date Issued **1/7/2021** Date Expires **1/7/2023**