



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY
BREATH ALCOHOL PROGRAM
INTOX DMT MAINTENANCE REPORT

REPORT #1

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 35 days).
Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service.
Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

INTOX DMT SN 500297	NAME OF AGENCY Pacific Police Department	DATE OF INSPECTION 08/23/2021
LOCATION OF INSTRUMENT (STREET AND CITY) 300 Hoven Pacific, MO 63069		TIME OF INSPECTION 08:08:37

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or is operating within established limits. (Write in observed values where determined). Unmarked items must be corrected before using instrument.

DIAGNOSTIC RECORD

DATE AND TIME <u>08/23/2021 08:08:39</u>	<input checked="" type="checkbox"/> DETECTOR
<input checked="" type="checkbox"/> PROGRAM	<input checked="" type="checkbox"/> FILTER 1
<input checked="" type="checkbox"/> SAMPLE CHAMBER <u>48.7°C</u>	<input checked="" type="checkbox"/> FILTER 2
<input checked="" type="checkbox"/> BREATH TUBE <u>48.1°C</u>	<input checked="" type="checkbox"/> FILTER 3
<input checked="" type="checkbox"/> PUMP	<input checked="" type="checkbox"/> INTERNAL STANDARD

BREATH ANALYZER ACCURACY STANDARDS

<input type="checkbox"/> SIMULATOR STANDARD	<input checked="" type="checkbox"/> COMPRESSED ETHANOL-GAS MIXTURE
---	--

<input checked="" type="checkbox"/> STANDARD SUPPLIER <u>INTOXIMETERS</u>	LOT # <u>AG019502</u>	EXP. DATE <u>07/13/2022</u>
<input type="checkbox"/> SIMULATOR TEMP (34°C ± 0.2°C) _____	SIM. SN _____	SIM. NIST EXP DATE _____

CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)
Run three tests using a standard. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard being used.

0.10% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE

0.08% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE

0.04% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

TEST 1: 0.102	TEST 2: 0.101	TEST 3: 0.101
----------------------	----------------------	----------------------

PERFORM R.F.I. TEST

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:

REFUSALS: 0	0-.04: 0	.05-.09: 0	.10-.14: 0	.15-.19: 0	OVER .19: 0
--------------------	-----------------	-------------------	-------------------	-------------------	--------------------

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY)

INSPECTING OFFICER

SIGNATURE <i>JH Klingle</i>	PRINT FULL NAME JAMES H KLINGLER	
TYPE II PERMIT NUMBER 200236	EXPIRATION DATE 08/25/2022	TELEPHONE NUMBER 636-221-1546

RETURN COMPLETED REPORT TO THE **Breath Alcohol Program, Missouri Department of Health and Senior Services**
by mail, fax, or email



Airgas USA LLC (LAB)

3500 Bernard Street

St. Louis, Mo. 63103

Ph: (314) 533-3100

Fax: (314) 533-7328

Certificate of Analysis

Customer Name

Exclusive Supplier

Intoximeters, Inc.

2081 Craig Road

St. Louis, Mo 63146

Test Date: 15-Jul-2020

Lot # AG019502 **Model** 108caccd

Exp. Date

13-Jul-2022

Cyl. Type

108

Component

Ethanol

Nitrogen

Certified Concentration

0.100 ± 2% BrAC (260 ppm)

Balance

Certification Traceable to N.I.S.T. RGM and to CRM Ethanol Standards:

RGM Serial No.

EB0010581

Concentration

392.1 ppm

EB0010570

259.8 ppm

EB0010285

208.0 ppm

EB0010561

103.6 ppm

EB0010681

52.12 ppm

RGM Serial No.

EB0010603

Concentration

393.0 ppm

EB0010559

258.2 ppm

EB0010595

208.3 ppm

EB0010562

104.2 ppm

EB0010579

52.81 ppm

CRM Serial No.

CC434668

Concentration

800.0 ppm

CC234503

253.0 ppm

CRM Serial No.

0056649

Concentration

390.1 ppm

0056662

150.2 ppm

Analytical Method:

NDIR

Digitally signed by Quality Control
Date: 2020.07.15 14:50:01 -05:00
Reason: Dry gas standard certification of analysis
Location: Airgas USA LLC (Lab)

Approved for Release:

Rod Marsala

ISO 17025:2005 A2LA accredited. Certificate Number 3082.06
ISO 17034:2016 A2LA accredited. Certificate Number 3082.07



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

2

**PERMIT
TYPE II**

JAMES H. KLINGLER

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

INTOX DMT

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 8/25/2020

NUMBER 200236

EXPIRES 8/25/2022

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.



Operator KLINGLER, JAMES
Permit No 200236
Date Issued 8/25/2020 Date Expires 8/25/2022

STATE OF MISSOURI)
)
COUNTY OF FRANKLIN) SS

AFFIDAVIT FOR RECORDS


Before me, the undersigned authority personally appeared, Sgt. James H Klingler 131, and upon being duly sworn by me, deposed as follows:

My name is Sgt. James Klingler, 131, I am of sound mind, capable of making this affidavit, and personally acquainted with the facts herein stated:

I am the custodian of records of the Pacific Police Department's, **INTOX DMT SN# 500297**. Attached hereto are **3** pages of records for the **INTOX DMT SN# 500297** from the Pacific Police Department for the month of July 23rd 2021 to August 23rd 2021. These pages for the instrument are kept by the Pacific Police Department in the regular course of business, and it is with the regular course of business that an employee of representative to make the record or to transmit information thereof to be included in such records. The records attached hereto are the original or exact duplicates of any and all original records kept by the Pacific Police Department in the regular course of business.

Sgt. James Klingler

Affiant's Name-typed of printed



Affiant's Signature

In witness whereof, I have hereunto subscribed my name affixed my official seal this,

24th day of AUGUST, 2021

My Commission expires: 5/31/25

Angela B. Downing

Notary Public

