

RECEIVED

By Tracy Crews at 7:50 am, Nov 30, 2021



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY
CMI INTOXILYZER 8000 MAINTENANCE REPORT

REPORT #2

Complete this report in duplicate at the time of the regular monthly preventive maintenance check, and whenever instrument is repaired. Send one copy to Department of Health and Senior Services, and retain one copy in department file.

INSTRUMENT SERIAL NUMBER 80-005829	LOCATION OF INSTRUMENT JASPER CO. SHERIFF	DATE OF INSPECTION 09/14/2021	TIME OF INSPECTION 07:50
---------------------------------------	----------------------------------------------	----------------------------------	-----------------------------

CALIBRATION CHECK RESULTS			CALIBRATION CHECK SUMMARY		
Test	g/210L	Time	STANDARD TYPE	STANDARD LOT #	STANDARD EXPIRATION DATE
Air Blank	0.000	07:52	DRY	AG016702	06/15/2022
Cal Check	0.079	07:52	SIM TEMPERATURE	SIM SERIAL NUMBER	SIM CERTIFICATE EXPIRATION
Air Blank	0.000	07:53	N/A	N/A	N/A
Cal Check	0.078	07:53	STANDARD VALUE	STANDARD SUPPLIER	
Air Blank	0.000	07:53	0.080	INTOXIMETERS	
Cal Check	0.079	07:54	CALIBRATION CHECK RESULT 1		
Air Blank	0.000	07:54	0.079		
Cal Check	0.079	07:54	CALIBRATION CHECK RESULT 2		
Air Blank	0.000	07:54	0.078		
Pass			CALIBRATION CHECK RESULT 3		
			0.079		
			MAXIMUM DEVIATION (MUST BE WITHIN 5%)	SPREAD (MUST BE .005 OR LESS)	
			2.5%	0.001	

DIAGNOSTIC TEST RESULTS			RFI TEST RESULTS		
Test	Pass	Time	Test	g/210L	Time
Voltage/Current Test	Pass		Air Blank	0.000	07:55
RAM Test	Pass		Subject Test	RFI*	07:55
EEPROM Checksum Test	Pass		Air Blank	0.000	07:55
Real Time Clock Test	Pass		*RFI Detect		
DSP Test	Pass		Pass		
Analytical Stability Test	Pass				
Modem Test	Pass				
Temperature Regulation Test	Pass				

NUMBER OF REFUSALS AND SUBJECT BREATH TESTS IN EACH RANGE SINCE LAST MAINTENANCE REPORT						
REFUSALS	.00-.04	.05-.09	.10-.14	.15-.19	OVER .19	
0	0	0	0	0	0	0

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

INSPECTING OFFICER		
SIGNATURE 	PRINT NAME JOHN HICKS	
TYPE II PERMIT NUMBER 290216	EXPIRATION DATE 09/25/2021	TELEPHONE NUMBER 4173588177