

RECEIVED

By Tracy Crews at 3:36 pm, Dec 20, 2021



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY
CMI INTOXILYZER 8000 MAINTENANCE REPORT

REPORT #2

Complete this report in duplicate at the time of the regular monthly preventive maintenance check, and whenever instrument is repaired. Send one copy to Department of Health and Senior Services, and retain one copy in department file.

INSTRUMENT SERIAL NUMBER 80-005834	LOCATION OF INSTRUMENT KCMO POLICE DEPARTME	DATE OF INSPECTION 12/01/2021	TIME OF INSPECTION 22:20
---------------------------------------	--	----------------------------------	-----------------------------

CALIBRATION CHECK RESULTS			CALIBRATION CHECK SUMMARY		
Test	g/210L	Time	STANDARD TYPE	STANDARD LOT #	STANDARD EXPIRATION DATE
Air Blank	0.000	22:22	DRY	14020080A2	07/05/2022
Cal Check	0.079	22:22	SIM TEMPERATURE	SIM SERIAL NUMBER	SIM CERTIFICATE EXPIRATION
Air Blank	0.000	22:23	N/A	N/A	N/A
Cal Check	0.080	22:23	STANDARD VALUE	STANDARD SUPPLIER	
Air Blank	0.000	22:24	0.080	INTOXIMETERS	
Cal Check	0.079	22:24	CALIBRATION CHECK RESULT 1		
Air Blank	0.000	22:25	0.079		
Pass			CALIBRATION CHECK RESULT 2		
			0.080		
			CALIBRATION CHECK RESULT 3		
			0.079		
			MAXIMUM DEVIATION (MUST BE WITHIN 5%)	SPREAD (MUST BE .005 OR LESS)	
			1.2%	0.001	

DIAGNOSTIC TEST RESULTS			RFI TEST RESULTS		
			Test	g/210L	Time
Voltage/Current Test	Pass		Air Blank	0.000	22:25
RAM Test	Pass		Subject Test	RFI*	22:25
EEPROM Checksum Test	Pass		Air Blank	0.000	22:26
Real Time Clock Test	Pass		*RFI Detect		
DSP Test	Pass		Pass		
Analytical Stability Test	Pass				
Modem Test	Pass				
Temperature Regulation Test	Pass				

NUMBER OF REFUSALS AND SUBJECT BREATH TESTS IN EACH RANGE SINCE LAST MAINTENANCE REPORT						
REFUSALS	.00-.04	.05-.09	.10-.14	.15-.19	OVER .19	
0	0	0	0	0	0	0

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).
TESTED AND CERTIFIED

INSPECTING OFFICER

SIGNATURE: *[Signature]* #5771 200231 08/20/2022 PRINT NAME: JEREMY WHITE

TYPE II PERMIT NUMBER: 200231 EXPIRATION DATE: 08/20/2022 TELEPHONE NUMBER: 8164828141



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM



PERMIT
TYPE II

JEREMY A. WHITE

herby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER, INTOXILYZER 8000

-the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 7.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE: 8/20/2020

PERMIT NUMBER: 200231

ISSUED: 8/20/2022

ISSUED BY: [Signature]

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

LS-4 (8-10)



7 Eastgate Dr. • P.O. Box 790 • Jacksonvile, IL 62651-0790
217-245-2183 • Fax: 217-243-7634 • www.ilmo-products.com

Certificate of Analysis

Certificate ID: 13921
Part #: BAC105L080T
Cylinder Size: 105L
Lot Number: 14020808A2
Expiration: 7/5/2022

0.080 BAC (for the calibration of instruments used to determine breath alcohol concentration)

Contents: 105 Liters @ 1000 psig 70°F (21°C)

Component	Reported Concentration	Analytical Accuracy (U, k=2)	Analytical Method
Ethanol	208 ppm	± 4.86% (6.26% NDIR)	NDIR
Nitrogen	Balance	± 3.2 ppm	

Distributed by:

CMi Inc
316 East Ninth Street
Owensboro, KY 42303
Phone 856-835-0690
www.alcoholtest.com

*Not acceptable to:
Certified Reference Material - 262.4 µmol/mol
Ethanol in Nitrogen - Serial No. GN0015026 Lot No. 050319E11

Store in dry area, away from sources of heat, ignition and direct sunlight. Do not allow storage area to exceed 52 °C (125 °F).

Specialty Gas Lab Tech
[Signature]

06-15-2020
Issuance Date



The calibration results within this certificate were obtained using equipment and standards capable of providing analytical results traceable to NIST and apply only to the items specified. The information use is at the sole discretion and risk of the user. Liability shall be limited to established replacement cost of this material or service.

ISO/IEC 17025:2017 Accredited Laboratory