

RECEIVED

By Tracy Crews at 7:46 am, Jun 04, 2021

MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY**CMI INTOXILYZER 8000 MAINTENANCE REPORT**

MPD

REPORT #2

Complete this report in duplicate at the time of the regular monthly preventive maintenance check, and whenever instrument is repaired. Send one copy to Department of Health and Senior Services, and retain one copy in department file.

INSTRUMENT SERIAL NUMBER 80-005838		LOCATION OF INSTRUMENT KCMO POLICE DEPT.		DATE OF INSPECTION 05/15/2021		TIME OF INSPECTION 19:31	
CALIBRATION CHECK RESULTS				CALIBRATION CHECK SUMMARY			
Test	g/210L	Time		STANDARD TYPE DRY	STANDARD LOT # 14020080A2	STANDARD EXPIRATION DATE 07/05/2022	
Air Blank	0.000	19:32		SIM TEMPERATURE N/A	SIM SERIAL NUMBER N/A	SIM CERTIFICATE EXPIRATION N/A	
Cal Check	0.082	19:33		STANDARD VALUE 0.080	STANDARD SUPPLIER CMI		
Air Blank	0.000	19:33		CALIBRATION CHECK RESULT 1 0.082			
Cal Check	0.080	19:34		CALIBRATION CHECK RESULT 2 0.080			
Air Blank	0.000	19:34		CALIBRATION CHECK RESULT 3 0.080			
Cal Check	0.080	19:34		MAXIMUM DEVIATION (MUST BE WITHIN 5%) 2.5%		SPREAD (MUST BE .005 OR LESS) 0.002	
Air Blank	0.000	19:35					
Pass							
DIAGNOSTIC TEST RESULTS				RFI TEST RESULTS			
Voltage/Current Test Pass				Test	g/210L	Time	
RAM Test Pass				Air Blank	0.000	19:35	
EEPROM Checksum Test Pass				Subject Test	RFI*	19:36	
Real Time Clock Test Pass				Air Blank	RFI*	19:36	
DSP Test Pass				*RFI Detect			
Analytical Stability Test Pass				Pass			
Modem Test Pass							
Temperature Regulation Test Pass							
Pass							
NUMBER OF REFUSALS AND SUBJECT BREATH TESTS IN EACH RANGE SINCE LAST MAINTENANCE REPORT							
REFUSALS	.00-.04	.05-.09	.10-.14	.15-.19	OVER .19		
0	2	4	2	3	5		
List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary). TESTED AND CERTIFIED.							
INSPECTING OFFICER							
SIGNATURE <i>PO. S. K. 5396</i>				PRINT NAME SHAWN DAVIS			
TYPE II PERMIT NUMBER 210059		EXPIRATION DATE 04/06/2023		TELEPHONE NUMBER 816-234-5000			