



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES  
STATE PUBLIC HEALTH LABORATORY  
**CMI INTOXILYZER 8000 MAINTENANCE REPORT**

REPORT #2

Complete this report in duplicate at the time of the regular monthly preventive maintenance check, and whenever instrument is repaired. Send one copy to Department of Health and Senior Services, and retain one copy in department file.

INSTRUMENT SERIAL NUMBER 80-005837	LOCATION OF INSTRUMENT KANSAS CITY PD	DATE OF INSPECTION 02/05/2021	TIME OF INSPECTION 19:44
---------------------------------------	--	----------------------------------	-----------------------------

CALIBRATION CHECK RESULTS			CALIBRATION CHECK SUMMARY		
Test	g/210L	Time	STANDARD TYPE	STANDARD LOT #	STANDARD EXPIRATION DATE
Air Blank	0.000	19:49	DRY	14020080A2	07/05/2022
Cal Check	0.082	19:49	SIM TEMPERATURE	SIM SERIAL NUMBER	SIM CERTIFICATE EXPIRATION
Air Blank	0.000	19:49	N/A	N/A	N/A
Cal Check	0.083	19:50	STANDARD VALUE	STANDARD SUPPLIER	
Air Blank	0.000	19:50	0.080	CMI	
Cal Check	0.082	19:51	CALIBRATION CHECK RESULT 1		
Air Blank	0.000	19:51	0.082		
<b>Pass</b>			CALIBRATION CHECK RESULT 2		
			0.083		
			CALIBRATION CHECK RESULT 3		
			0.082		
			MAXIMUM DEVIATION (MUST BE WITHIN 5%)		SPREAD (MUST BE .005 OR LESS)
			3.7%		0.001

DIAGNOSTIC TEST RESULTS			RFI TEST RESULTS		
Test	Pass/Fail	Time	Test	g/210L	Time
Voltage/Current Test	Pass		Air Blank	0.000	19:52
RAM Test	Pass		Subject Test	RFI*	19:52
EEPROM Checksum Test	Pass		Air Blank	0.000	19:52
Real Time Clock Test	Pass		*RFI Detect		
DSP Test	Pass		<b>Pass</b>		
Analytical Stability Test	Pass				
Modem Test	Pass				
Temperature Regulation Test	Pass				
<b>Pass</b>					

**NUMBER OF REFUSALS AND SUBJECT BREATH TESTS IN EACH RANGE SINCE LAST MAINTENANCE REPORT**

REFUSALS	.00-.04	.05-.09	.10-.14	.15-.19	OVER .19
0	4	0	3	4	6

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

TESTED AND CERTIFIED

**INSPECTING OFFICER**

SIGNATURE: *[Signature]* #5771 2022/08/20/2022 PRINT NAME: JEREMY WHITE

TYPE II PERMIT NUMBER 200231	EXPIRATION DATE 08/20/2022	TELEPHONE NUMBER 8164828141
---------------------------------	-------------------------------	--------------------------------

