

RECEIVED

By Tracy Crews at 7:46 am, Jun 04, 2021



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY
CMI INTOXILYZER 8000 MAINTENANCE REPORT

Bat Cave

REPORT #2

Complete this report in duplicate at the time of the regular monthly preventive maintenance check, and whenever instrument is repaired. Send one copy to Department of Health and Senior Services, and retain one copy in department file.

INSTRUMENT SERIAL NUMBER 80-005839	LOCATION OF INSTRUMENT KCMO POLICE DEPT	DATE OF INSPECTION 05/15/2021	TIME OF INSPECTION 17:50
---------------------------------------	--	----------------------------------	-----------------------------

CALIBRATION CHECK RESULTS			CALIBRATION CHECK SUMMARY		
Test	g/210L	Time	STANDARD TYPE DRY	STANDARD LOT # 14020080A2	STANDARD EXPIRATION DATE 07/05/2022
Air Blank	0.000	17:51	SIM TEMPERATURE N/A	SIM SERIAL NUMBER N/A	SIM CERTIFICATE EXPIRATION N/A
Cal Check	0.081	17:52	STANDARD VALUE 0.080	STANDARD SUPPLIER CMI	
Air Blank	0.000	17:52	CALIBRATION CHECK RESULT 1 0.081		
Cal Check	0.080	17:52	CALIBRATION CHECK RESULT 2 0.080		
Air Blank	0.000	17:53	CALIBRATION CHECK RESULT 3 0.080		
Cal Check	0.080	17:53	MAXIMUM DEVIATION (MUST BE WITHIN 5%) 1.2%	SPREAD (MUST BE .005 OR LESS) 0.001	
Air Blank	0.000	17:54			

Pass

DIAGNOSTIC TEST RESULTS			RFI TEST RESULTS		
Voltage/Current Test	Pass		Test	g/210L	Time
RAM Test	Pass		Air Blank	0.000	17:54
EEPROM Checksum Test	Pass		Subject Test	RFI*	17:55
Real Time Clock Test	Pass		Air Blank	0.000	17:55
DSP Test	Pass		*RFI Detect		
Analytical Stability Test	Pass				
Modem Test	Pass				
Temperature Regulation Test	Pass				

Pass**Pass**

NUMBER OF REFUSALS AND SUBJECT BREATH TESTS IN EACH RANGE SINCE LAST MAINTENANCE REPORT						
REFUSALS	.00-.04	.05-.09	.10-.14	.15-.19	OVER .19	
0	0	0	0	0	0	0

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

TESTED AND CERTIFIED.

INSPECTING OFFICER		
SIGNATURE 	PRINT NAME SHAWN DAVIS	
TYPE II PERMIT NUMBER 210059	EXPIRATION DATE 04/06/2023	TELEPHONE NUMBER 816-234-5000