



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY
CMI INTOXILYZER 8000 MAINTENANCE REPORT

REPORT #2

Complete this report in duplicate at the time of the regular monthly preventive maintenance check, and whenever instrument is repaired. Send one copy to Department of Health and Senior Services, and retain one copy in department file.

INSTRUMENT SERIAL NUMBER 80-005839	LOCATION OF INSTRUMENT KCMO POLICE DEPT	DATE OF INSPECTION 09/07/2021	TIME OF INSPECTION 04:13
---------------------------------------	--	----------------------------------	-----------------------------

CALIBRATION CHECK RESULTS			CALIBRATION CHECK SUMMARY		
Test	g/210L	Time	STANDARD TYPE DRY	STANDARD LOT # 14020080A2	STANDARD EXPIRATION DATE 07/05/2022
Air Blank	0.000	04:15	SIM TEMPERATURE N/A	SIM SERIAL NUMBER N/A	SIM CERTIFICATE EXPIRATION N/A
Cal Check	0.081	04:15	STANDARD VALUE 0.080	STANDARD SUPPLIER CMI	
Air Blank	0.000	04:15	CALIBRATION CHECK RESULT 1 0.081		
Cal Check	0.080	04:16	CALIBRATION CHECK RESULT 2 0.080		
Air Blank	0.000	04:16	CALIBRATION CHECK RESULT 3 0.080		
Cal Check	0.080	04:17	MAXIMUM DEVIATION (MUST BE WITHIN 5%) 1.2%	SPREAD (MUST BE .005 OR LESS) 0.001	
Air Blank	0.000	04:17	<h1>Pass</h1>		

DIAGNOSTIC TEST RESULTS			RFI TEST RESULTS		
Voltage/Current Test	Pass		Test	g/210L	Time
RAM Test	Pass		Air Blank	RFI*	04:17
EEPROM Checksum Test	Pass		Air Blank	0.000	04:18
Real Time Clock Test	Pass		<h1>Pass</h1>		
DSP Test	Pass				
Analytical Stability Test	Pass				
Modem Test	Pass				
Temperature Regulation Test	Pass				

NUMBER OF REFUSALS AND SUBJECT BREATH TESTS IN EACH RANGE SINCE LAST MAINTENANCE REPORT						
REFUSALS	.00-.04	.05-.09	.10-.14	.15-.19	OVER .19	
1	0	0	0	0	0	0

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).
TESTED AND CERTIFIED

INSPECTING OFFICER		
SIGNATURE 	PRINT NAME INFRANCA, JORDAN	
TYPE II PERMIT NUMBER 210128	EXPIRATION DATE 06/21/2023	TELEPHONE NUMBER 816-382-5897



7 Eastgate Dr. • P.O. Box 790 • Jackson, IL 62651-0790
217-245-2183 • Fax: 217-243-7634 • www.ilmo-products.com

Certificate of Analysis

certificate ID: 13821
part #: BAC1051080T
cylinder Size: 105L
lot Number: 14020080A2
expiration: 7/5/2022

080 BAC (for the calibration of instruments used to determine breath alcohol concentration)

Component	Reported Concentration	Analytical Accuracy (U, I=2)	Analytical Method
nitrogen	288 ppm	+/- 0.02 BAC(0.20L)	NDR
nitrogen	Balance	(5.2 ppm)	

Distributed by:
CMI Inc.
316 East Ninth Street
Owensboro, KY 42303
Phone 866-835-0690
www.alcoholtest.com

receivable to:
certified Reference Material - 2624 µmol/mol
nitrogen in Nitrogen - Serial No. GN0015026 Lot No. 050319E11

Store in dry area, away from sources of heat,
ignition and direct sunlight. Do not allow storage
area to exceed 52 °C (125 °F).

[Signature]
Quality Gas Lab Tech

06-15-2020
Issuance Date



ISO/IEC 17025:2017 Accredited Laboratory

calibration results within this certificate were obtained using equipment and methods capable of producing analytical results traceable to NIST, and apply only to the items stated on this certificate. LHO Products Company makes no warranty or representation as to the suitability of the use of any information provided for any particular purpose. The information is to be used at the discretion and risk of the user, liability shall be limited or established replacement cost of the material or service.



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

PERMIT TYPE II

JORDAN INFRANCA

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER, INTOXILYZER 8000

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE: 6/21/2010
NUMBER: 290130
EXPIRES: 6/21/2021
MO 580-071 (6-10)

[Signature]
DIRECTOR OF STATE PUBLIC HEALTH LABORATORY
DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES
LAB-4 (6-10)

STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM
INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate and maintain breath alcohol instruments for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator: INFRANCA, JORDAN
Permit No: 290130
Date Issued: 6/21/2010
Date Expires: 6/21/2021