



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY
CMI INTOXILYZER 8000 MAINTENANCE REPORT

REPORT #2

Complete this report in duplicate at the time of the regular monthly preventive maintenance check, and whenever instrument is repaired. Send one copy to Department of Health and Senior Services, and retain one copy in department file.

INSTRUMENT SERIAL NUMBER 80-005840	LOCATION OF INSTRUMENT KCMO POLICE DEPT.	DATE OF INSPECTION 02/05/2021	TIME OF INSPECTION 17:26
---------------------------------------	---------------------------------------------	----------------------------------	-----------------------------

CALIBRATION CHECK RESULTS			CALIBRATION CHECK SUMMARY		
Test	g/210L	Time	STANDARD TYPE	STANDARD LOT #	STANDARD EXPIRATION DATE
Air Blank	0.000	17:28	DRY	14020080A2	07/05/2022
Cal Check	0.083	17:28	SIM TEMPERATURE	SIM SERIAL NUMBER	SIM CERTIFICATE EXPIRATION
Air Blank	0.000	17:29	N/A	N/A	N/A
Cal Check	0.083	17:29	STANDARD VALUE	STANDARD SUPPLIER	
Air Blank	0.000	17:29	0.080	CMI	
Cal Check	0.083	17:30	CALIBRATION CHECK RESULT 1		
Air Blank	0.000	17:30	0.083		
			CALIBRATION CHECK RESULT 2		
			0.083		
			CALIBRATION CHECK RESULT 3		
			0.083		
			MAXIMUM DEVIATION (MUST BE WITHIN 5%)	SPREAD (MUST BE .005 OR LESS)	
			3.7%	0.000	

Pass

DIAGNOSTIC TEST RESULTS			RFI TEST RESULTS		
			Test	g/210L	Time
Voltage/Current Test	Pass		Air Blank	0.000	17:31
RAM Test	Pass		Subject Test	RFI*	17:31
EEPROM Checksum Test	Pass		Air Blank	0.000	17:31
Real Time Clock Test	Pass		*RFI Detect		
DSP Test	Pass				
Analytical Stability Test	Pass				
Modem Test	Pass				
Temperature Regulation Test	Pass				

Pass

NUMBER OF REFUSALS AND SUBJECT BREATH TESTS IN EACH RANGE SINCE LAST MAINTENANCE REPORT							
REFUSALS	.00-.04	.05-.09	.10-.14	.15-.19	OVER .19		
0	0	0	0	0	0	0	0

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

TESTED AND CERTIFIED

INSPECTING OFFICER		
SIGNATURE	PRINT NAME	
	JEREMY WHITE	
TYPE II PERMIT NUMBER	EXPIRATION DATE	TELEPHONE NUMBER
200231	08/20/2022	8164828141



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM



PERMIT
TYPE II

JEREMY A. WHITE

I hereby authorize to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER, INTOXILYZER 8000

or the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 8/20/2020

NUMBER 200231

EXPIRES 8/20/2022

3 280-071 (6-10)

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

[Signature]

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES
Lab-1 (pre-10)

STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM
INSTRUMENT OPERATOR CARD

The instrument operator is authorized to operate the instrument and to collect and analyze samples for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator: **WHITE, JEREMY**
Permit No. **200231**
Date Issued **8/20/2020** Date Expires **8/20/2022**



7 Eastgate Dr. • P.O. Box 790 • Jacksonville, IL 62651-0790
217-245-2183 • Fax: 217-243-7634 • www.ilmproducts.com

Certificate of Analysis

Certificate ID: 13021
Part #: BAC105L080T
Cylinder Size: 105L
Lot Number: 14020080A2
Expiration: 7/5/2022

0.080 BAC (For the calibration of instruments used to determine breath alcohol concentration)

Contents: 105 Liters @ 1000 psig 70°F (21°C)

Component	Reported Concentration:	Accuracy (U, k=2):	Analytical Method:
Ethanol	208 ppm	±0.002 BAC(g/230L)	NDIR
Nitrogen	Balance	±1.2 ppm	

Distributed by:
CM1 Inc
316 East Ninth Street
Owensboro, KY 42303
Phone 866-835-0690
www.alcoholcert.com

*Traceable to:
Certified Reference Material - 2624 μmol/mol
Ethanol in Nitrogen - Serial No. GN0015026 Lot No. 050319E11

Store in dry area, away from sources of heat, ignition and direct sunlight. Do not allow storage area to exceed 52 °C (125 °F).

[Signature]
Specialty Gas Lab Tech

06-15-2020
Issuance Date



The calibration results apply only to the specific instrument used for the analysis. The information is provided for informational purposes only. The information is not to be used as a substitute for the manufacturer's instructions. The information is not to be used as a substitute for the manufacturer's instructions.

ISO/IEC 17025:2017 Accredited Laboratory