

RECEIVED

By Tracy Crews at 1:41 pm, Oct 14, 2021

MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY**CMI INTOXILYZER 8000 MAINTENANCE REPORT**

Bat Cave

REPORT #2

Complete this report in duplicate at the time of the regular monthly preventive maintenance check, and whenever instrument is repaired. Send one copy to Department of Health and Senior Services, and retain one copy in department file.

INSTRUMENT SERIAL NUMBER: 80-005840	LOCATION OF INSTRUMENT KCMO POLICE DEPT.	DATE OF INSPECTION 10/08/2021	TIME OF INSPECTION 23:37
--	---	----------------------------------	-----------------------------

CALIBRATION CHECK RESULTS

Test	g/210L	Time
Air Blank	0.000	23:38
Cal Check	0.082	23:38
Air Blank	0.000	23:39
Cal Check	0.081	23:39
Air Blank	0.000	23:39
Cal Check	0.081	23:40
Air Blank	0.000	23:40

Pass

CALIBRATION CHECK SUMMARY

STANDARD TYPE DRY	STANDARD LOT # 14020080A2	STANDARD EXPIRATION DATE 07/05/2022
SIM TEMPERATURE N/A	SIM SERIAL NUMBER N/A	SIM CERTIFICATE EXPIRATION N/A
STANDARD VALUE 0.080	STANDARD SUPPLIER CMI	
CALIBRATION CHECK RESULT 1 0.082		
CALIBRATION CHECK RESULT 2 0.081		
CALIBRATION CHECK RESULT 3 0.081		
MAXIMUM DEVIATION (MUST BE WITHIN 5%) 2.5%		SPREAD (MUST BE .005 OR LESS) 0.001

DIAGNOSTIC TEST RESULTS

Voltage/Current Test	Pass
RAM Test	Pass
EEPROM Checksum Test	Pass
Real Time Clock Test	Pass
DSP Test	Pass
Analytical Stability Test	Pass
Modem Test	Pass
Temperature Regulation Test	Pass

Pass

RFI TEST RESULTS

Test	g/210L	Time
Air Blank	0.000	23:41
Subject Test	RFI*	23:41
Air Blank	0.000	23:41
*RFI Detect		

Pass

NUMBER OF REFUSALS AND SUBJECT BREATH TESTS IN EACH RANGE SINCE LAST MAINTENANCE REPORT

REFUSALS	.00-.04	.05-.09	.10-.14	.15-.19	OVER .19
0	28	0	0	0	0

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

TESTED AND CERTIFIED.

INSPECTING OFFICER

SIGNATURE P.G. SL R 5396	PRINT NAME SHAWN DAVIS	
TYPE II PERMIT NUMBER 210059	EXPIRATION DATE 04/06/2023	TELEPHONE NUMBER 8162345000



7 Eastgate Dr. • P.O. Box 790 • Jacksonville, IL 62651-0790
217-245-2183 • Fax: 217-243-7634 • www.ilmoproducts.com

Certificate of Analysis

Certificate ID: 13021
Part #: BAC165L080T
Cylinder Size: 105L
Lot Number: 14820080A2
Expiration: 7/5/2022

0.080 BAC (For the calibration of instruments used to determine breath alcohol concentration)

Component	Reported Concentration	Accuracy (U, K=?)	Analytical Method
Ethanol	288 ppm	±0.893 BAC(7246)	NDIR
Nitrogen	Balance	±1.8 ppm	

Distributed by:
CMI Inc.
316 East Ninth Street
Owensboro, KY 42303
Phone 866-835-0690
www.alcoholtest.com

Traceable to:
Certified Reference Material - 2624 umol/mol
Ethanol in Nitrogen - Serial No. GM0015026 Lot No. 050319E11

Store in dry area, away from sources of heat,
ignition and direct sunlight. Do not allow storage
area to exceed 52 °C (125 °F).

[Signature]
Spectrolyt Gas Lab Tech

06-15-2020
Issuance Date



Calibration and Testing
Accreditation 61385

ISO/IEC 17025:2017 Accredited Laboratory



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

PERMIT

TYPE II

SHAWN E. DAVIS

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs,
and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER, INTOXILYZER 8000

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections
577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 4/6/2021
NUMBER 210059
EXPIRES 4/6/2023

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY
[Signature]
DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 96-0771 (6-10)

STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM
INSTRUMENT OPERATOR CARD

The instrument operator is authorized to operate the specified breath alcohol
analyzer for the administration of the machine, conduct its breath test or analyze air
in Missouri.

Operator: **DAVIS SHAWN**
ID Number: **210059**
Date Issued: **4/6/2021**
Date Expires: **4/6/2023**