



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY
CMI INTOXILYZER 8000 MAINTENANCE REPORT

REPORT #2

Complete this report in duplicate at the time of the regular monthly preventive maintenance check, and whenever instrument is repaired. Send one copy to Department of Health and Senior Services, and retain one copy in department file.

INSTRUMENT SERIAL NUMBER 80-005841	LOCATION OF INSTRUMENT KCMO POLICE DEPT.	DATE OF INSPECTION 12/11/2021	TIME OF INSPECTION 17:06
---------------------------------------	---	----------------------------------	-----------------------------

CALIBRATION CHECK RESULTS			CALIBRATION CHECK SUMMARY		
Test	g/210L	Time	STANDARD TYPE	STANDARD LOT #	STANDARD EXPIRATION DATE
Air Blank	0.000	17:08	DRY	14020080A2	07/05/2022
Cal Check	0.081	17:08	SIM TEMPERATURE	SIM SERIAL NUMBER	SIM CERTIFICATE EXPIRATION
Air Blank	0.000	17:09	N/A	N/A	N/A
Cal Check	0.081	17:09	STANDARD VALUE	STANDARD SUPPLIER	
Air Blank	0.000	17:09	0.080	CMI	
Cal Check	0.081	17:10	CALIBRATION CHECK RESULT 1		
Air Blank	0.000	17:10	0.081		
Cal Check	0.081	17:10	CALIBRATION CHECK RESULT 2		
Air Blank	0.000	17:10	0.081		
Pass			CALIBRATION CHECK RESULT 3		
			0.081		
			MAXIMUM DEVIATION (MUST BE WITHIN 5%)	SPREAD (MUST BE .005 OR LESS)	
			1.2%	0.000	

DIAGNOSTIC TEST RESULTS			RFI TEST RESULTS		
			Test	g/210L	Time
Voltage/Current Test	Pass		Air Blank	0.000	17:11
RAM Test	Pass		Subject Test	RFI*	17:11
EEPROM Checksum Test	Pass		Air Blank	0.000	17:12
Real Time Clock Test	Pass		*RFI Detect		
DSP Test	Pass				
Analytical Stability Test	Pass				
Modem Test	Pass				
Temperature Regulation Test	Pass				
Pass			Pass		

NUMBER OF REFUSALS AND SUBJECT BREATH TESTS IN EACH RANGE SINCE LAST MAINTENANCE REPORT

REFUSALS	.00-.04	.05-.09	.10-.14	.15-.19	OVER .19
0	0	0	0	0	0

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

TESTED AND CERTIFIED

INSPECTING OFFICER

SIGNATURE 	PRINT NAME JEREMY WHITE
---------------	----------------------------

TYPE II PERMIT NUMBER 200231	EXPIRATION DATE 08/20/2022	TELEPHONE NUMBER 8164828141
---------------------------------	-------------------------------	--------------------------------



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM



**PERMIT
TYPE II**

JEREMY A. WHITE

hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER, INTOXILYZER 8000

the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 7.020 through 577.041, RSMo and 306.111 through 306.118 RSMo.

VTE 8/20/2020

NUMBER 200231

PIRES 8/20/2022

Jeremy A. White
DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

[Signature]
DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

380-0771 (6-18) LAB-4 (PE-10)

STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named operator is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content of breath form of expired air in Missouri.

Operator: **WHITE, JEREMY**
Permit No: **200231**
Date Issued: **8/20/2020** Date Expires: **8/20/2022**



7 Eastgate Dr. • P.O. Box 790 • Jackson, IL 62651-0790
217-245-2183 • Fax: 217-243-7634 • www.ilmo-products.com

Certificate of Analysis

Certificate ID: 13021

Part #: BAC105L080T

Cylinder Size: 105L

Lot Number: 14020080A2

Expiration: 7/5/2022

0.080 BAC (For the calibration of instruments used to determine breath alcohol concentration)

Contents: 105 Liters @ 1000 psig 70°F (21°C)

Analytical

Reported Accuracy (U, V=2): Method: NOIR

Component: Ethanol Concentration: 208 ppm Balance

Mitrogen

Distributed by:

CMi Inc.
316 East Ninth Street
Owensboro, KY 42303
Phone 866-835-0650
www.cmi-tech.com

*Traceable to:

Certified Reference Material - 2624 μmol/mol

Ethanol in Nitrogen - Serial No. GIN0015026 Lot No. 050319E11

Store in dry area, away from sources of heat, ignition and direct sunlight. Do not allow storage area to exceed 52 °C (125 °F).

[Signature]
Specialty Gas Lab Tech

06-15-2020
Issuance Date



The calibration certificate with this certificate were obtained using equipment and standards capable of producing analytical results traceable to NIST, and apply only to the items listed. The information is provided for your information as to the suitability of the use of any information provided for any particular purpose. The information is not intended to be used as a replacement for the manufacturer's instructions or to establish replacement cost of the material or service.

ISO/IEC 17025:2017 Accredited Laboratory