



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY
CMI INTOXILYZER 8000 MAINTENANCE REPORT

REPORT #2

Complete this report in duplicate at the time of the regular monthly preventive maintenance check, and whenever instrument is repaired. Send one copy to Department of Health and Senior Services, and retain one copy in department file.

INSTRUMENT SERIAL NUMBER 80-005842	LOCATION OF INSTRUMENT KANSAS CITY POLICE	DATE OF INSPECTION 02/05/2021	TIME OF INSPECTION 20:28
--	---	---	------------------------------------

CALIBRATION CHECK RESULTS			CALIBRATION CHECK SUMMARY		
Test	g/210L	Time	STANDARD TYPE DRY	STANDARD LOT # 14020080A2	STANDARD EXPIRATION DATE 07/05/2022
Air Blank	0.000	20:30	SIM TEMPERATURE N/A	SIM SERIAL NUMBER N/A	SIM CERTIFICATE EXPIRATION N/A
Cal Check	0.081	20:31	STANDARD VALUE 0.080	STANDARD SUPPLIER CMI	
Air Blank	0.000	20:31	CALIBRATION CHECK RESULT 1 0.081		
Cal Check	0.082	20:32	CALIBRATION CHECK RESULT 2 0.082		
Air Blank	0.000	20:32	CALIBRATION CHECK RESULT 3 0.082		
Cal Check	0.082	20:32	MAXIMUM DEVIATION (MUST BE WITHIN 5%) 2.5%		
Air Blank	0.000	20:33	SPREAD (MUST BE .005 OR LESS) 0.001		

Pass

DIAGNOSTIC TEST RESULTS			RFI TEST RESULTS		
Voltage/Current Test	Pass		Test	g/210L	Time
RAM Test	Pass		Air Blank	0.000	20:33
EEPROM Checksum Test	Pass		Subject Test	RFI*	20:34
Real Time Clock Test	Pass		Air Blank	0.000	20:34
DSP Test	Pass		*RFI Detect		
Analytical Stability Test	Pass				
Modem Test	Pass				
Temperature Regulation Test	Pass				

Pass

NUMBER OF REFUSALS AND SUBJECT BREATH TESTS IN EACH RANGE SINCE LAST MAINTENANCE REPORT						
REFUSALS	.00-.04	.05-.09	.10-.14	.15-.19	OVER .19	
0	1	1	2	1	0	

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

TESTED AND CERTIFIED

INSPECTING OFFICER

SIGNATURE 	PRINT NAME JEREMY WHITE
TYPE II PERMIT NUMBER 200231	EXPIRATION DATE 08/20/2022
TELEPHONE NUMBER 8164828141	



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM



PERMIT
TYPE II

JEREMY A. WHITE

I hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER, INTOXILYZER 8000

or the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 8/20/2020

NUMBER 200231

EXPIRES 8/20/2022

0 980-0771 (8-10)

Jeremy A. White
DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES
LMS-4 (08-10)

STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM
INSTRUMENT OPERATOR CARD

This permit authorizes the operator to operate an Intoxilizer breath analyzer instrument for the detection and/or estimation of the alcoholic content of expired air in Missouri.

Operator: **WHITE, JEREMY**
Permit No: **200231**
Date Issued: **8/20/2020** Date Expires: **8/20/2022**



7 Eastgate Dr. • P.O. Box 790 • Jacksonville, IL 62651-0790
217-245-2183 • Fax: 217-243-7634 • www.ilmoproducts.com

Certificate of Analysis

Certificate ID: 13021
Part #: BAC105L080T
Cylinder Size: 105L
Lot Number: 1A020808A2
Expiration: 7/5/2022

0.080 BAC (For the calibration of instruments used to determine breath alcohol concentration)

Contents: 105 Liters @ 1000 psig 70°F (21°C)

Component	Reported Concentration:	Accuracy (U, k=2):	Analytical Method:
Ethanol	288 ppm	±1.6 ppm BAC(728L)	NDIR
Nitrogen	Balance	±5.2 ppm	

Distributed by:
CPI Inc
316 East Ninth Street
Owensboro, KY 42303
Phone 866-835-0690
www.alcoholtest.com

*Traceable to:
Certified Reference Material - 262.4 µmol/mol
Ethanol in Nitrogen - Serial No. GND015026 Lot No. 050319E11

Score in dry area, away from sources of heat, ignition and direct sunlight. Do not allow storage area to exceed 52 °C (125 °F).

Specialty Gas Lab Tech
Jeremy A. White

08-15-2020
Issuance Date



The calibration made within this certificate was obtained using equipment and standards capable of producing analytical results traceable to NIST and apply only to the items considered on this certificate. ILMO Products Company makes no warranty or representation as to the suitability of the use of any information provided for any particular purpose. The information use is at the sole discretion and risk of the user. Liability shall be limited to established replacement cost of this material or service.

ISO/IEC 17025:2017 Accredited Laboratory