



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY
CMI INTOXILYZER 8000 MAINTENANCE REPORT

REPORT #2

Complete this report in duplicate at the time of the regular monthly preventive maintenance check, and whenever instrument is repaired. Send one copy to Department of Health and Senior Services, and retain one copy in department file.

INSTRUMENT SERIAL NUMBER 80-005844	LOCATION OF INSTRUMENT KCMO POLICE DEPT.	DATE OF INSPECTION 01/06/2021	TIME OF INSPECTION 17:32
---------------------------------------	---	----------------------------------	-----------------------------

CALIBRATION CHECK RESULTS			CALIBRATION CHECK SUMMARY		
Test	g/210L	Time	STANDARD TYPE	STANDARD LOT #	STANDARD EXPIRATION DATE
-----	-----	-----	DRY	14020080A2	07/05/2022
Air Blank	0.000	17:34	SIM TEMPERATURE	SIM SERIAL NUMBER	SIM CERTIFICATE EXPIRATION
Cal Check	0.080	17:35	N/A	N/A	N/A
Air Blank	0.000	17:35	STANDARD VALUE	STANDARD SUPPLIER	
Cal Check	0.081	17:35	0.080	CMI	01402008
Air Blank	0.000	17:36	CALIBRATION CHECK RESULT 1		
Cal Check	0.080	17:36	0.080		
Air Blank	0.000	17:36	CALIBRATION CHECK RESULT 2		
Cal Check	0.080	17:36	0.081		
Air Blank	0.000	17:37	CALIBRATION CHECK RESULT 3		
Pass			MAXIMUM DEVIATION (MUST BE WITHIN 5%)		SPREAD (MUST BE .005 OR LESS)
			1.2%		0.001

DIAGNOSTIC TEST RESULTS			RFI TEST RESULTS		
Test	Result	Time	Test	g/210L	Time
Voltage/Current Test	Pass		-----	-----	-----
RAM Test	Pass		Air Blank	RFI*	17:37
EEPROM Checksum Test	Pass		Air Blank	0.000	17:37
Real Time Clock Test	Pass		*RFI Detect		
DSP Test	Pass		Pass		
Analytical Stability Test	Pass				
Modem Test	Pass				
Temperature Regulation Test	Pass				

NUMBER OF REFUSALS AND SUBJECT BREATH TESTS IN EACH RANGE SINCE LAST MAINTENANCE REPORT							
REFUSALS	.00-.04	.05-.09	.10-.14	.15-.19	OVER .19		
0	0	0	0	0	0		0

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

TESTED AND CERTIFIED

INSPECTING OFFICER			
SIGNATURE <i>P.O. J. Infranca</i>		PRINT NAME JORDAN INFRANCA	
TYPE II PERMIT NUMBER 290130	EXPIRATION DATE 06/21/2021	TELEPHONE NUMBER 81-382-5897	



7 Eastgate Dr. • P.O. Box 790 • Jacksonville, IL 62651-0790
217-245-2183 • Fax: 217-243-7634 • www.ilmoproducts.com

Certificate of Analysis

Certificate ID: 13021
Part #: BAC105L080T
Cylinder Size: 105L
Lot Number: 14020080A2
Expiration: 7/5/2022

0.080 BAC (For the calibration of instruments used to determine breath alcohol concentration)

Contents: 105 Liters @ 1000 psig 70°F (21°C)

Component:	Reported Concentration:	Analytical Accuracy (U, k=2):	Analytical Method:
Ethanol	208 ppm	+/-0.092 BAC(G/210L) [5.2 ppm]	NDIR
Nitrogen	Balance		

Distributed by:
CMI Inc.
316 East Ninth Street
Owensboro, KY 42303
Phone 866-835-0690
www.alcoholtest.com

*Traceable to:
Certified Reference Material - 262.4 µmol/mol
Ethanol in Nitrogen - Serial No. GN0015026 Lot No. 050319E11

Store in dry area, away from sources of heat, ignition and direct sunlight. Do not allow storage area to exceed 52 °C (125 °F).

Specialty Gas Lab Tech

06-15-2020
Issuance Date



The calibration results within this certificate were obtained using equipment and standards capable of producing analytical results traceable to NIST, and apply only to the items contained on this certificate. ILMO Products Company makes no warranty or representation as to the suitability of the use of any information provided for any particular purpose. The information use is at the sole discretion and risk of the user. Liability shall be limited to established replacement cost of this material or service.

ISO/IEC 17025:2017 Accredited Laboratory



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM



PERMIT TYPE II

JORDAN INFRANCA

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER, INTOXILYZER 8000

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 6/21/2019

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

NUMBER 290130

EXPIRES 6/21/2021

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES
LAB-4 (PS-10)

MO 580-0771 (6-10)

STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator: INFRANCA, JORDAN
Permit No: 290130
Date Issued: 6/21/2019 Date Expires: 6/21/2021