



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY
CMi INTOXILYZER 8000 MAINTENANCE REPORT

REPORT #2

Complete this report in duplicate at the time of the regular monthly preventive maintenance check, and whenever instrument is repaired. Send one copy to Department of Health and Senior Services, and retain one copy in department file.

INSTRUMENT SERIAL NUMBER 80-005844	LOCATION OF INSTRUMENT KCMO POLICE DEPT.	DATE OF INSPECTION 02/05/2021	TIME OF INSPECTION 16:40
--	--	---	------------------------------------

CALIBRATION CHECK RESULTS			CALIBRATION CHECK SUMMARY		
Test	g/210L	Time	STANDARD TYPE	STANDARD LOT #	STANDARD EXPIRATION DATE
Air Blank	0.000	16:42	DRY	14020080A2	07/05/2022
Cal Check	0.081	16:43	SIM TEMPERATURE	SIM SERIAL NUMBER	SIM CERTIFICATE EXPIRATION
Air Blank	0.000	16:43	N/A	N/A	N/A
Cal Check	0.081	16:43	STANDARD VALUE	STANDARD SUPPLIER	01402008
Air Blank	0.000	16:44	0.080	CMI	
Cal Check	0.081	16:44	CALIBRATION CHECK RESULT 1		
Air Blank	0.000	16:44	0.081		
Cal Check	0.081	16:44	CALIBRATION CHECK RESULT 2		
Air Blank	0.000	16:45	0.081		
Pass			CALIBRATION CHECK RESULT 3		
			0.081		
			MAXIMUM DEVIATION (MUST BE WITHIN 5%)	SPREAD (MUST BE .005 OR LESS)	
			1.2%	0.000	

0A2

DIAGNOSTIC TEST RESULTS			RFI TEST RESULTS		
			Test	g/210L	Time
Voltage/Current Test	Pass		Air Blank	0.000	16:45
RAM Test	Pass		Subject Test	RFI*	16:46
EEPROM Checksum Test	Pass		Air Blank	0.000	16:46
Real Time Clock Test	Pass		*RFI Detect		
DSP Test	Pass				
Analytical Stability Test	Pass				
Modem Test	Pass				
Temperature Regulation Test	Pass				
Pass			Pass		

NUMBER OF REFUSALS AND SUBJECT BREATH TESTS IN EACH RANGE SINCE LAST MAINTENANCE REPORT						
REFUSALS	.00-.04	.05-.09	.10-.14	.15-.19	OVER .19	
0	0	0	0	0	0	0

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).
TESTED AND CERTIFIED

INSPECTING OFFICER		
SIGNATURE	PRINT NAME	
	JEREMY WHITE	
TYPE II PERMIT NUMBER	EXPIRATION DATE	TELEPHONE NUMBER
200231	08/20/2022	8164828141



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM



**PERMIT
TYPE II**

JEREMY A. WHITE

I am hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER, INTOXILYZER 8000

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.041, RSMo and 506.111 through 506.119 RSMo.

DATE 8/20/2020

NUMBER 200231

EXPIRES 8/20/2022

SP-0771 (6-10)

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

LAB-4 (PE-10)

STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named operator is authorized to operate an alcohol breathalyzer instrument for the determination of the alcoholic content in breath form of expired air in Missouri.



Operator WHITE, JEREMY
Permit No 200231
Date Issued 8/20/2020 Date Expires 8/20/2022



7 Eastgate Dr. • P.O. Box 790 • Jacksonville, IL 62651-0790
217-245-2183 • Fax: 217-243-7634 • www.ilmo-products.com

Certificate of Analysis

Certificate ID: 13021
Part #: BAC10510880T
Cylinder Size: 105L
Lot Number: 14020080A2
Expiration: 7/5/2022

0.080 BAC (For the calibration of instruments used to determine breath alcohol concentration)

Contents: 105 Liters @ 1000 psig 70°F (21°C)
Analytical Accuracy (U, k=2): ±0.002 BAC (67.20μL) (5.2 ppm)
Reported Concentration: 288 ppm
Component: Ethanol Nitrogen
Analytical Method: NDIR
Balance

Distributed by:
CMI Inc
316 East Ninth Street
Owensboro, KY 42303
Phone 866-835-0690
www.alcoholtest.com

*Traceable to:
Certified Reference Material - 262.4 μmol/mol
Ethanol in Nitrogen - Serial No. GN0015026 Lot No. 050319E11

Store in dry area, away from sources of heat, ignition and direct sunlight. Do not allow storage area to exceed 52 °C (125 °F).

SpeciLab Gas Lab Tech
06/15/2020
Issuance Date



The calibration results within this certificate were obtained using equipment and standards capable of producing analytical results traceable to NIST, and apply only to the items identified in this certificate. The information is provided for informational purposes only and is not intended for legal or regulatory purposes. The information use is at the sole discretion and risk of the user. Liability shall be limited to established replacement costs of this material or service.

ISO/IEC 17025:2017 Accredited Laboratory