



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY
CMI INTOXILYZER 8000 MAINTENANCE REPORT

REPORT #2

Complete this report in duplicate at the time of the regular monthly preventive maintenance check, and whenever instrument is repaired. Send one copy to Department of Health and Senior Services, and retain one copy in department file.

INSTRUMENT SERIAL NUMBER 80-005844	LOCATION OF INSTRUMENT KCMO POLICE DEPT.	DATE OF INSPECTION 08/07/2021	TIME OF INSPECTION 19:38
---------------------------------------	---	----------------------------------	-----------------------------

CALIBRATION CHECK RESULTS			CALIBRATION CHECK SUMMARY		
Test	g/210L	Time	STANDARD TYPE	STANDARD LOT #	STANDARD EXPIRATION DATE
Air Blank	0.000	19:39	DRY	14020080A2	07/05/2022
Cal Check	0.078	19:40	SIM TEMPERATURE	SIM SERIAL NUMBER	SIM CERTIFICATE EXPIRATION
Air Blank	0.000	19:40	N/A	N/A	N/A
Cal Check	0.078	19:40	STANDARD VALUE	STANDARD SUPPLIER	
Air Blank	0.000	19:41	0.080	CMI	01402008
Cal Check	0.079	19:41	CALIBRATION CHECK RESULT 1		
Air Blank	0.000	19:42	0.078		
Pass			CALIBRATION CHECK RESULT 2		
			0.078		
			CALIBRATION CHECK RESULT 3		
			0.079		
			MAXIMUM DEVIATION (MUST BE WITHIN 5%)	SPREAD (MUST BE .005 OR LESS)	
			2.5%	0.001	

DIAGNOSTIC TEST RESULTS			RFI TEST RESULTS		
Test	Pass/Fail		Test	g/210L	Time
Voltage/Current Test	Pass		Air Blank	0.000	19:42
RAM Test	Pass		Subject Test	RFI*	19:43
EEPROM Checksum Test	Pass		Air Blank	0.000	19:43
Real Time Clock Test	Pass		*RFI Detect		
DSP Test	Pass				
Analytical Stability Test	Pass				
Modem Test	Pass				
Temperature Regulation Test	Pass				
Pass			Pass		

NUMBER OF REFUSALS AND SUBJECT BREATH TESTS IN EACH RANGE SINCE LAST MAINTENANCE REPORT							
REFUSALS	.00-.04	.05-.09	.10-.14	.15-.19	OVER .19		
0	0	0	0	0	0	0	

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).
TESTED AND CERTIFIED

INSPECTING OFFICER			
SIGNATURE 		PRINT NAME NATHAN MAGERS	
TYPE II PERMIT NUMBER 210105	EXPIRATION DATE 05/18/2023	TELEPHONE NUMBER 8162345000	

