

RECEIVED

By Tracy Crews at 10:57 am, Sep 13, 2021



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY
CMI INTOXILYZER 8000 MAINTENANCE REPORT

Batcave

REPORT #2

Complete this report in duplicate at the time of the regular monthly preventive maintenance check, and whenever instrument is repaired. Send one copy to Department of Health and Senior Services, and retain one copy in department file.

INSTRUMENT SERIAL NUMBER 80-005844	LOCATION OF INSTRUMENT KCMO POLICE DEPT.	DATE OF INSPECTION 09/06/2021	TIME OF INSPECTION 23:53
---------------------------------------	---	----------------------------------	-----------------------------

CALIBRATION CHECK RESULTS			CALIBRATION CHECK SUMMARY		
Test	g/210L	Time	STANDARD TYPE	STANDARD LOT #	STANDARD EXPIRATION DATE
Air Blank	0.000	23:55	DRY	14020080A2	07/05/2022
Cal Check	0.079	23:55	SIM TEMPERATURE	SIM SERIAL NUMBER	SIM CERTIFICATE EXPIRATION
Air Blank	0.000	23:55	N/A	N/A	N/A
Cal Check	0.080	23:56	STANDARD VALUE	STANDARD SUPPLIER	
Air Blank	0.000	23:56	0.080	CMI 01402008	
Cal Check	0.079	23:57	CALIBRATION CHECK RESULT 1		
Air Blank	0.000	23:57	0.079		
Cal Check	0.079	23:57	CALIBRATION CHECK RESULT 2		
Air Blank	0.000	23:57	0.080		
			CALIBRATION CHECK RESULT 3		
			0.079		
			MAXIMUM DEVIATION (MUST BE WITHIN 5%)	SPREAD (MUST BE .005 OR LESS)	
			1.2%	0.001	

Pass

DIAGNOSTIC TEST RESULTS		RFI TEST RESULTS	
Test	Result	Test	Time
Voltage/Current Test	Pass	Air Blank	23:57
RAM Test	Pass	Air Blank	23:58
EEPROM Checksum Test	Pass	*RFI Detect	
Real Time Clock Test	Pass		
DSP Test	Pass		
Analytical Stability Test	Pass		
Modem Test	Pass		
Temperature Regulation Test	Pass		

Pass**Pass****NUMBER OF REFUSALS AND SUBJECT BREATH TESTS IN EACH RANGE SINCE LAST MAINTENANCE REPORT**

REFUSALS	.00-.04	.05-.09	.10-.14	.15-.19	OVER .19
0	0	0	0	0	0

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

TESTED AND CERTIFIED

INSPECTING OFFICER

SIGNATURE 	PRINT NAME INFRANCA, JORDAN	
TYPE II PERMIT NUMBER 210128	EXPIRATION DATE 06/21/2023	TELEPHONE NUMBER 816-382-5897



7 Eastgate Dr. • P.O. Box 790 • Jackson, IL 62651-0790
217-245-2183 • Fax: 217-243-7634 • www.ilmo-products.com

Certificate of Analysis

Certificate ID: 13821
Part #: BAC105L080T
Cylinder Size: 105L
Lot Number: 140208080A2
Expiration: 7/5/2022

080 BAC (For the calibration of instruments used to determine breath alcohol concentration)

Contents: 105 Liters @ 1000 psig 70°F (21°C)
Reported Analytical Accuracy Analytical
Concentration: (U, K=2): Method:
Nitrogen 288 ppm +/- 0.002 BAC(6/20A1) NDIR
Balance (5.2 ppm)

Distributed by:
CMI Inc.
316 East Ninth Street
Owensboro, KY 42303
Phone 866-835-0690
www.alcoholtest.com

Traceable to:
Certified Reference Material - 2624 µmol/mol
Nitrogen in Nitrogen - Serial No. GN0015026 Lot No. 050319E11

Store in dry area, away from sources of heat,
ignition and direct sunlight. Do not allow storage
area to exceed 52 °C (125 °F).

[Signature]
Quality Gas Lab Tech

06-15-2020
Issuance Date



ISO/IEC 17025:2017 Accredited Laboratory

Information: This certificate was obtained using equipment that is inherently capable of producing analytical results traceable to NIST, and apply only to the material stated on this certificate. ILMO Products Company makes no warranty or representation as to the suitability of the use of any information provided for any particular purpose. The information is for the information of the user. Liability shall be limited to the maximum extent permitted by law.



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

PERMIT TYPE II

JORDAN INFRANCA

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER, INTOXILYZER 8000

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE: 6/21/2019
NUMBER: 290130
EXPIRES: 6/21/2021
MO 560-071 (6-10)

[Signature]
DIRECTOR OF STATE PUBLIC HEALTH LABORATORY
DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES
LH&S (PS-10)

STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM
INSTRUMENT OPERATOR CARD

This permit cardholder is authorized to operate an Intoxilizer breathalyzer instrument for the determination of the alcoholic content in breath from an expired air in Missouri.

Operator: INFRANCA, JORDAN
Permit No: 290130
Date Issued: 6/21/2019
Date Expires: 6/21/2021