



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY
CMI INTOXILYZER 8000 MAINTENANCE REPORT

REPORT #2

Complete this report in duplicate at the time of the regular monthly preventive maintenance check, and whenever instrument is repaired. Send one copy to Department of Health and Senior Services, and retain one copy in department file.

INSTRUMENT SERIAL NUMBER 80-005845	LOCATION OF INSTRUMENT KANSAS CITY PD	DATE OF INSPECTION 02/05/2021	TIME OF INSPECTION 16:24
---------------------------------------	--	----------------------------------	-----------------------------

CALIBRATION CHECK RESULTS			CALIBRATION CHECK SUMMARY		
Test	g/210L	Time	STANDARD TYPE	STANDARD LOT #	STANDARD EXPIRATION DATE
Air Blank	0.000	16:27	DRY	14020080A2	07/05/2022
Cal Check	0.080	16:27	SIM TEMPERATURE	SIM SERIAL NUMBER	SIM CERTIFICATE EXPIRATION
Air Blank	0.000	16:28	N/A	N/A	N/A
Cal Check	0.082	16:28	STANDARD VALUE	STANDARD SUPPLIER	
Air Blank	0.000	16:29	0.080	CMI	
Cal Check	0.083	16:29	CALIBRATION CHECK RESULT 1		
Air Blank	0.000	16:29	0.080		
Pass			CALIBRATION CHECK RESULT 2		
			0.082		
			CALIBRATION CHECK RESULT 3		
			0.083		
			MAXIMUM DEVIATION (MUST BE WITHIN 5%)	SPREAD (MUST BE .005 OR LESS)	
			3.7%	0.003	

DIAGNOSTIC TEST RESULTS		RFI TEST RESULTS		
		Test	g/210L	Time
Voltage/Current Test	Pass	Air Blank	0.000	16:30
RAM Test	Pass	Subject Test	RFI*	16:30
EEPROM Checksum Test	Pass	Air Blank	0.000	16:31
Real Time Clock Test	Pass	*RFI Detect		
DSP Test	Pass			
Analytical Stability Test	Pass			
Modem Test	Pass			
Temperature Regulation Test	Pass			
Pass		Pass		

NUMBER OF REFUSALS AND SUBJECT BREATH TESTS IN EACH RANGE SINCE LAST MAINTENANCE REPORT							
REFUSALS	.00-.04	.05-.09	.10-.14	.15-.19	OVER .19		
0	2	1	3	0	6		

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

TESTED AND CERTIFIED

INSPECTING OFFICER

SIGNATURE 	PRINT NAME JEREMY WHITE
TYPE II PERMIT NUMBER 200231	EXPIRATION DATE 08/20/2022
TELEPHONE NUMBER 8164828141	



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

**PERMIT
TYPE II**

JEREMY A. WHITE



7 Eastgate Dr. • P.O. Box 790 • Jacksonvillle, IL 62651-0790
217-245-2183 • Fax 217-245-7634 • www.ilmo.com



Certificate of Analysis

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER, INTOXILYZER 8000

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 8/20/2020

NUMBER 200231

EXPIRES 8/20/2022

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

[Signature]

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 280-0771 (6-19)

LAB 698-10

STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The instrument operator is authorized to operate an evidential breath analyzer in Missouri for the determination of the alcoholic content of expired air.

Operator: **WHITE, JEREMY**
Permit No: **200231**
Date Issued: **8/20/2020** Date Expires: **8/20/2022**

Certificate ID: 13021
Part #: BAC105L080T
Cylinder Size: 105L
Lot Number: 14020080A2
Expiration: 7/5/2022

0.080 BAC (For the calibration of instruments used to determine breath alcohol concentration)

Contents: 105 Liters @ 1000 psig 70°F (21 °C)

Component	Reported Concentration:	Analytical Accuracy (U, K=2):	Analytical Method:
Ethanol	208 ppm	+/-0.002 BAC(g/200L)	NDIR
Nitrogen	Balance	[3.2 ppm]	

Distributed by:
CMI Inc
316 East Ninth Street
Owensboro, KY 42303
Phone 866-835-0690
www.alcoholtest.com

*Traceable to:
Certified Reference Material - 2624 μmol/mol
Ethanol in Nitrogen - Serial No. GNM015026 Lot No. 050319E11

Score in dry area, away from sources of heat, ignition and direct sunlight. Do not allow storage area to exceed 52 °C (125 °F).

Specialty Gas Lab Tech
[Signature]

06-15-2020
Issuance Date



ISO/IEC 17025:2017 Accredited Laboratory

The calibration results within this certificate were obtained using equipment and standards capable of producing analytical results traceable to NIST, and apply only to the items compared on this certificate. ILMO Product Company makes no warranty or representation as to the suitability of the use of any information provided for any particular purpose. The information on this certificate is for informational purposes only and does not constitute a contract or any other legal relationship.