

RECEIVED

By Tracy Crews at 7:46 am, Jun 04, 2021



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY
CMI INTOXILYZER 8000 MAINTENANCE REPORT

SPD

REPORT #2

Complete this report in duplicate at the time of the regular monthly preventive maintenance check, and whenever instrument is repaired. Send one copy to Department of Health and Senior Services, and retain one copy in department file.

INSTRUMENT SERIAL NUMBER 80-005845	LOCATION OF INSTRUMENT KANSAS CITY PD	DATE OF INSPECTION 05/15/2021	TIME OF INSPECTION 17:05
---------------------------------------	--	----------------------------------	-----------------------------

CALIBRATION CHECK RESULTS			CALIBRATION CHECK SUMMARY		
Test	g/210L	Time	STANDARD TYPE	STANDARD LOT #	STANDARD EXPIRATION DATE
Air Blank	0.000	17:06	DRY	14020080A2	07/05/2022
Cal Check	0.081	17:06	SIM TEMPERATURE	SIM SERIAL NUMBER	SIM CERTIFICATE EXPIRATION
Air Blank	0.000	17:07	N/A	N/A	N/A
Cal Check	0.079	17:07	STANDARD VALUE	STANDARD SUPPLIER	
Air Blank	0.000	17:08	0.080	CMI	
Cal Check	0.081	17:08	CALIBRATION CHECK RESULT 1		
Air Blank	0.000	17:09	0.081		
Pass			CALIBRATION CHECK RESULT 2		
			0.079		
			CALIBRATION CHECK RESULT 3		
			0.081		
			MAXIMUM DEVIATION (MUST BE WITHIN 5%)	SPREAD (MUST BE .005 OR LESS)	
			1.2%	0.002	

DIAGNOSTIC TEST RESULTS		RFI TEST RESULTS		
Test	Pass	Test	g/210L	Time
Voltage/Current Test	Pass	Air Blank	0.000	17:09
RAM Test	Pass	Subject Test	RFI*	17:09
EEPROM Checksum Test	Pass	Air Blank	0.000	17:10
Real Time Clock Test	Pass	*RFI Detect		
DSP Test	Pass	Pass		
Analytical Stability Test	Pass			
Modem Test	Pass			
Temperature Regulation Test	Pass			

NUMBER OF REFUSALS AND SUBJECT BREATH TESTS IN EACH RANGE SINCE LAST MAINTENANCE REPORT							
REFUSALS	.00-.04	.05-.09	.10-.14	.15-.19	OVER .19		
0	1	1	3	0	2		

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).
TESTED AND CERTIFIED.

INSPECTING OFFICER		SIGNATURE		PRINT NAME	
		PO, SL R 5396		SHAWN DAVIS	
TYPE II PERMIT NUMBER		EXPIRATION DATE		TELEPHONE NUMBER	
210059		04/06/2023		816-234-5000	