

**RECEIVED**

By Tracy Crews at 8:59 am, Sep 15, 2021



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES  
STATE PUBLIC HEALTH LABORATORY  
**CMI INTOXILYZER 8000 MAINTENANCE REPORT**

REPORT #2

Complete this report in duplicate at the time of the regular monthly preventive maintenance check, and whenever instrument is repaired. Send one copy to Department of Health and Senior Services, and retain one copy in department file.

|                                       |   |                                  |                             |
|---------------------------------------|---|----------------------------------|-----------------------------|
| INSTRUMENT SERIAL NUMBER<br>80-005847 | LOCATION OF INSTRUMENT<br>IPD DETENTION | DATE OF INSPECTION<br>09/09/2021 | TIME OF INSPECTION<br>14:18 |
|---------------------------------------|---|----------------------------------|-----------------------------|

| CALIBRATION CHECK RESULTS |        |       | CALIBRATION CHECK SUMMARY             |                   |                            |
|---------------------------|--------|-------|---------------------------------------|-------------------|----------------------------|
| Test                      | g/210L | Time  | STANDARD TYPE                         | STANDARD LOT #    | STANDARD EXPIRATION DATE   |
| -----                     | -----  | ----- | DRY                                   | AG928002          | 10/07/2021                 |
| Air Blank                 | 0.000  | 14:20 | SIM TEMPERATURE                       | SIM SERIAL NUMBER | SIM CERTIFICATE EXPIRATION |
| Cal Check                 | 0.079  | 14:20 | N/A                                   | N/A               | N/A                        |
| Air Blank                 | 0.000  | 14:21 | STANDARD VALUE                        | STANDARD SUPPLIER |                            |
| Cal Check                 | 0.080  | 14:21 | 0.080                                 | INTOXIMETERS, INC |                            |
| Air Blank                 | 0.000  | 14:22 | CALIBRATION CHECK RESULT 1            |                   |                            |
| Cal Check                 | 0.080  | 14:22 | 0.079                                 |                   |                            |
| Air Blank                 | 0.000  | 14:22 | CALIBRATION CHECK RESULT 2            |                   |                            |
| Cal Check                 | 0.080  | 14:22 | 0.080                                 |                   |                            |
| Air Blank                 | 0.000  | 14:23 | CALIBRATION CHECK RESULT 3            |                   |                            |
| <b>Pass</b>               |        |       | MAXIMUM DEVIATION (MUST BE WITHIN 5%) |                   |                            |
|                           |        |       | 1.2%                                  |                   |                            |
|                           |        |       | 0.001                                 |                   |                            |

| DIAGNOSTIC TEST RESULTS     |      |  | RFI TEST RESULTS |        |       |
|-----------------------------|------|--|------------------|--------|-------|
|                             |      |  | Test             | g/210L | Time  |
| Voltage/Current Test        | Pass |  | -----            | -----  | ----- |
| RAM Test                    | Pass |  | Air Blank        | 0.000  | 14:23 |
| EEPROM Checksum Test        | Pass |  | Subject Test     | RFI*   | 14:23 |
| Real Time Clock Test        | Pass |  | Air Blank        | 0.000  | 14:24 |
| DSP Test                    | Pass |  | *RFI Detect      |        |       |
| Analytical Stability Test   | Pass |  | <b>Pass</b>      |        |       |
| Modem Test                  | Pass |  |                  |        |       |
| Temperature Regulation Test | Pass |  |                  |        |       |
|                             |      |  |                  |        |       |

| NUMBER OF REFUSALS AND SUBJECT BREATH TESTS IN EACH RANGE SINCE LAST MAINTENANCE REPORT |         |         |         |         |          |  |  |
|---|---------|---------|---------|---------|----------|--|--|
| REFUSALS  | .00-.04 | .05-.09 | .10-.14 | .15-.19 | OVER .19 |  |  |
| 2   | 1       | 1       | 2       | 1       | 4        |  |  |

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

INSTRUMENT OPERATED WITHIN STATE STANDARDS

**INSPECTING OFFICER**

|                                       |                               |                                  |
|---------------------------------------|-------------------------------|----------------------------------|
| SIGNATURE<br><i>Ron Baltzer #1460</i> | PRINT NAME<br>RON BALTZER     |                                  |
| TYPE II PERMIT NUMBER<br>210144       | EXPIRATION DATE<br>07/16/2023 | TELEPHONE NUMBER<br>816-325-7300 |



Airgas USA LLC (LAB)  
 3500 Bernard Street  
 St. Louis, Mo. 63103  
 P: (314) 533-3100  
 Fax: (314) 533-7328

## Certificate of Analysis

Customer Name  
 Exclusive Supplier  
 Intoximeters, Inc.  
 2081 Craig Road  
 St. Louis, Mo 63146

Test Date: 7-Oct-2019

**Lot # AG928002 Model 108caccd**

| <u>Exp. Date</u> | <u>Cyl. Type</u> | <u>Component</u> | <u>Certified Concentration</u> |
|------------------|------------------|------------------|--------------------------------|
| 7-Oct-2021       | 108              | Ethanol          | 0.080 ± 0.002 BrAC (208 ppm)   |
|                  |                  | Nitrogen         | Balance                        |

Certification Traceable to N.I.S.T. RGM and to CRM Ethanol Standards:

| <u>RGM Serial No.</u> | <u>Concentration</u> | <u>RGM Serial No.</u> | <u>Concentration</u> |
|-----------------------|----------------------|-----------------------|----------------------|
| EB0010581             | 392.1 ppm            | EB0010603             | 393.0 ppm            |
| EB0010570             | 259.8 ppm            | EB0010559             | 258.2 ppm            |
| EB0010285             | 208.0 ppm            | EB0010595             | 208.3 ppm            |
| EB0010561             | 103.6 ppm            | EB0010562             | 104.2 ppm            |
| EB0010681             | 52.12 ppm            | EB0010579             | 52.81 ppm            |

  

| <u>CRM Serial No.</u> | <u>Concentration</u> | <u>CRM Serial No.</u> | <u>Concentration</u> |
|-----------------------|----------------------|-----------------------|----------------------|
| CC434668              | 800.0 ppm            | 0056649               | 390.1 ppm            |
| CC234503              | 253.0 ppm            | 0056662               | 150.2 ppm            |

Analytical Method: NDIR

Digitally signed by Quality Control  
 Date: 2019.10.07 16:34:11 -05:00  
 Reason: Dry gas standard certification of analysis  
 Location: Airgas USA LLC (Lab)

Approved for Release: *Rod Marsala*  
 Rod Marsala

ISO 17025:2005 A2LA accredited. Certificate Number 3082.06  
 ISO 17034:2016 A2LA accredited. Certificate Number 3082.07



STATE OF MISSOURI  
 DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 BREATH ALCOHOL PROGRAM

2

**PERMIT**  
**TYPE II**  
**RON BALTZER**

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

**INTOXILYZER 8000**

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 7/16/2021

*Laura Q. Wag*

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

NUMBER 210144

*Ron Kuehl*

EXPIRES 7/16/2023

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (6-10)

LAB-4 (R6-10)



STATE OF MISSOURI  
 DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 BREATH ALCOHOL PROGRAM

**INSTRUMENT OPERATOR CARD**

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator BALTZER, RON  
 Permit No 210144  
 Date Issued 7/16/2021 Date Expires 7/16/2023

