

RECEIVED

By Tracy Crews at 3:36 pm, Dec 20, 2021



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY
CMI INTOXILYZER 8000 MAINTENANCE REPORT

REPORT #2

Complete this report in duplicate at the time of the regular monthly preventive maintenance check, and whenever instrument is repaired. Send one copy to Department of Health and Senior Services, and retain one copy in department file.

INSTRUMENT SERIAL NUMBER 80-006267	LOCATION OF INSTRUMENT KCPD	DATE OF INSPECTION 12/07/2021	TIME OF INSPECTION 19:13
---------------------------------------	--------------------------------	----------------------------------	-----------------------------

CALIBRATION CHECK RESULTS			CALIBRATION CHECK SUMMARY		
Test	g/210L	Time	STANDARD TYPE	STANDARD LOT #	STANDARD EXPIRATION DATE
Air Blank	0.000	19:15	DRY	14020080A2	07/05/2022
Cal Check	0.084	19:15	SIM TEMPERATURE	SIM SERIAL NUMBER	SIM CERTIFICATE EXPIRATION
Air Blank	0.000	19:16	N/A	N/A	N/A
Cal Check	0.084	19:16	STANDARD VALUE	STANDARD SUPPLIER	
Air Blank	0.000	19:17	0.080	INTOXIMETERS	
Cal Check	0.083	19:17	CALIBRATION CHECK RESULT 1		
Air Blank	0.000	19:18	0.084		
Pass			CALIBRATION CHECK RESULT 2		
			0.084		
			CALIBRATION CHECK RESULT 3		
			0.083		
			MAXIMUM DEVIATION (MUST BE WITHIN 5%)	SPREAD (MUST BE .005 OR LESS)	
			5.0%	0.001	

DIAGNOSTIC TEST RESULTS			RFI TEST RESULTS		
Test	Pass	Time	Test	g/210L	Time
Voltage/Current Test	Pass		Air Blank	0.000	19:18
RAM Test	Pass		Subject Test	RFI*	19:18
EEPROM Checksum Test	Pass		Air Blank	0.000	19:19
Real Time Clock Test	Pass		*RFI Detect		
DSP Test	Pass				
Analytical Stability Test	Pass				
Modem Test	Pass				
Temperature Regulation Test	Pass				
Pass			Pass		

NUMBER OF REFUSALS AND SUBJECT BREATH TESTS IN EACH RANGE SINCE LAST MAINTENANCE REPORT

REFUSALS	.00-.04	.05-.09	.10-.14	.15-.19	OVER .19
0	0	0	0	1	0

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

TESTED AND CERTIFIED

INSPECTING OFFICER

SIGNATURE 	PRINT NAME JEREMY WHITE
TYPE II PERMIT NUMBER 200231	EXPIRATION DATE 08/20/2022
	TELEPHONE NUMBER 8164828141



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM



PERMIT
TYPE II

JEREMY A. WHITE

hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER, INTOXILYZER 8000

- the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 7.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 8/20/2020

NUMBER 200231

PIPERES 8/20/2022

390-0771 (6-09)

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

LS-4 (86-49)



7 Eastgate Dr. • P.O. Box 790 • Jacksonville, IL 62651-0790
217-245-2183 • Fax: 217-243-7634 • www.ilmproducts.com

Certificate of Analysis

Certificate ID: 13021
Part #: BAC105L080T
Cylinder Size: 105L
Lot Number: 14020080A2
Expiration: 7/5/2022

0.080 BAC (for the calibration of instruments used to determine breath alcohol concentration)

Contents: 105 Liters @ 1000 psig 70°F (21°C)

Component	Reported Concentration:	Analytical Accuracy (U, k=2):	Analytical Method:
Ethanol	208 ppm	4.9-0.003 det.(6/20/04)	NDIR
Nitrogen	Balance	[5.2 ppm]	

Distributed by:
CMI Inc
316 East Ninth Street
Owensboro, KY 42303
Phone 866-835-0690
www.alcoholtest.com

*Traceable to:
Certified Reference Material - 262.4 µmol/l
Ethanol in Nitrogen - Serial No. GN0015026 Lot No. 050319E11

Store in dry area, away from sources of heat, ignition and direct sunlight. Do not allow storage area to exceed 52 °C (125 °F).

Speedy Gas Lab Tech
[Signature]

Issuance Date 06-15-2020



The detector results within this certificate were obtained using equipment and standards capable of producing analytical results traceable to NIST, and apply only to the items specified in the certificate. LINTO Products Company makes no warranty or representation as to the suitability of the use of any information provided for any particular purpose. The information that is the sole discretion and risk of the user. Liability shall be limited to substituted replacement cost of this material or service.

ISO/IEC 17025:2017 Accredited Laboratory