

AS IV Serial no: 034874
Version no: 532C

TEST RECORD 00175

Temp Date Time ^{s/} 210L

Air Blank:
10/12/22 07:41 .000
Calibration Check:
23 10/12/22 07:41 .000

Subject Name

~~Test #1~~ Test #1

Subject I.D.

Operator Name, I.D.

Cpl. Johnson 220150

Location

SCCPD

AS IV Serial no: 034874
Version no: 532C

TEST RECORD 00176

Temp Date Time ^{s/} 210L

Air Blank:
10/12/22 07:43 .000
Calibration Check:
24 10/12/22 07:43 .079

Subject Name

~~Test #2~~ Test #2

Subject I.D.

Operator Name, I.D.

Cpl. Johnson 220150

Location

SCCPD

AS IV Serial no: 034874
Version no: 532C

TEST RECORD 00177

Temp Date Time ^{s/} 210L

Air Blank:
10/12/22 07:44 .000
Calibration Check:
25 10/12/22 07:44 .079

Subject Name

~~Test #3~~ Test #3

Subject I.D.

Operator Name, I.D.

Cpl. Johnson 220150

Location

SCCPD

AS IV Serial no: 034874
Version no: 532C

TEST RECORD 00178

Temp Date Time ^{s/} 210L

VOID: PFI
12 10/12/22 07:45

Subject Name

~~RST test~~

Subject I.D.

Operator Name, I.D.

Cpl. Johnson 220150

Location

SCCPD

AS IV Serial no: 034874
Version no: 532C

TEST RECORD 00179

Temp Date Time ^{s/} 210L

Air Blank:
10/12/22 07:46 .000
Subject Test: Auto
26 10/12/22 07:46 .000

Subject Name

~~Sub-test #~~

Subject I.D.

Operator Name, I.D.

Cpl. Johnson 220150

Location

SCCPD



Airgas USA LLC (LAB)
 3500 Bernard Street
 St. Louis, Mo. 63103
 Ph: (314) 533-3100
 Fax: (314) 533-7328

Certificate of Analysis

Customer Name
 Exclusive Supplier
 Intoximeters, Inc.
 2081 Craig Road
 St. Louis, Mo 63146

Test Date: 7-Apr-2021

Lot # AG109701 Model 108cacc

Exp. Date 7-Apr-2023	Cyl. Type 108	Component Ethanol Nitrogen	Certified Concentration 0.082 ± 0.002 BrAC (223 ppm) Balance
--------------------------------	-------------------------	---	---

Certification Traceable to N.I.S.T. RGM and to CRM Ethanol Standards:

<u>RGM Serial No.</u>	<u>Concentration</u>
EB0010581	392.1 ppm
EB0010570	259.8 ppm
EB0010285	208.0 ppm
EB0010561	103.6 ppm
EB0010681	52.12 ppm

<u>RGM Serial No.</u>	<u>Concentration</u>
EB0010603	393.0 ppm
EB0010559	258.2 ppm
EB0010595	208.3 ppm
EB0010562	104.2 ppm
EB0010579	52.81 ppm

<u>CRM Serial No.</u>	<u>Concentration</u>
CC727481	800.0 ppm
CC727496	253.0 ppm

<u>CRM Serial No.</u>	<u>Concentration</u>
CC727493	390.0 ppm
CC727498	150.0 ppm

Analytical Method: NDIR

Digitally signed by Quality Control
 Date: 2021.04.07 17:42:01 -05:00
 Reason: Dry gas standard certification of analysis
 Location: Airgas USA LLC (Lab)

Approved for Release:

Rod Marsala

ISO 17025:2005 A2LA accredited. Certificate Number 3082.06
 ISO 17034:2016 A2LA accredited. Certificate Number 3082.07



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM



PERMIT
TYPE II
MIKE JOHNSON

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 5/25/2022

NUMBER 220150

EXPIRES 5/25/2024

Mike Massman

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

Dave J. Nicholson

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 680-0771 (6-10)

LAB-4 (R6-10)

 STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator JOHNSON, MIKE
 Permit No 220150
 Date Issued 5/25/2022 Date Expires 5/25/2024

