



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 STATE PUBLIC HEALTH LABORATORY  
**ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT**

**RECEIVED**  
 By Tracy Crews at 11:28 am, Jul 01, 2022

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 035484	NAME OF AGENCY Saint Charles County Police Department	DATE OF INSPECTION 06/27/2022
-----------------------------	--	----------------------------------

LOCATION OF INSTRUMENT (STREET AND CITY) 301 North Second Street, Saint Charles, MO 63301	TIME OF INSPECTION 9:36 am
--	-------------------------------

**CHECKLIST:** Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)
- TEMPERATURE OF ALCO SENSOR (10°C - 40°C)
- PRINTER WORKING PROPERLY
- TIME AND DATE DISPLAYING PROPERLY

**BREATH ALCOHOL ACCURACY STANDARDS**

<input type="checkbox"/> SIMULATOR SOLUTION	<input checked="" type="checkbox"/> COMPRESSED ETHANOL-GAS MIXTURE
<input checked="" type="checkbox"/> STANDARD SUPPLIER <u>Intoximeters</u>	LOT # <u>AG109701</u> EXP. DATE <u>04/07/2023</u>
<input type="checkbox"/> SIMULATOR TEMPERATURE (34°C ± 0.2°C)	SIM. SN _____ SIM. NIST EXP DATE _____

CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)  
 Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
- 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
- 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 ➔ .078	TEST 2 ➔ .079	TEST 3 ➔ .078
---------------	---------------	---------------

RFI DETECTOR OPERATING

**INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)**

REFUSALS	0	(0-.04)	0	(.05-.09)	0	(.10-.14)	1	(.15-.19)	1	(OVER .19)	0
----------	---	---------	---	-----------	---	-----------	---	-----------	---	------------	---

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

Calibration was performed on instrument

<b>INSPECTING OFFICER</b>	
SIGNATURE Cpl.  663	PRINT NAME Mike Johnson
TYPE II PERMIT NUMBER/EXPIRATION DATE 220150 / 05-25-2024	TELEPHONE NUMBER (636) 949-3000

**Return completed report to the:** Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office by mail, fax, or email.

AS IV Serial no: 035484  
Version no: 532C

TEST RECORD 00221

Temp Date Time <sup>s/</sup> 210L

Air Blank:  
06/27/77 09:36 .000  
Calibration Check:  
25 06/27/72 09:36 .078

Subject Name

Test #1

Subject I.D.

Operator Name: I.D.

Cpl. Johnson 220150

Location

SCCOOL

AS IV Serial no: 035484  
Version no: 532C

TEST RECORD 00222

Temp Date Time <sup>s/</sup> 210L

Air Blank:  
06/27/72 09:38 .000  
Calibration Check:  
26 06/27/72 09:38 .079

Subject Name

Test #2

Subject I.D.

Operator Name: I.D.

Cpl. Johnson 220150

Location

SCCOOL

AS IV Serial no: 035484  
Version no: 532C

TEST RECORD 00223

Temp Date Time <sup>s/</sup> 210L

Air Blank:  
06/27/72 09:40 .000  
Calibration Check:  
26 06/27/72 09:40 .078

Subject Name

Test #3

Subject I.D.

Operator Name: I.D.

Cpl. Johnson 220150

Location

SCCOOL

AS IV Serial no: 035484  
Version no: 532C

TEST RECORD 00224

Temp Date Time <sup>s/</sup> 210L

WIND RTI  
12 06/27/72 09:41

Subject Name

RTI Test

Subject I.D.

Operator Name: I.D.

Cpl. Johnson 220150

Location

SCCOOL

AS IV Serial no: 035484  
Version no: 532C

TEST RECORD 00225

Temp Date Time <sup>s/</sup> 210L

Air Blank:  
06/27/77 09:42 .000  
Subject Test: Auto  
27 06/27/77 09:42 .000

Subject Name

Self-Test

Subject I.D.

Operator Name: I.D.

Cpl. Johnson 220150

Location

SCCOOL



Airgas USA LLC (LAB)  
 3500 Bernard Street  
 St. Louis, Mo. 63103  
 Ph: (314) 533-3100  
 Fax: (314) 533-7328

## Certificate of Analysis

**Customer Name**  
 Exclusive Supplier  
 Intoximeters, Inc.  
 2081 Craig Road  
 St. Louis, Mo 63146

**Test Date:** 7-Apr-2021

**Lot # AG109701 Model 108cacc**

<b>Exp. Date</b> 7-Apr-2023	<b>Cyl. Type</b> 108	<b>Component</b> Ethanol Nitrogen	<b>Certified Concentration</b> 0.082 ± 0.002 BrAC (223 ppm) Balance
--------------------------------	-------------------------	---	---

Certification Traceable to N.I.S.T. RGM and to CRM Ethanol Standards:

<table border="0"> <tr><td><b><u>RGM Serial No.</u></b></td><td><b><u>Concentration</u></b></td></tr> <tr><td>EB0010581</td><td>392.1 ppm</td></tr> <tr><td>EB0010570</td><td>259.8 ppm</td></tr> <tr><td>EB0010285</td><td>208.0 ppm</td></tr> <tr><td>EB0010561</td><td>103.6 ppm</td></tr> <tr><td>EB0010681</td><td>52.12 ppm</td></tr> </table>	<b><u>RGM Serial No.</u></b>	<b><u>Concentration</u></b>	EB0010581	392.1 ppm	EB0010570	259.8 ppm	EB0010285	208.0 ppm	EB0010561	103.6 ppm	EB0010681	52.12 ppm	<table border="0"> <tr><td><b><u>RGM Serial No.</u></b></td><td><b><u>Concentration</u></b></td></tr> <tr><td>EB0010603</td><td>393.0 ppm</td></tr> <tr><td>EB0010559</td><td>258.2 ppm</td></tr> <tr><td>EB0010595</td><td>208.3 ppm</td></tr> <tr><td>EB0010562</td><td>104.2 ppm</td></tr> <tr><td>EB0010579</td><td>52.81 ppm</td></tr> </table>	<b><u>RGM Serial No.</u></b>	<b><u>Concentration</u></b>	EB0010603	393.0 ppm	EB0010559	258.2 ppm	EB0010595	208.3 ppm	EB0010562	104.2 ppm	EB0010579	52.81 ppm
<b><u>RGM Serial No.</u></b>	<b><u>Concentration</u></b>																								
EB0010581	392.1 ppm																								
EB0010570	259.8 ppm																								
EB0010285	208.0 ppm																								
EB0010561	103.6 ppm																								
EB0010681	52.12 ppm																								
<b><u>RGM Serial No.</u></b>	<b><u>Concentration</u></b>																								
EB0010603	393.0 ppm																								
EB0010559	258.2 ppm																								
EB0010595	208.3 ppm																								
EB0010562	104.2 ppm																								
EB0010579	52.81 ppm																								
<table border="0"> <tr><td><b><u>CRM Serial No.</u></b></td><td><b><u>Concentration</u></b></td></tr> <tr><td>CC727481</td><td>800.0 ppm</td></tr> <tr><td>CC727496</td><td>253.0 ppm</td></tr> </table>	<b><u>CRM Serial No.</u></b>	<b><u>Concentration</u></b>	CC727481	800.0 ppm	CC727496	253.0 ppm	<table border="0"> <tr><td><b><u>CRM Serial No.</u></b></td><td><b><u>Concentration</u></b></td></tr> <tr><td>CC727493</td><td>390.0 ppm</td></tr> <tr><td>CC727498</td><td>150.0 ppm</td></tr> </table>	<b><u>CRM Serial No.</u></b>	<b><u>Concentration</u></b>	CC727493	390.0 ppm	CC727498	150.0 ppm												
<b><u>CRM Serial No.</u></b>	<b><u>Concentration</u></b>																								
CC727481	800.0 ppm																								
CC727496	253.0 ppm																								
<b><u>CRM Serial No.</u></b>	<b><u>Concentration</u></b>																								
CC727493	390.0 ppm																								
CC727498	150.0 ppm																								

**Analytical Method:** NDIR

Digitally signed by Quality Control  
 Date: 2021.04.07 17:42:01 -05:00  
 Reason: Dry gas standard certification of analysis  
 Location: Airgas USA LLC (Lab)

Approved for Release: \_\_\_\_\_



Rod Marsala

**ISO 17025:2005 A2LA accredited. Certificate Number 3082.06**  
**ISO 17034:2016 A2LA accredited. Certificate Number 3082.07**



STATE OF MISSOURI  
 DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 BREATH ALCOHOL PROGRAM



**PERMIT**  
**TYPE II**  
**MIKE JOHNSON**

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

**ALCO-SENSOR IV WITH PRINTER**

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 5/25/2022

NUMBER 220150

EXPIRES 5/25/2024

*Mike Mason*

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

*Paula J. Nielson*

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 590-0771 (6-10)

LAB-4 (R6-10)

 **STATE OF MISSOURI**  
 DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 BREATH ALCOHOL PROGRAM

**INSTRUMENT OPERATOR CARD**

*The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.*

**Operator** JOHNSON, MIKE  
**Permit No** 220150  
**Date Issued** 5/25/2022 **Date Expires** 5/25/2024

