



**ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT**

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 042311	NAME OF AGENCY Saint Charles County Police Department	DATE OF INSPECTION 01/02/2022
-----------------------------	--	----------------------------------

LOCATION OF INSTRUMENT (STREET AND CITY) 101 Sheriff Dierker Court, O'Fallon, Missouri 63366	TIME OF INSPECTION 5:01 pm
---	-------------------------------

**CHECKLIST:** Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)
- TEMPERATURE OF ALCO SENSOR (10°C - 40°C)
- PRINTER WORKING PROPERLY
- TIME AND DATE DISPLAYING PROPERLY

**BREATH ALCOHOL ACCURACY STANDARDS**

<input type="checkbox"/> SIMULATOR SOLUTION	<input checked="" type="checkbox"/> COMPRESSED ETHANOL-GAS MIXTURE
<input checked="" type="checkbox"/> STANDARD SUPPLIER <u>Intoximeters</u>	LOT # <u>AG109701</u> EXP. DATE <u>04/07/2023</u>
<input type="checkbox"/> SIMULATOR TEMPERATURE (34°C ± 0.2°C)	SIM. SN _____ SIM. NIST EXP DATE _____

- CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)**  
Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)
- 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
  - 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
  - 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 • .083	TEST 2 • .084	TEST 3 • .083
---------------	---------------	---------------

- RFI DETECTOR OPERATING

**INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)**

REFUSALS	0	(0-.04)	0	(.05-.09)	0	(.10-.14)	0	(.15-.19)	0	(OVER .19)	0
----------	---	---------	---	-----------	---	-----------	---	-----------	---	------------	---

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

**INSPECTING OFFICER**

SIGNATURE ▶ <u>Capl. Johnson 663</u>	PRINT NAME Mike Johnson
TYPE II PERMIT NUMBER/EXPIRATION DATE 200198 / July 8, 2022	TELEPHONE NUMBER (636) 949-3000

**Return completed report to the:** Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office by mail, fax, or email.

AS IV Serial no: 042311  
Version no: 532C

TEST RECORD 00285

Temp Date Time <sup>s/</sup> 210L

Air Blank:  
01/02/22 16:50 .000  
Calibration Check:  
26 01/02/22 16:50 .083

Subject Name

Test #1  
Subject I.D.

Operator Name, I.D.

Cpl. Johnson 200199  
Location

SCCPD

AS IV Serial no: 042311  
Version no: 532C

TEST RECORD 00286

Temp Date Time <sup>s/</sup> 210L

Air Blank:  
01/02/22 16:51 .000  
Calibration Check:  
26 01/02/22 16:51 .084

Subject Name

Test #2  
Subject I.D.

Operator Name, I.D.

Cpl. Johnson 200199  
Location

SCCPD

AS IV Serial no: 042311  
Version no: 532C

TEST RECORD 00287

Temp Date Time <sup>s/</sup> 210L

Air Blank:  
01/02/22 16:53 .000  
Calibration Check:  
27 01/02/22 16:53 .083

Subject Name

Test #3  
Subject I.D.

Operator Name, I.D.

Cpl. Johnson 200199  
Location

SCCPD

AS IV Serial no: 042311  
Version no: 532C

TEST RECORD 00288

Temp Date Time <sup>s/</sup> 210L

VOID: RFI  
12 01/02/22 16:53

Subject Name

RFI Test

Subject I.D.

Operator Name, I.D.

Cpl. Johnson 200199  
Location

SCCPD

AS IV Serial no: 042311  
Version no: 532C

TEST RECORD 00289

Temp Date Time <sup>s/</sup> 210L

Air Blank:  
01/02/22 16:55 .000  
Subject Test: Auto  
27 01/02/22 16:55 .000

Subject Name

SW Test

Subject I.D.

Operator Name, I.D.

Cpl. Johnson 200199  
Location

SCCPD



Airgas USA LLC (LAB)

3500 Bernard Street

St. Louis, Mo. 63103

Ph: (314) 533-3100

Fax: (314) 533-7328

### Certificate of Analysis

**Customer Name**

Exclusive Supplier  
Intoximeters, Inc.  
2081 Craig Road  
St. Louis, Mo 63146

**Test Date:** 7-Apr-2021

**Lot # AG109701 Model 108cacd**

<u>Exp. Date</u>	<u>Cyl. Type</u>	<u>Component</u>	<u>Certified Concentration</u>
7-Apr-2023	108	Ethanol	0.082 ± 0.002 BrAC (223 ppm)
		Nitrogen	Balance

Certification Traceable to N.I.S.T. RGM and to CRM Ethanol Standards:

<u>RGM Serial No.</u>	<u>Concentration</u>	<u>RGM Serial No.</u>	<u>Concentration</u>
EB0010581	392.1 ppm	EB0010603	393.0 ppm
EB0010570	259.8 ppm	EB0010559	258.2 ppm
EB0010285	208.0 ppm	EB0010595	208.3 ppm
EB0010561	103.6 ppm	EB0010562	104.2 ppm
EB0010681	52.12 ppm	EB0010579	52.81 ppm

  

<u>CRM Serial No.</u>	<u>Concentration</u>	<u>CRM Serial No.</u>	<u>Concentration</u>
CC727481	800.0 ppm	CC727493	390.0 ppm
CC727496	253.0 ppm	CC727498	150.0 ppm

**Analytical Method:** NDIR

Digitally signed by Quality Control  
Date: 2021.04.07 17:42:01 -05:00  
Reason: Dry gas standard certification of analysis  
Location: Airgas USA LLC (Lab)

Approved for Release:

Rod Marsala

**ISO 17025:2005 A2LA accredited. Certificate Number 3082.06**  
**ISO 17034:2016 A2LA accredited. Certificate Number 3082.07**



STATE OF MISSOURI  
 DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 BREATH ALCOHOL PROGRAM

2

**PERMIT**  
**TYPE II**  
**MIKE JOHNSON**

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

**ALCO-SENSOR IV WITH PRINTER**

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 7/8/2020

NUMBER 200198

EXPIRES 7/8/2022

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES



**STATE OF MISSOURI**  
 DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 BREATH ALCOHOL PROGRAM

**INSTRUMENT OPERATOR CARD**

*The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.*

Operator JOHNSON, MIKE  
 Permit No 200198  
 Date Issued 7/8/2020 Date Expires 7/8/2022

