



AS IV Serial no: 043580  
Version no: 532C

TEST RECORD 00453

Temp Date Time %/ 210L

Air Blank:  
10/08/22 19:50 .000  
Calibration Check:  
23 10/08/22 19:50 .076

Subject Name

Test #1

Subject I.D.

Operator Name, I.D.

Cpl. Johnson 220150

Location

5CCPD

AS IV Serial no: 043580  
Version no: 532C

TEST RECORD 00454

Temp Date Time %/ 210L

Air Blank:  
10/08/22 19:51 .000  
Calibration Check:  
24 10/08/22 19:51 .088

Subject Name

Test #2

Subject I.D.

Operator Name, I.D.

Cpl. Johnson 220150

Location

5CCPD

AS IV Serial no: 043580  
Version no: 532C

TEST RECORD 00455

Temp Date Time %/ 210L

Air Blank:  
10/08/22 19:52 .000  
Calibration Check:  
26 10/08/22 19:52 .000

Subject Name

Test #3

Subject I.D.

Operator Name, I.D.

Cpl. Johnson 220150

Location

5CCPD

AS IV Serial no: 043580  
Version no: 532C  
TEST RECORD 00455  
Temp Date Time %/ 210L  
HOLD: RTI  
12 10/08/22 19:54  
Subject Name  
Operator Name, I.D.  
Cpl. Johnson 220150  
Location  
5CCPD

AS IV Serial no: 043580  
Version no: 532C  
TEST RECORD 00457  
Temp Date Time %/ 210L  
Air Blank:  
10/08/22 19:55 .000  
Calibration Check:  
27 10/08/22 19:55 .000  
Subject Name  
Operator Name, I.D.  
Cpl. Johnson 220150  
Location  
5CCPD



Airgas USA LLC (LAB)  
 3500 Bernard Street  
 St. Louis, Mo. 63103  
 Ph: (314) 533-3100  
 Fax: (314) 533-7328

## Certificate of Analysis

**Customer Name**  
Exclusive Supplier  
 Intoximeters, Inc.  
 2081 Craig Road  
 St. Louis, Mo. 63146

**Test Date:** 7-Apr-2021

**Lot # AG109701 Model 108cacc**

<b>Exp. Date</b> 7-Apr-2023	<b>Cyl. Type</b> 108	<b>Component</b> Ethanol Nitrogen	<b>Certified Concentration</b> 0.082 ± 0.002 BrAC (223 ppm) Balance
--------------------------------	-------------------------	---	---

Certification Traceable to N.I.S.T. RGM and to CRM Ethanol Standards:

<u>RGM Serial No.</u>	<u>Concentration</u>	<u>RGM Serial No.</u>	<u>Concentration</u>
EB0010581	392.1 ppm	EB0010603	393.0 ppm
EB0010570	259.8 ppm	EB0010559	258.2 ppm
EB0010285	208.0 ppm	EB0010595	208.3 ppm
EB0010561	103.6 ppm	EB0010562	104.2 ppm
EB0010681	52.12 ppm	EB0010579	52.81 ppm

  

<u>CRM Serial No.</u>	<u>Concentration</u>	<u>CRM Serial No.</u>	<u>Concentration</u>
CC727481	800.0 ppm	CC727493	390.0 ppm
CC727496	253.0 ppm	CC727498	150.0 ppm

**Analytical Method:** NDIR

Digitally signed by Quality Control  
 Date: 2021.04.07 17:42:01 -05:00  
 Reason: Dry gas standard certification of analysis  
 Location: Airgas USA LLC (Lab)

Approved for Release: \_\_\_\_\_

*Rod Marsala*  
 \_\_\_\_\_  
 Rod Marsala

ISO 17025:2005 A2LA accredited. Certificate Number 3082.06  
 ISO 17034:2016 A2LA accredited. Certificate Number 3082.07



STATE OF MISSOURI  
 DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 BREATH ALCOHOL PROGRAM

2

**PERMIT**  
**TYPE II**  
**MIKE JOHNSON**

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

**ALCO-SENSOR IV WITH PRINTER**

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 5/25/2022

NUMBER 220150

EXPIRES 5/25/2024

*Mike Massman*

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

*Paula J. Nicholson*

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (6-10)

LAB-4 (R6-10)

 STATE OF MISSOURI  
 DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 BREATH ALCOHOL PROGRAM

**INSTRUMENT OPERATOR CARD**

*The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.*

Operator JOHNSON, MIKE  
 Permit No 220150  
 Date Issued 5/25/2022 Date Expires 5/25/2024

