



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY

ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 062092	PRINTER SN 03A.2436.036	DATE OF INSPECTION 02/04/2022
-----------------------------	----------------------------	----------------------------------

LOCATION OF INSTRUMENT (STREET AND CITY) 4001 NE Lakewood CT Lees Summit MO 64064	TIME OF INSPECTION 9:57 pm
--	-------------------------------

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)
- TEMPERATURE OF ALCO SENSOR (10°C - 40°C)
- PRINTER WORKING PROPERLY
- TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

<input type="checkbox"/> SIMULATOR SOLUTION	<input checked="" type="checkbox"/> COMPRESSED ETHANOL-GAS MIXTURE
---	--

<input checked="" type="checkbox"/> STANDARD SUPPLIER <u>Intoximeters</u> LOT # <u>AG006306</u> EXP. DATE <u>03/03/2022</u>

<input type="checkbox"/> SIMULATOR TEMPERATURE (34°C ± 0.2°C) _____ SIMULATOR SN _____ SIMULATOR EXP DATE _____

CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
- 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
- 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 ➡ .105	TEST 2 ➡ .104	TEST 3 ➡ .104
---------------	---------------	---------------

RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	0	(0-.04)	0	(.05-.09)	0	(.10-.14)	0	(.15-.19)	0	(OVER .19)	0
----------	---	---------	---	-----------	---	-----------	---	-----------	---	------------	---

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).
Instrument meets all DHSS standards and guidelines.

INSPECTING OFFICER

SIGNATURE 	PRINT NAME DEPUTY DUSTIN B. LOVE #61/0430
---------------	--

TYPE II PERMIT NUMBER/EXPIRATION DATE 220006 01-06-2024	TELEPHONE NUMBER (816) 541-8017
--	------------------------------------

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
2875 James Boulevard
Ponlar Bluff MO 63901

AS IU Serial no: 062092
Version no: 532B

TEST RECORD 01572

Temp Date Time 210L

Air Blank: 02/04/22 21:57 .000

Calibration Check: 23 02/04/22 21:57 .105

Subject Name

Monthly Maint

Subject I.D.

Operator Name, I.D.

Dep. D Lou #61

Location

UCSD GH2

AS IU Serial no: 062092
Version no: 532B

TEST RECORD 01573

Temp Date Time 210L

Air Blank: 02/04/22 22:00 .000

Calibration Check: 23 02/04/22 22:00 .104

Subject Name

Monthly Maint

Subject I.D.

Operator Name, I.D.

Dep. D Lou #61

Location

UCSD GH2

AS IU Serial no: 062092
Version no: 532B

TEST RECORD 01574

Temp Date Time 210L

Air Blank: 02/04/22 22:02 .000

Calibration Check: 24 02/04/22 22:02 .104

Subject Name

Monthly Maint

Subject I.D.

Operator Name, I.D.

D Lou #61

Location

UCSD GH2

AS IU Serial no: 062092
Version no: 532B

TEST RECORD 01575

Temp Date Time 210L

VOID: RPI

12 02/04/22 22:04

Subject Name

Monthly Maint

Subject I.D.

Operator Name, I.D.

Dep. D Lou #61

Location

UCSD GH2



Airgas USA LLC (LAB)
 3500 Bernard Street
 St. Louis, Mo. 63103
 Ph: (314) 533-3100
 Fax: (314) 533-7328

Certificate of Analysis

Customer Name
Exclusive Supplier
 Intoximeters, Inc.
 2081-Craig Road
 St. Louis, Mo 63146

Test Date: 4-Mar-2020

Lot # AG006306 Model 108cacc

Exp. Date 3-Mar-2022	Cyl. Type 108	Component Ethanol Nitrogen	Certified Concentration 0.100 ± 2% BrAC (272 ppm) Balance
--------------------------------	-------------------------	---	--

Certification Traceable to N.I.S.T. RGM and to CRM Ethanol Standards:

<table border="0"> <tr><td><u>RGM Serial No.</u></td><td><u>Concentration</u></td></tr> <tr><td>EB0010581</td><td>392.1 ppm</td></tr> <tr><td>EB0010570</td><td>259.8 ppm</td></tr> <tr><td>EB0010285</td><td>208.0 ppm</td></tr> <tr><td>EB0010561</td><td>103.6 ppm</td></tr> <tr><td>EB0010681</td><td>52.12 ppm</td></tr> </table>	<u>RGM Serial No.</u>	<u>Concentration</u>	EB0010581	392.1 ppm	EB0010570	259.8 ppm	EB0010285	208.0 ppm	EB0010561	103.6 ppm	EB0010681	52.12 ppm	<table border="0"> <tr><td><u>RGM Serial No.</u></td><td><u>Concentration</u></td></tr> <tr><td>EB0010603</td><td>393.0 ppm</td></tr> <tr><td>EB0010559</td><td>258.2 ppm</td></tr> <tr><td>EB0010595</td><td>208.3 ppm</td></tr> <tr><td>EB0010562</td><td>104.2 ppm</td></tr> <tr><td>EB0010579</td><td>52.81 ppm</td></tr> </table>	<u>RGM Serial No.</u>	<u>Concentration</u>	EB0010603	393.0 ppm	EB0010559	258.2 ppm	EB0010595	208.3 ppm	EB0010562	104.2 ppm	EB0010579	52.81 ppm
<u>RGM Serial No.</u>	<u>Concentration</u>																								
EB0010581	392.1 ppm																								
EB0010570	259.8 ppm																								
EB0010285	208.0 ppm																								
EB0010561	103.6 ppm																								
EB0010681	52.12 ppm																								
<u>RGM Serial No.</u>	<u>Concentration</u>																								
EB0010603	393.0 ppm																								
EB0010559	258.2 ppm																								
EB0010595	208.3 ppm																								
EB0010562	104.2 ppm																								
EB0010579	52.81 ppm																								
<table border="0"> <tr><td><u>CRM Serial No.</u></td><td><u>Concentration</u></td></tr> <tr><td>CC434668</td><td>800.0 ppm</td></tr> <tr><td>CC234503</td><td>253.0 ppm</td></tr> </table>	<u>CRM Serial No.</u>	<u>Concentration</u>	CC434668	800.0 ppm	CC234503	253.0 ppm	<table border="0"> <tr><td><u>CRM Serial No.</u></td><td><u>Concentration</u></td></tr> <tr><td>0056649</td><td>390.1 ppm</td></tr> <tr><td>0056662</td><td>150.2 ppm</td></tr> </table>	<u>CRM Serial No.</u>	<u>Concentration</u>	0056649	390.1 ppm	0056662	150.2 ppm												
<u>CRM Serial No.</u>	<u>Concentration</u>																								
CC434668	800.0 ppm																								
CC234503	253.0 ppm																								
<u>CRM Serial No.</u>	<u>Concentration</u>																								
0056649	390.1 ppm																								
0056662	150.2 ppm																								

Analytical Method: NDIR

Digitally signed by Quality Control
 Date: 2020.03.05 13:27:24 -06:00
 Reason: Dry gas standard certification of analysis
 Location: Airgas USA LLC (Lab)

Approved for Release: _____



Rod Marsala

ISO 17025:2005 A2LA accredited. Certificate Number 3082.06
ISO 17034:2016 A2LA accredited. Certificate Number 3082.07



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM



PERMIT
TYPE II
DUSTIN B. LOVE

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 1/6/2022

NUMBER 220006

EXPIRES 1/6/2024

Laura E. Noy

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

Donald A. Ramsey

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (6-10)

LAB-4 (R6-10)

 **STATE OF MISSOURI**
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator LOVE, DUSTIN
Permit No 220006
Date Issued 1/6/2022 Date Expires 1/6/2024

