



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES  
STATE PUBLIC HEALTH LABORATORY

**RECEIVED**

By Tracy Crews at 11:40 am, Feb 15, 2022

**ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT**

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 062093	NAME OF AGENCY Sugar Creek Police Department	DATE OF INSPECTION 02/15/2022
LOCATION OF INSTRUMENT (STREET AND CITY) 1001 Heroes Way, Sugar Creek		TIME OF INSPECTION 10:51 am

**CHECKLIST:** Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)

TEMPERATURE OF ALCO SENSOR (10°C - 40°C)

PRINTER WORKING PROPERLY

TIME AND DATE DISPLAYING PROPERLY

**BREATH ALCOHOL ACCURACY STANDARDS**

SIMULATOR SOLUTION

COMPRESSED ETHANOL-GAS MIXTURE

STANDARD SUPPLIER Guth Laboratories LOT # 1460001 EXP. DATE 03/06/2024

SIMULATOR TEMPERATURE (34°C ± 0.2°C) SIM. SN SIM. NIST EXP DATE

CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
- 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
- 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1  .101

TEST 2  .100

TEST 3  .100

RFI DETECTOR OPERATING

**INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)**

REFUSALS	2	(0-.04)	0	(.05-.09)	0	(.10-.14)	0	(.15-.19)	0	(OVER .19)	1
----------	---	---------	---	-----------	---	-----------	---	-----------	---	------------	---

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

Maintenance conducted in accordance with DHSS guidelines.

**INSPECTING OFFICER**

SIGNATURE  
*Amanda Akers*

PRINT NAME  
Amanda Akers

TYPE II PERMIT NUMBER/EXPIRATION DATE  
210216 09/16/2023

TELEPHONE NUMBER  
816-521-7923 Ext. 1235

**Return completed report to the:** Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office by mail, fax, or email.

# SUGAR CREEK MISSOURI POLICE DEPARTMENT

## ALCO-SENSOR IV WITH PRINTER TEST RESULTS

TEST # 1

AS IV Serial no: 062093  
Version no: 532B

TEST RECORD 01512

Temp Date Time 210L

Air Blank:  
02/15/22 10:51 .000  
Calibration Check:  
23 02/15/22 10:51 .101

Subject Name

Test

Subject I.D.

# 1

Operator Name, I.D.

Akers # 210216

Location

1001 Heroes Way

Sugar Creek, MO  
64054

TEST # 2

AS IV Serial no: 062093  
Version no: 532B

TEST RECORD 01513

Temp Date Time 210L

Air Blank:  
02/15/22 10:53 .000  
Calibration Check:  
24 02/15/22 10:53 .100

Subject Name

Test

Subject I.D.

# 2

Operator Name, I.D.

Akers # 210216

Location

1001 Heroes Way

Sugar Creek, MO  
64054

TEST # 3

AS IV Serial no: 062093  
Version no: 532B

TEST RECORD 01514

Temp Date Time 210L

Air Blank:  
02/15/22 10:55 .000  
Calibration Check:  
25 02/15/22 10:55 .100

Subject Name

Test

Subject I.D.

# 3

Operator Name, I.D.

Akers # 210216

Location

1001 Heroes Way

Sugar Creek, MO  
64054

RFI TEST

AS IV Serial no: 062093  
Version no: 532B

TEST RECORD 01515

Temp Date Time 210L

VOID: RFI  
12 02/15/22 10:57

Subject Name

Test

Subject I.D.

RFI

Operator Name, I.D.

Akers # 210216

Location

1001 Heroes Way

Sugar Creek, MO  
64054



# CERTIFICATE OF ANALYSIS

## EBS - ETHANOL BREATH STANDARD

GUTH LABORATORIES INC  
590 NORTH 67TH STREET  
HARRISBURG, PA 17111

INVOICE#: 20833403  
PO#: 2033679  
CUST. ITEM #: DG-U100-10  
DATE: Mar. 09, 2021

METHOD OF ANALYSIS: IR Breath Alcohol Analyzer  
ANALYTICAL ACCURACY: +/-0.002 BrAC or +/-2% whichever is greater.  
CALGAZ LOT#: 1460001  
ETHANOL IN NITROGEN

PRODUCT EXPIRATION: Mar. 06, 2024

COMPONENT	PPM	( BrAC )
ETHANOL	260.5	(0.100)
NITROGEN	BAL	
AVERAGE ANALYTICAL VALUE	PPM	( BrAC )
ETHANOL	262.1	(0.101)

REFERENCE STANDARD	CYLINDER	CONCENTRATION PPM
M.I. TRACEABLE STANDARDS*	ND50144	260.6

CERTIFICATION TRACEABLE TO National Metrology Institute Traceable Standards.

### TRACEABILITY

#### Preparation:

Gas mixtures manufactured with balances calibrated by an ISO 17025 accredited company using NIST traceable weights and meets or exceeds the requirements of NIST Handbook 44.

Calibration test 121088, 121097, 121091, or 121100 dated, 18th January 2019 applies.

#### Analytical:

Analytical Instruments Calibrated Using NMI Traceable Standards.

Certification Numbers: ND50144-20201218, A679, ND18363-20191203, A650

No affecting environmental conditions during analysis.

\*NMI is recognized by NIST through the Mutual Recognition Agreement (CIPM MRA).

CALGAZ calibration devices were found to meet all applicable requirements of the National Highway Traffic Safety Administration Model Specifications for calibrating units for breath alcohol testers.

MANUFACTURED DATE: Mar. 06, 2021

CALGAZ CYLINDER SIZE: 6D

APPROVED BY :

"We certify that all the cylinders for the Lot numbers identified herein are manufactured and tested within the requirements of CFR 49 part 178.65 and that physical and chemical test reports are on file and copies will be furnished upon request."

CALGAZ, a division of Airgas USA LLC  
821 Chesapeake Drive, Cambridge, MD 21613-0149  
Phone: (410)228-6400 Fax: (410)228-4251



STATE OF MISSOURI  
 DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 BREATH ALCOHOL PROGRAM

2

**PERMIT**  
**TYPE II**  
**AMANDA AKERS**

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

**ALCO-SENSOR IV WITH PRINTER**

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 9/16/2021

NUMBER 210216

EXPIRES 9/16/2023

*Laura Q. Day*

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

*Donald A. Ramsey*

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

 STATE OF MISSOURI  
 DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 BREATH ALCOHOL PROGRAM

**INSTRUMENT OPERATOR CARD**

*The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.*

Operator AKERS, AMANDA  
 Permit No 210216  
 Date Issued 9/16/2021 Date Expires 9/16/2023

