



AS IV Serial no: 087965  
Version no: 532B

TEST RECORD 00585

Temp	Date	Time	a/ 210L
-----			
Air Blank:			
	10/03/22	12:13	.000
Calibration Check:			
	21 10/03/22	12:13	.101

Subject Name

Test 1  
Subject I.D.

Operator Name, I.D.

Rev 220231 09/09/24

Location

2501 W. Medicine

AS IV Serial no: 087965  
Version no: 532B

TEST RECORD 00586

Temp	Date	Time	a/ 210L
-----			
Air Blank:			
	10/03/22	12:15	.000
Calibration Check:			
	21 10/03/22	12:15	.102

Subject Name

Test 2  
Subject I.D.

Operator Name, I.D.

Rev 220231 9/9/24

Location

2501 W. Medicine

AS IV Serial no: 087965  
Version no: 532B

TEST RECORD 00587

Temp	Date	Time	a/ 210L
-----			
Air Blank:			
	10/03/22	12:18	.000
Calibration Check:			
	22 10/03/22	12:18	.102

Subject Name

Test 3  
Subject I.D.

Operator Name, I.D.

Rev 220231 9/9/24

Location

2501 W. Medicine

AS IV Serial no: 087965  
Version no: 532B

TEST RECORD 00588

Temp	Date	Time	a/ 210L
-----			
VOID: RFI			
	12 10/03/22	12:20	

Subject Name

Test Rfi  
Subject I.D.

Operator Name, I.D.

Rev 220231 9/9/24

Location

2501 W. Medicine

AS IV Serial no: 087965  
Version no: 532B

TEST RECORD 00589

Temp	Date	Time	a/ 210L
-----			
Air Blank:			
	10/03/22	12:22	.000
Subject Test: Auto			
	23 10/03/22	12:22	.000

Subject Name

Test Blank  
Subject I.D.

Operator Name, I.D.

Rev 220231 9/9/24

Location

2501 W. Medicine



Airgas USA LLC (LAB)  
 3500 Bernard Street  
 St. Louis, Mo. 63103  
 Ph: (314) 533-3100  
 Fax: (314) 533-7328

## Certificate of Analysis

Customer Name  
 Exclusive Supplier  
 Intoximeters, Inc.  
 2081 Craig Road  
 St. Louis, Mo 63146

Test Date: 14-Dec-2021

Lot # AG134807 Model 34

Exp Date	Cyl. Type	Component	Certified Concentration
14-Aug-2023	34	Ethanol Nitrogen	0.100 ± 2% BrAC (272 ppm)

Certification Traceable to N.I.S.T. RGM and to CRM Ethanol Standards:

RGM Serial No.	Concentration	RGM Serial No.	Concentration
EB0010581	392.1 ppm	EB0010603	393.0 ppm
EB0010570	259.8 ppm	EB0010559	258.2 ppm
EB0010285	208.0 ppm	EB0010595	208.3 ppm
EB0010561	103.6 ppm	EB0010562	104.2 ppm
EB0010681	52.12 ppm	EB0010579	52.81 ppm

CRM Serial No.	Concentration	CRM Serial No.	Concentration
CC434668	800.0 ppm	0056649	390.1 ppm
CC234503	253.0 ppm	0056662	150.2 ppm

Analytical Method: NDIR

Digitally signed by: Quality Control  
 Reason: Dry gas standard certification of analysis  
 Location: Airgas USA LLC (Lab)  
 Date: 12.16.2021 14:30

Approved for Release:

Rod Marsala

ISO 17025:2017 A2LA accredited. Certificate Number 3082.06  
 ISO 17034:2016 A2LA accredited. Certificate Number 3082.07



STATE OF MISSOURI  
 DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 BREATH ALCOHOL PROGRAM



**PERMIT  
 TYPE II**

**JAMES N. REW**

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

**ALCO-SENSOR IV WITH PRINTER, INTOX EC/IR II**

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 9/9/2022

NUMBER 220231

EXPIRES 9/9/2024

*Laura Q. Way*

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

*Donald A. Ramsey*

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES