

RECEIVED

By Tracy Crews at 7:51 am, Jun 03, 2022



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY

ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN: 087967 PRINTER SN: 08C.3556.20 DATE OF INSPECTION: 06/01/22

LOCATION OF INSTRUMENT (STREET AND CITY): 303 E 3rd St, Joplin, MO, 64801 TIME OF INSPECTION: 2114

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)

TEMPERATURE OF ALCO SENSOR (10°C - 40°C)

PRINTER WORKING PROPERLY

TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

SIMULATOR SOLUTION

COMPRESSED ETHANOL-GAS MIXTURE

STANDARD SUPPLIER: Intoxic meters LOT # 46133404 EXP. DATE 11-30-23

SIMULATOR TEMPERATURE (34°C ± 0.2°C) SIMULATOR SN SIMULATOR EXP DATE

CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE

0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE

0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 = .104

TEST 2 = .103

TEST 3 = .103

RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS

(0-.04)

(.05-.09)

(.10-.14)

(.15-.19)

(OVER .19)

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

INSPECTING OFFICER

SIGNATURE

PRINT NAME

K. Hinkle 6082

TYPE II PERMIT NUMBER/EXPIRATION DATE

220221

11/26/23

TELEPHONE NUMBER

417-623-3131

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
2875 James Boulevard
Poplar Bluff, MO 63901

AS IV Serial no: 087967
Version no: 532B

TEST RECORD 00798

Temp Date Time 210L

Air Blank:
06/01/23 21:17 .000
Calibration Check:
22 06/01/23 21:17 .103

Subject Name

Test 3

Subject I.D.

Operator Name, I.D.

H. Hinkle 1082

Location

303 E 3rd St

Joplin Mo 64801

AS IV Serial no: 087967
Version no: 532B

TEST RECORD 00796

Temp Date Time 210L

Air Blank:
06/01/23 21:14 .000
Subject Test: Man
20 06/01/23 21:14 .104

Subject Name

Test 1

Subject I.D.

Operator Name, I.D.

H. Hinkle 1082

Location

303 E 3rd St

Joplin Mo 64801

AS IV Serial no: 087967
Version no: 532B

TEST RECORD 00799

Temp Date Time 210L

VOID: RFI
12 06/01/23 21:19

Subject Name

Test 4

Subject I.D.

Operator Name, I.D.

H. Hinkle 1082

Location

303 E 3rd St

Joplin Mo 64801

AS IV Serial no: 087967
Version no: 532B

TEST RECORD 00797

Temp Date Time 210L

Air Blank:
06/01/23 21:15 .000
Calibration Check:
21 06/01/23 21:15 .103

Subject Name

Test 2

Subject I.D.

Operator Name, I.D.

H. Hinkle 1082

Location

303 E 3rd St

Joplin Mo 64801



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

2

PERMIT
TYPE II
HAYDEN HINKLE

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER, INTOXILYZER 8000

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 11/26/2021

NUMBER 210271

EXPIRES 11/26/2023

Laura E. Day

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

Donald A. Korman

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (5-10)

LAB-4 (R6-10)



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator HINKLE, HAYDEN
Permit No 210271
Date Issued 11/26/2021 **Date Expires** 11/26/2023

