



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY
ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 087971	NAME OF AGENCY Carrollton Police Department	DATE OF INSPECTION 02/03/2022
LOCATION OF INSTRUMENT (STREET AND CITY) 106 S. Folger Street, Carrollton		TIME OF INSPECTION 2:59 pm

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)
- TEMPERATURE OF ALCO SENSOR (10°C - 40°C)
- PRINTER WORKING PROPERLY
- TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

- SIMULATOR SOLUTION COMPRESSED ETHANOL-GAS MIXTURE

STANDARD SUPPLIER Guth Labs LOT # 21080 EXP. DATE 03/08/2023

SIMULATOR TEMPERATURE (34°C ± 0.2°C) 34.01 SIM. SN MP3879 SIM. NIST EXP DATE 06/28/2022

- CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)
 Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)
 - 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
 - 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
 - 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 .101 TEST 2 .100 TEST 3 .100

- RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	0	(0-.04)	0	(.05-.09)	0	(.10-.14)	0	(.15-.19)	0	(OVER .19)	0
----------	---	---------	---	-----------	---	-----------	---	-----------	---	------------	---

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

INSPECTING OFFICER

SIGNATURE Brian C. Woods PRINT NAME Brian C. Woods

TYPE II PERMIT NUMBER/EXPIRATION DATE 200226 08/13/2022 TELEPHONE NUMBER (660) 542-2200

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office by mail, fax, or email.

AS IV Serial no: 087971
Version no: 532B

TEST RECORD 00243

Temp Date Time 210L

Air Blank:
02/03/22 4:59 .000
Calibration Check:
22 02/03/22 4:59 .101

Subject Name

Test #1

Subject I.D.

Operator Name, I.D.

BC Woods, 782

Location

AS IV Serial no: 087971
Version no: 532B

TEST RECORD 00244

Temp Date Time 210L

Air Blank:
02/03/22 15:01 .000
Calibration Check:
23 02/03/22 15:01 .100

Subject Name

Test #2

Subject I.D.

Operator Name, I.D.

BC Woods, 782

Location

AS IV Serial no: 087971
Version no: 532B

TEST RECORD 00245

Temp Date Time 210L

Air Blank:
02/03/22 15:03 .000
Calibration Check:
23 02/03/22 15:03 .100

Subject Name

Test #3

Subject I.D.

Operator Name, I.D.

BC Woods, 782

Location

AS IV Serial no: 087971
Version no: 532B

TEST RECORD 00246

Temp Date Time 210L

VOID: RFI
12 02/03/22 15:06

Subject Name

RFI Test

Subject I.D.

Operator Name, I.D.

BC Woods, 782

Location

AS IV Serial no: 087971
Version no: 532B

TEST RECORD 00247

Temp Date Time 210L

Air Blank:
02/03/22 15:08 .000
Subject Test: Auto
24 02/03/22 15:08 .000

Subject Name

Blank Test

Subject I.D.

Operator Name, I.D.

BC Woods, 782

Location



GUTH LABORATORIES, INC.

590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number **21080** of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on **March 10, 2021**, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain **0.1214%** (w/vol) ethyl alcohol. The expiration date for this lot number is **March 8, 2023** at 11:59 PM.

When used in a calibrated Simulator, operating at $34^{\circ}\text{C} \pm .2^{\circ}\text{C}$, this solution will give a breath alcohol analysis instrument reading of **0.100 g/210L \pm 3%**.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number FN03052002 whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.

