



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY
ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN Nixa 095961	NAME OF AGENCY Nixa Police Department	DATE OF INSPECTION 02/23/2022
LOCATION OF INSTRUMENT (STREET AND CITY) 715 W Center Circle, Nixa, MO 65714		TIME OF INSPECTION 12:30 am

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)
- TEMPERATURE OF ALCO SENSOR (10°C - 40°C)
- PRINTER WORKING PROPERLY
- TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

- SIMULATOR SOLUTION .100 COMPRESSED ETHANOL-GAS MIXTURE
- STANDARD SUPPLIER RepCo Marketing Inc LOT # 21001 EXP. DATE 06/16/2023
- SIMULATOR TEMPERATURE (34°C ± 0.2°C) 34.0 SIM. SN MP 5537 SIM. NIST EXP DATE 09/03/2022

- CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)
 Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)
 - 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
 - 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
 - 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 ← .100	TEST 2 ← .100	TEST 3 ← .100
---------------	---------------	---------------

- RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	0	(0-.04)	0	(.05-.09)	0	(.10-.14)	0	(.15-.19)	0	(OVER .19)	0
----------	---	---------	---	-----------	---	-----------	---	-----------	---	------------	---

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

INSPECTING OFFICER	
SIGNATURE 403	PRINT NAME Sgt. J Barton
TYPE II PERMIT NUMBER/EXPIRATION DATE 210015 02/01/2023	TELEPHONE NUMBER (417) 725-2510

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office by mail, fax, or email.

Nixa Police Department

Calibration Check slip's

AS IV Serial no: 095961
Version no: 532B

TEST RECORD 02424

Temp	Date	Time	s/
		210L	

Air Blank:
02/23/22 00:30 .000
Calibration Check:
25 02/23/22 00:30 .100

Subject Name
TEST 2

Subject I.D.

Operator Name, I.D.
JH 403

Location

AS IV Serial no: 095961
Version no: 532B

TEST RECORD 02425

Temp	Date	Time	s/
		210L	

Air Blank:
02/23/22 00:33 .000
Calibration Check:
27 02/23/22 00:33 .100

Subject Name
TEST 2

Subject I.D.

Operator Name, I.D.
JH 403

Location

AS IV Serial no: 095961
Version no: 532B

TEST RECORD 02426

Temp	Date	Time	s/
		210L	

Air Blank:
02/23/22 00:35 .000
Calibration Check:
28 02/23/22 00:35 .100

Subject Name
TEST 3

Subject I.D.

Operator Name, I.D.
JH 403

Location

Nixa Police Department

RFI Evidence slip

AG IV Serial no: 095961			
Version no: 532B			
TEST RECORD 02427			
Temp	Date	Time	210L

VOID: RFI			
12 02/23/22 00:38			

Subject Name			
RFI			

Subject I.D.			


Operator Name, I.D.			
403			

Location			

Nixa Police Department

Blank (Zero) test Evidence slip

AS IV Serial no: 095961			
Version no: 532B			
TEST RECORD 02428			
Temp	Date	Time	210L ^{9/}
Air Blank:			
	02/23/22	00:39	.000
Subject Test: Auto			
28	02/23/22	00:39	.000
Subject Name			
BLANK			
Subject I.D.			
Operator Name, I.D.			
[Signature] 403			
Location			



RepCo Marketing Co
3101-188 Stony Brook Drive
Raleigh, NC 27604
919-876-5480

CERTIFICATE OF ANALYSIS

MANUFACTURER AND SUPPLIER: RepCo Marketing Co.

LOT NUMBER: 21001

EXPIRATION DATE: June 16, 2023 at 11:59 p.m.

RepCo Marketing Co. certifies the following:

RepCo Marketing Co. prepared, tested and supplied Lot Number 21001 of Alcohol Certified Solution for simulators. This solution was manufactured and tested by RepCo Marketing Co., with confirmation by ISO 17025 and ISO 17034 accredited institution Alcohol Countermeasure Systems, using NIST standards. Random samples were analyzed by Alcohol Countermeasure Systems utilizing a gas chromatograph and found to contain .1222 gms/dl +/- .003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of .100 +/- 3% gms/210L breath when heated to 34 Degrees Celsius +/- 0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is June 17, 2021 The expiration date for this lot number is June 16, 2023 at 11:59 p.m.

This document is a true representation of the original Certificate of Analysis.

Alma Palmer, Operations Manager
RepCo Marketing Co.



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

2

**PERMIT
TYPE II**

JOSHUA C. BARTON

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 2/1/2021

NUMBER 210015

EXPIRES 2/1/2023

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (6-10)

LAB-4 (R6-10)



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator BARTON, JOSHUA
Permit No 210015
Date Issued 2/1/2021 Date Expires 2/1/2023

