



ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 096228	NAME OF AGENCY NORMANDY POLICE DEPARTMENT	DATE OF INSPECTION 12/11/2022
LOCATION OF INSTRUMENT (STREET AND CITY) 7700 NATURAL BRIDGE ROAD, NORMANDY		TIME OF INSPECTION 2:52 am

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)

TEMPERATURE OF ALCO SENSOR (10°C - 40°C)

PRINTER WORKING PROPERLY

TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

SIMULATOR SOLUTION COMPRESSED ETHANOL-GAS MIXTURE

STANDARD SUPPLIER INTOXIMETER LOT # AG127802 EXP. DATE 10/05/2023

SIMULATOR TEMPERATURE (34°C ± 0.2°C) _____ SIM. SN _____ SIM. NIST EXP DATE _____

CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)
Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE

0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE

0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 ➔ .083	TEST 2 ➔ .083	TEST 3 ➔ .083
---------------	---------------	---------------

RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	2	(0-.04)	(.05-.09)	(.10-.14)	(.15-.19)	(OVER .19)
----------	---	---------	-----------	-----------	-----------	------------

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

INSPECTING OFFICER

SIGNATURE 	PRINT NAME CPL THOMAS M. MOORE #325
TYPE II PERMIT NUMBER/EXPIRATION DATE 220193 / 08-08-2024	TELEPHONE NUMBER (314) 385-3300

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office by mail, fax, or email.

AS IV Serial no: 096228
Version no: 532B

TEST RECORD - REPRINT

TEST RECORD 01032

Temp Date Time ^{9/} 210L

VOID: RFI
12 12/11/22 02:52

Subject Name

BAC MAINT

Subject I.D.

Operator Name, I.D.

MOORE 325/220193

Location

7700 NATURAL BR RD

NORMANDY, MD 63121

AS IV Serial no: 096228
Version no: 532B

TEST RECORD 01034

Temp Date Time ^{9/} 210L

Air Blank:
12/11/22 02:56 .000
Calibration Check:
21 12/11/22 02:56 .083

Subject Name

BAC MAINT

Subject I.D.

Operator Name, I.D.

MOORE 325/220193

Location

7700 NATURAL BR RD

NORMANDY, MD 63121

AS IV Serial no: 096228
Version no: 532B

TEST RECORD 01035

Temp Date Time ^{9/} 210L

Air Blank:
12/11/22 02:58 .000
Subject Test: Man
21 12/11/22 02:58 .083

Subject Name

BAC MAINT

Subject I.D.

Operator Name, I.D.

MOORE 325/220193

Location

7700 NATURAL BR RD

NORMANDY, MD 63121

AS IV Serial no: 096228
Version no: 532B

TEST RECORD 01036

Temp Date Time ^{9/} 210L

Air Blank:
12/11/22 03:00 .000
Subject Test: Man
21 12/11/22 03:00 .083

Subject Name

BAC MAINT

Subject I.D.

Operator Name, I.D.

MOORE 325/220193

Location

7700 NATURAL BR RD

NORMANDY, MD 63121



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

2

**PERMIT
TYPE II**

THOMAS M. MOORE

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER, INTOX DMT

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 7/8/2020

NUMBER 200200

EXPIRES 7/8/2022

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (6-10)

LAB-4 (R6-10)

STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator **MOORE, THOMAS**
Permit No **200200**
Date Issued **7/8/2020** Date Expires **7/8/2022**



Airgas USA LLC (LAB)

3500 Bernard Street

St. Louis, Mo. 63103

Ph: (314) 533-3100

Fax: (314) 533-7328

Certificate of Analysis

Customer Name

Exclusive Supplier

Intoximeters, Inc.

2081 Craig Road

St. Louis, Mo 63146

Test Date: 5-Oct-2021

Lot # AG127802 Model 108cadd

Exp. Date

5-Oct-2023

Cyl. Type

108

Component

Ethanol

Nitrogen

Certified Concentration

0.080 ± 0.002 BrAC (218 ppm)

Balance

Certification Traceable to N.I.S.T. RGM and to CRM Ethanol Standards:

RGM Serial No.

EB0010581

EB0010570

EB0010285

EB0010561

EB0010681

Concentration

392.1 ppm

259.8 ppm

208.0 ppm

103.6 ppm

52.12 ppm

RGM Serial No.

EB0010603

EB0010559

EB0010595

EB0010562

EB0010579

Concentration

393.0 ppm

258.2 ppm

208.3 ppm

104.2 ppm

52.81 ppm

CRM Serial No.

CC434668

CC234503

Concentration

800.0 ppm

253.0 ppm

CRM Serial No.

0056649

0056662

Concentration

390.1 ppm

150.2 ppm

Analytical Method:

NDIR

Digitally signed by Quality Control
Date: 2021.10.05 18:55:05 -05:00
Reason: Dry gas standard certification of analysis
Location: Airgas USA LLC (Lab)

Approved for Release:

Rod Marsala

ISO 17025:2017 A2LA accredited. Certificate Number 3082.06

ISO 17034:2016 A2LA accredited. Certificate Number 3082.07