

AS 10 Serial no: 892400
Version no: 5025

TEST RECORD 66419

Temp Date Time ^{5/} 210L

Air Blndg
05/21/77 14:04 .000
Calibration Check
23 05/21/77 14:04 .100

Subject Name

MAINT, CHECK

Subject I.N.

S.K. TALBUNT 1110

Operator Name: I.N.

Location

TROOP E HQ

POPLAR BLUFF, MO

AS 10 Serial no: 892400
Version no: 5025

TEST RECORD 66420

Temp Date Time ^{5/} 210L

Air Blndg
05/21/77 14:03 .000
Calibration Check
23 05/21/77 14:03 .099

Subject Name

MAINT CHECK

Subject I.N.

S.K. TALBUNT 1110

Operator Name: I.N.

Location

TROOP E HQ

POPLAR BLUFF, MO

2

AS 10 Serial no: 892400
Version no: 5025

TEST RECORD 66421

Temp Date Time ^{5/} 210L

Air Blndg
05/21/77 14:06 .000
Calibration Check
24 05/21/77 14:06 .100

Subject Name

MAINT CHECK

Subject I.N.

S.K. TALBUNT 1110

Operator Name: I.N.

Location

TROOP E HQ

POPLAR BLUFF, MO

1

AS 10 Serial no: 892400
Version no: 5025

TEST RECORD 66422

Temp Date Time ^{5/} 210L

(DATE PD)
12 05/21/77 14:00

Subject Name

MAINT CHECK

Subject I.N.

S.K. TALBUNT 1110

Operator Name: I.N.

Location

TROOP E HQ

POPLAR BLUFF, MO



RepCo Marketing Co
3101-188 Stony Brook Dr
Raleigh, NC 27604
919-876-5480

CERTIFICATE OF ANALYSIS

MANUFACTURER AND SUPPLIER: RepCo Marketing Co.
LOT NUMBER: 20001
EXPIRATION DATE: October 7, 2022 at 11:59 p.m.

RepCo Marketing Co. certifies the following:


RepCo Marketing Co. prepared, tested and supplied Lot Number 20001 of Alcohol Certified Solution for simulators. This solution was manufactured and tested by RepCo Marketing Co., with confirmation by ISO 17025 and ISO 17034 accredited institution Alcohol Countermeasure Systems, using NIST standards. Random samples were analyzed by Alcohol Countermeasure Systems utilizing a gas chromatograph and found to contain .1227 gms/dl +/- .003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of .100 +/- 3% gms/210L breath when heated to 34 Degrees Celsius +/- 0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is October 8, 2020 The expiration date for this lot number is October 7, 2022 at 11:59 p.m.

This document is a true representation of the original Certificate of Analysis.


Alma Palmer, Operations Manager
RepCo Marketing Co.



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM



PERMIT
TYPE II
SHAYNE K. TALBURT

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER, INTOX DMT

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 12/7/2021

NUMBER 210293

EXPIRES 12/7/2023

Laura P. Way

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

Donald S. Korman

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator TALBURT, SHAYNE
Permit No 210293
Date Issued 12/7/2021 **Date Expires** 12/7/2023

