



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY
ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

RECEIVED
 By Tracy Crews at 11:33 am, Aug 16, 2022

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 097400	NAME OF AGENCY MSHP-E	DATE OF INSPECTION 08/14/2022
LOCATION OF INSTRUMENT (STREET AND CITY) Troop E Headquarters, Poplar Bluff, Missouri 63901		TIME OF INSPECTION 8:42 am

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)

TEMPERATURE OF ALCO SENSOR (10°C - 40°C)

PRINTER WORKING PROPERLY

TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

SIMULATOR SOLUTION COMPRESSED ETHANOL-GAS MIXTURE

STANDARD SUPPLIER RepCo Marketing Inc. LOT # 20001 EXP. DATE 10/07/2022

SIMULATOR TEMPERATURE (34°C ± 0.2°C) 34.01 SIM. SN MP2217 SIM. NIST EXP DATE 03/15/2023

CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
- 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
- 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 .099

TEST 2 .100

TEST 3 .099

RFI DETECTOR OPERATING

**INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:
 (DO NOT INCLUDE SELF-ADMINISTERED TESTS)**

REFUSALS	0	(0-.04)	0	(.05-.09)	0	(.10-.14)	0	(.15-.19)	0	(OVER .19)	0
----------	---	---------	---	-----------	---	-----------	---	-----------	---	------------	---

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

This instrument operated within DHSS standards. It should be noted on Test 00435 there was an RFI status message. Test 00436 tested within DHSS standards.

INSPECTING OFFICER

SIGNATURE <i>Shayne K Talburt</i>	PRINT NAME Master Sergeant Shayne K. Talburt
--------------------------------------	-------------------------------------------------

TYPE II PERMIT NUMBER/EXPIRATION DATE 210293 12/07/2023	TELEPHONE NUMBER (573) 840-9500
------------------------------------------------------------	------------------------------------

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office by mail, fax, or email.

AS TU Serial no: 092400
Version no: 502B

TEST RECORD: 00437

Temp Date Time 210L

QCID: P1
12 08/14/22 09:17

Subject Name

MAINT CHECK

Subject J.J.

Operator Name: J.J.

S.K. TALBURT #1110

Location

TROOP E HQ

POPLAR BLUFF, MO

AS TU Serial no: 092400
Version no: 502B

TEST RECORD: 00436

Temp Date Time 210L

Air Blasts

08/14/22 09:00 .000
Calibration Check:
23 08/14/22 09:00 .099

Subject Name

MAINT CHECK

Subject J.J.

Operator Name: J.J.

S.K. TALBURT #1110

Location

TROOP E HQ

POPLAR BLUFF, MO

AS TU Serial no: 092400
Version no: 502B

TEST RECORD: 00434

Temp Date Time 210L

Air Blasts

08/14/22 09:04 .000
Calibration Check:
22 08/14/22 09:04 .100

Subject Name

MAINT CHECK

Subject J.J.

Operator Name: J.J.

S.K. TALBURT #1110

Location

TROOP E HQ

POPLAR BLUFF, MO

TEST RECORD: 00433

Temp Date Time 210L

Air Blasts
08/14/22 09:59 .000
Calibration Check:
21 08/14/22 09:59 .099

Subject Name

MAINT CHECK

Subject J.J.

Operator Name: J.J.

S.K. TALBURT #1110

Location

TROOP E HQ

POPLAR BLUFF, MO



RepCo Marketing Co
3101-188 Stony Brook Dri
Raleigh, NC 27604
919-876-5480

CERTIFICATE OF ANALYSIS

MANUFACTURER AND SUPPLIER: RepCo Marketing Co.
LOT NUMBER: 20001
EXPIRATION DATE: October 7, 2022 at 11:59 p.m.

RepCo Marketing Co. certifies the following:


RepCo Marketing Co. prepared, tested and supplied Lot Number 20001 of Alcohol Certified Solution for simulators. This solution was manufactured and tested by RepCo Marketing Co., with confirmation by ISO 17025 and ISO 17034 accredited institution Alcohol Countermeasure Systems, using NIST standards. Random samples were analyzed by Alcohol Countermeasure Systems utilizing a gas chromatograph and found to contain .1227 gms/dl +/- .003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of .100 +/- 3% gms/210L breath when heated to 34 Degrees Celsius +/- 0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is October 8, 2020 The expiration date for this lot number is October 7, 2022 at 11:59 p.m.

This document is a true representation of the original Certificate of Analysis.


Alma Palmer, Operations Manager
RepCo Marketing Co.



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM



PERMIT
TYPE II
SHAYNE K. TALBURT

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER, INTOX DMT

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 12/7/2021

NUMBER 210293

EXPIRES 12/7/2023

Laura P. Way

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

Donald S. Kowmy

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

 STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator TALBURT, SHAYNE
 Permit No 210293
 Date Issued 12/7/2021 Date Expires 12/7/2023

