



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY
ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 097416	NAME OF AGENCY Greene County Sheriff's Office	DATE OF INSPECTION 12/21/2022
LOCATION OF INSTRUMENT (STREET AND CITY) 5100 W. Division Springfield, MO 65802		TIME OF INSPECTION 10:34 am

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)
- TEMPERATURE OF ALCO SENSOR (10°C - 40°C)
- PRINTER WORKING PROPERLY
- TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

<input type="checkbox"/> SIMULATOR SOLUTION	<input checked="" type="checkbox"/> COMPRESSED ETHANOL-GAS MIXTURE
<input checked="" type="checkbox"/> STANDARD SUPPLIER <u>Intoximeters</u> LOT # <u>AG215102</u> EXP. DATE <u>05/31/2024</u>	
<input type="checkbox"/> SIMULATOR TEMPERATURE (34°C ± 0.2°C) _____ SIM. SN _____ SIM. NIST EXP DATE _____	

CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)
Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
- 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
- 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 ☛ .096	TEST 2 ☛ .098	TEST 3 ☛ .100
---------------	---------------	---------------

RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	(0-.04)	(.05-.09)	(.10-.14)	(.15-.19)	(OVER .19)
----------	---------	-----------	-----------	-----------	------------

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

Replace Paper Roll

INSPECTING OFFICER	
SIGNATURE <i>Ronald L. Killingsworth</i>	PRINT NAME Ronald L. Killingsworth
TYPE II PERMIT NUMBER/EXPIRATION DATE 210003 / 01-07-2023	TELEPHONE NUMBER (417) 829-6216

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office by mail, fax, or email.



Airgas USA LLC (LAB)
 3500 Bernard Street
 St. Louis, Mo. 63103
 Ph: (314) 533-3100
 Fax: (314) 533-7328

Certificate of Analysis

Customer Name
 Exclusive Supplier
 Intoximeters, Inc.
 2081 Craig Road
 St. Louis, Mo 63146

Test Date: 1-Jun-2022

Lot # AG215102 **Model** 108

Exp Date 31-May-2024	Cyl. Type 108	Component Ethanol Nitrogen	Certified Concentration 0.100 ± 2% BrAC (272 ppm)
--------------------------------	-------------------------	---	---

Certification Traceable to N.I.S.T. RGM and to CRM Ethanol Standards:

RGM Serial No.	Concentration	RGM Serial No.	Concentration
EB0010581	391.8 ppm	EB0010603	392.5 ppm
EB0010570	259.8 ppm	EB0010559	258.9 ppm
EB0010285	209.0 ppm	EB0010562	104.2 ppm
EB0010561	103.7 ppm	EB0010579	52.94 ppm
EB0010681	52.22 ppm		

CRM Serial No.	Concentration	CRM Serial No.	Concentration
CC727481	800.0 ppm	CC727493	390.0 ppm
CC727496	253.0 ppm	CC727498	150.0 ppm

Analytical Method: NDIR

Digitally signed by:Quality Control
 Reason:Dry gas standard certification of analysis
 Location:Airgas USA LLC (Lab)
 Date:06.02.2022 17:18

Approved for Release: _____

 Rod Marsala

ISO 17025:2017 A2LA accredited. Certificate Number 3082.06
ISO 17034:2016 A2LA accredited. Certificate Number 3082.07



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

2

**PERMIT
 TYPE II**

RONALD L. KILLINGSWORTH

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER, INTOX EC/IR II

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 1/7/2021

NUMBER 210003

EXPIRES 1/7/2023

MO 580-0771 (6-10)

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

LAB-4 (R6-10)



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator KILLINGSWORTH, RONALD
Permit No 210003
Date Issued 1/7/2021 **Date Expires** 1/7/2023



AS IV Serial no: 09 416
Version no: 532B

TEST RECORD 009 3

Temp Date Time ^{g/} 210L

Air Blank:
12/21/22 10:34 300
Calibration Check:
21 12/21/22 10:34 396

Subject Name

Test 1

Subject I.D.

Operator Name, I.D.

Ronald Z. Clayton

Location

AS IV Serial no: 09 416
Version no: 532B

TEST RECORD 009 4

Temp Date Time ^{g/} 210L

Air Blank:
12/21/22 10:36 300
Calibration Check:
21 12/21/22 10:36 398

Subject Name

Test 2

Subject I.D.

Operator Name, I.D.

Ronald Z. Clayton

Location

AS IV Serial no: 09 416
Version no: 532B

TEST RECORD 009 5

Temp Date Time ^{g/} 210L

Air Blank:
12/21/22 10:38 300
Calibration Check:
22 12/21/22 10:38 100

Subject Name

Test 3

Subject I.D.

Operator Name, I.D.

Ronald Z. Clayton

Location

AS IV Serial no: 09 416
Version no: 532B

TEST RECORD 009 7

Temp Date Time ^{g/} 210L

Air Blank:
12/21/22 10:40 300
Subject Test: Auto
24 12/21/22 10:40 300

Subject Name

Test 5 / Sober

Subject I.D.

Operator Name, I.D.

Ronald Z. Clayton

Location

AS IV Serial no: 09 416
Version no: 532B

TEST RECORD 009 6

Temp Date Time ^{g/} 210L

VOID: RFI
12 12/21/22 10:39

Subject Name

Test 4 / RFI

Subject I.D.

Operator Name, I.D.

Ronald Z. Clayton

Location