



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 STATE PUBLIC HEALTH LABORATORY  
**ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT**

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 097421	NAME OF AGENCY BRANSON WEST	DATE OF INSPECTION 02/21/2022
-----------------------------	--------------------------------	----------------------------------

LOCATION OF INSTRUMENT (STREET AND CITY) 110 SILVER LADY LN., BRANSON WEST MO 65737	TIME OF INSPECTION 4:33 pm
--	-------------------------------

**CHECKLIST:** Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)
- TEMPERATURE OF ALCO SENSOR (10°C - 40°C)
- PRINTER WORKING PROPERLY
- TIME AND DATE DISPLAYING PROPERLY

**BREATH ALCOHOL ACCURACY STANDARDS**

- SIMULATOR SOLUTION  COMPRESSED ETHANOL-GAS MIXTURE
- STANDARD SUPPLIER INTOXIMETERS LOT # AG127903 EXP. DATE 10/06/2023
- SIMULATOR TEMPERATURE (34°C ± 0.2°C) \_\_\_\_\_ SIM. SN \_\_\_\_\_ SIM. NIST EXP DATE \_\_\_\_\_

- CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)  
 Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)
  - 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
  - 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
  - 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1  .096	TEST 2  .097	TEST 3  .096
--------------	--------------	--------------

- RFI DETECTOR OPERATING

**INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)**

REFUSALS	0	(0-.04)	0	(.05-.09)	0	(.10-.14)	0	(.15-.19)	0	(OVER .19)	0
----------	---	---------	---	-----------	---	-----------	---	-----------	---	------------	---

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

NEW RIBBON FOR PRINTER INSTALLED, NEW COMPRESSED GAS AND CALIBRATION DONE

<b>INSPECTING OFFICER</b>	
SIGNATURE 	PRINT NAME SGT AARON HOEFT
TYPE II PERMIT NUMBER/EXPIRATION DATE 210259 11/18/2023	TELEPHONE NUMBER (417) 272-3400

**Return completed report to the:** Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office by mail, fax, or email.

AS IV Serial no: 097421  
Version no: 532B

TEST RECORD 00816

Temp	Date	Time	s/ 210L
------	------	------	------------

Air Blank:  
02/21/22 16:33 .000  
Calibration Check:  
25 02/21/22 16:33 .096

Subject Name

Test 1

Subject I.D.

Test 1

Operator Name, I.D.

Hoelt 210259

Location

Branson West MO

110 Silver Lady LN

AS IV Serial no: 097421  
Version no: 532B

TEST RECORD 00817

Temp	Date	Time	s/ 210L
------	------	------	------------

Air Blank:  
02/21/22 16:35 .000  
Calibration Check:  
26 02/21/22 16:35 .097

Subject Name

Test 2

Subject I.D.

Test 2

Operator Name, I.D.

Hoelt 210259

Location

Branson West MO

110 Silver Lady LN

AS IV Serial no: 097421  
Version no: 532B

TEST RECORD 00818

Temp	Date	Time	s/ 210L
------	------	------	------------

Air Blank:  
02/21/22 16:37 .000  
Calibration Check:  
27 02/21/22 16:37 .096

Subject Name

Test 3

Subject I.D.

Test 3

Operator Name, I.D.

Hoelt 210259

Location

Branson West

110 Silver Lady LN

AS IV Serial no: 097421  
Version no: 532B

TEST RECORD 00819

Temp	Date	Time	s/ 210L
------	------	------	------------

VOID: RFI  
12 02/21/22 16:39

Subject Name

RFI Test

Subject I.D.

RFI test

Operator Name, I.D.

Hoelt 210259

Location

Branson West MO

110 Silver Lady LN



Airgas USA LLC (LAB)

3500 Bernard Street

St. Louis, Mo. 63103

Ph: (314) 533-3100

Fax: (314) 533-7328

## Certificate of Analysis

**Customer Name**

*Exclusive Supplier*

Intoximeters, Inc.  
2081 Craig Road  
St. Louis, Mo 63146

**Test Date:** 6-Oct-2021

**Lot # AG127903 Model 108cadd**

**Exp. Date**

6-Oct-2023

**Cyl. Type**

108

**Component**

Ethanol  
Nitrogen

**Certified Concentration**

0.100 ± 2% BrAC (272 ppm)  
Balance

**Certification Traceable to N.I.S.T. RGM and to CRM Ethanol Standards:**

**RGM Serial No.**

EB0010581

EB0010570

EB0010285

EB0010561

EB0010681

**Concentration**

392.1 ppm

259.8 ppm

208.0 ppm

103.6 ppm

52.12 ppm

**RGM Serial No.**

EB0010603

EB0010559

EB0010595

EB0010562

EB0010579

**Concentration**

393.0 ppm

258.2 ppm

208.3 ppm

104.2 ppm

52.81 ppm

**CRM Serial No.**

CC434668

CC234503

**Concentration**

800.0 ppm

253.0 ppm

**CRM Serial No.**

0056649

0056662

**Concentration**

390.1 ppm

150.2 ppm

**Analytical Method:**

NDIR

Digitally signed by Quality Control  
Date: 2021.10.07 19:22:40 -05:00  
Reason: Dry gas standard certification of analysis  
Location: Airgas USA LLC (Lab)

Approved for Release:

Rod Marsala

**ISO 17025:2017 A2LA accredited. Certificate Number 3082.06**

**ISO 17034:2016 A2LA accredited. Certificate Number 3082.07**



STATE OF MISSOURI  
DEPARTMENT OF HEALTH AND SENIOR SERVICES  
BREATH ALCOHOL PROGRAM



**PERMIT**  
**TYPE II**  
**AARON HOEFT**

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

**ALCO-SENSOR IV WITH PRINTER, INTOX DMT**

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 11/18/2021

NUMBER 210259

EXPIRES 11/18/2023

*Laura Q. Noy*

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

*Donald A. Kamm*

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 680-0771 (6-10)

LAB-4 (R6-10)



**STATE OF MISSOURI**  
DEPARTMENT OF HEALTH AND SENIOR SERVICES  
BREATH ALCOHOL PROGRAM

**INSTRUMENT OPERATOR CARD**

*The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.*

Operator HOEFT, AARON  
Permit No 210259  
Date Issued 11/18/2021    Date Expires 11/18/2023

