



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY
ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 097422	NAME OF AGENCY Newton County Sheriff's Office	DATE OF INSPECTION 07/19/2022
LOCATION OF INSTRUMENT (STREET AND CITY) 208 W. Coler Street Neosho, MO 64850		TIME OF INSPECTION 8:17 am

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

<input checked="" type="checkbox"/> DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)
<input checked="" type="checkbox"/> TEMPERATURE OF ALCO SENSOR (10°C - 40°C)
<input checked="" type="checkbox"/> PRINTER WORKING PROPERLY
<input checked="" type="checkbox"/> TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

<input checked="" type="checkbox"/> SIMULATOR SOLUTION	<input type="checkbox"/> COMPRESSED ETHANOL-GAS MIXTURE
<input checked="" type="checkbox"/> STANDARD SUPPLIER <u>Guth Laboratories</u> LOT # <u>21380</u> EXP. DATE <u>09/13/2023</u>	
<input checked="" type="checkbox"/> SIMULATOR TEMPERATURE (34°C ± 0.2°C) <u>34.0</u> SIM. SN <u>DR6930</u> SIM. NIST EXP DATE <u>10/07/2022</u>	
<input checked="" type="checkbox"/> CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT) Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)	
<input checked="" type="checkbox"/> 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE	
<input type="checkbox"/> 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE	
<input type="checkbox"/> 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE	

TEST 1 .099	TEST 2 .099	TEST 3 .099
--------------	--------------	--------------

RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	(0-.04)	(.05-.09)	(.10-.14)	(.15-.19)	(OVER .19)
----------	---------	-----------	-----------	-----------	------------

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

INSPECTING OFFICER

SIGNATURE 	PRINT NAME Joshua Fort
TYPE II PERMIT NUMBER/EXPIRATION DATE 220097 03/18/2024	TELEPHONE NUMBER (417) 451-8300

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office by mail, fax, or email.

AS IV Serial no: 897422
Version no: 532B

TEST RECORD 00524

Temp	Date	Time	s/	210L
------	------	------	----	------

Air Blank:
07/19/22 08:17 .000
Calibration Check:
20 07/19/22 08:17 .099

Subject Name

Test

Subject I.D.

#1

Operator Name, I.D.

Joshua Fort 220097

Location Exp. 3/18/24

208 W. Coler St

Neosho, mo 64850

AS IV Serial no: 897422
Version no: 532B

TEST RECORD 00525

Temp	Date	Time	s/	210L
------	------	------	----	------

Air Blank:
07/19/22 08:19 .000
Calibration Check:
21 07/19/22 08:19 .099

Subject Name

Test

Subject I.D.

#2

Operator Name, I.D.

Joshua Fort 220097

Location Exp. 3/18/24

208 W. Coler St

Neosho, mo 64850

AS IV Serial no: 897422
Version no: 532B

TEST RECORD 00526

Temp	Date	Time	s/	210L
------	------	------	----	------

Air Blank:
07/19/22 08:21 .000
Calibration Check:
21 07/19/22 08:21 .099

Subject Name

Test

Subject I.D.

#3

Operator Name, I.D.

Joshua Fort 220097

Location Exp. 3/18/24

208 W. Coler St

Neosho, mo 64850

AS IV Serial no: 897422
Version no: 532B

TEST RECORD 00527

Temp	Date	Time	s/	210L
------	------	------	----	------

Air Blank:
07/19/22 08:22 .000
Subject Test: Auto
22 07/19/22 08:22 .000

Subject Name

Test

Subject I.D.

Blank

Operator Name, I.D.

Joshua Fort 220097

Location Exp. 3/18/24

208 W. Coler St

Neosho, mo 64850

AS IV Serial no: 897422
Version no: 532B

TEST RECORD 00528

Temp	Date	Time	s/	210L
------	------	------	----	------

VOID: RFI
12 07/19/22 08:24

Subject Name

Test

Subject I.D.

RFI

Operator Name, I.D.

Joshua Fort 220097

Location Exp. 3/18/24

208 W. Coler St

Neosho, mo 64850



GUTH LABORATORIES, INC.

590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number **21380** of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on **September 15, 2021**, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain **0.1214%** (w/vol) ethyl alcohol. The expiration date for this lot number is **September 13, 2023** at 11:59 PM.

When used in a calibrated Simulator, operating at $34^{\circ}\text{C} \pm .2^{\circ}\text{C}$, this solution will give a breath alcohol analysis instrument reading of **0.100 g/210L \pm 3%**.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number FN03052002 whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM



PERMIT
TYPE II
JOSHUA FORT

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 3/18/2022

Laura P. Wang

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

NUMBER 220097

Paula J. Nielsen

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

EXPIRES 3/18/2024

MQ 880-0771 (6-10)

LAB-4 (R6-10)



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator FORT, JOSHUA
 Permit No 220097
 Date Issued 3/18/2022 Date Expires 3/18/2024

