



**ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT**

REPORT #7

**RECEIVED**

By Brianna Medrano at 1:23 pm, Jan 24, 2022

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN <b>097425</b>	NAME OF AGENCY <b>St. Louis County Police Department</b>	DATE OF INSPECTION <b>01/21/2022</b>
------------------------------------	---	---

LOCATION OF INSTRUMENT (STREET AND CITY) <b>14301 South Outer 40 Rd, Chesterfield</b>	TIME OF INSPECTION <b>11:46 am</b>
--	---------------------------------------

**CHECKLIST:** Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)
- TEMPERATURE OF ALCO SENSOR (10°C - 40°C)
- PRINTER WORKING PROPERLY
- TIME AND DATE DISPLAYING PROPERLY

**BREATH ALCOHOL ACCURACY STANDARDS**

<input type="checkbox"/> SIMULATOR SOLUTION	<input checked="" type="checkbox"/> COMPRESSED ETHANOL-GAS MIXTURE
<input checked="" type="checkbox"/> STANDARD SUPPLIER <u>Intoximeters</u>	LOT # <u>AG015503</u> EXP. DATE <u>06/03/2022</u>
<input type="checkbox"/> SIMULATOR TEMPERATURE (34°C ± 0.2°C)	SIM. SN _____ SIM. NIST EXP DATE _____

- CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)  
Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)
  - 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
  - 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
  - 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 <b>.101</b>	TEST 2 <b>.101</b>	TEST 3 <b>.101</b>
--------------------	--------------------	--------------------

- RFI DETECTOR OPERATING

**INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)**

REFUSALS	0	(0-.04)	0	(.05-.09)	0	(.10-.14)	1	(.15-.19)	0	(OVER .19)	1
----------	---	---------	---	-----------	---	-----------	---	-----------	---	------------	---

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

<b>INSPECTING OFFICER</b>	
SIGNATURE <i>PO Neuman #3072</i>	PRINT NAME <b>PO NEUMAN, DSN 3072</b>
TYPE II PERMIT NUMBER/EXPIRATION DATE <b>210074 04/06/2023</b>	TELEPHONE NUMBER <b>(636) 529-8210</b>

**Return completed report to the:** Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office by mail, fax, or email.

AS TV Serial no: 697425  
Version no: 532B

TEST RECORD 66543

Temp Date Time <sup>9/</sup> 210L

Air Blank:  
01/21/22 11:46 .000  
Calibration Check:  
20 01/21/22 11:46 .101

Subject Name

Test #1

Subject I.D.

NA

Operator Name, I.D. #3072

PO D. Neuman

Location

MODOT TMC

AS TV Serial no: 697425  
Version no: 532B

TEST RECORD 66544

Temp Date Time <sup>9/</sup> 210L

Air Blank:  
01/21/22 11:48 .000  
Calibration Check:  
21 01/21/22 11:48 .101

Subject Name

Test #2

Subject I.D.

NA

Operator Name, I.D. #3072

PO D. Neuman

Location

MODOT TMC

AS TV Serial no: 697425  
Version no: 532B

TEST RECORD 66545

Temp Date Time <sup>9/</sup> 210L

Air Blank:  
01/21/22 11:50 .000  
Calibration Check:  
22 01/21/22 11:50 .101

Subject Name

Test #3

Subject I.D.

NA

Operator Name, I.D. #3072

PO D. Neuman

Location

MODOT TMC

AS TV Serial no: 697425  
Version no: 532B

TEST RECORD 66546

Temp Date Time <sup>9/</sup> 210L

VOID: RFI  
12 01/21/22 11:52

Subject Name

RFI!

Subject I.D.

NA

Operator Name, I.D. #3072

PO D. Neuman

Location

MODOT TMC



STATE OF MISSOURI  
 DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 BREATH ALCOHOL PROGRAM

2

**PERMIT  
 TYPE II**

**DAWN M. NEUMAN**

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

**ALCO-SENSOR IV WITH PRINTER, INTOX DMT**

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 4/6/2021

NUMBER 210074

EXPIRES 4/6/2023

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES



Airgas USA LLC (LAB)

3500 Bernard Street

St. Louis, Mo. 63103

Ph: (314) 533-3100

Fax: (314) 533-7328

## Certificate of Analysis

**Customer Name**

Exclusive Supplier

Intoximeters, Inc.

2081 Craig Road

St. Louis, Mo 63146

**Test Date:** 8-Jun-2020

**Lot #** AG015503 **Model** 108cadd

**Exp. Date**  
3-Jun-2022

**Cyl. Type**  
108

**Component**  
Ethanol  
Nitrogen

**Certified Concentration**  
0.100 ± 2% BrAC (272 ppm)  
Balance

Certification Traceable to N.I.S.T. RGM and to CRM Ethanol Standards:

<u>RGM Serial No.</u>	<u>Concentration</u>	<u>RGM Serial No.</u>	<u>Concentration</u>
EB0010581	392.1 ppm	EB0010603	393.0 ppm
EB0010570	259.8 ppm	EB0010559	258.2 ppm
EB0010285	208.0 ppm	EB0010595	208.3 ppm
EB0010561	103.6 ppm	EB0010562	104.2 ppm
EB0010681	52.12 ppm	EB0010579	52.81 ppm
<u>CRM Serial No.</u>	<u>Concentration</u>	<u>CRM Serial No.</u>	<u>Concentration</u>
CC434668	800.0 ppm	0056649	390.1 ppm
CC234503	253.0 ppm	0056662	150.2 ppm

**Analytical Method:** NDIR

Digitally signed by Quality Control  
Date: 2020.06.10 14:11:39 -05:00  
Reason: Dry gas standard certification of analysis  
Location: Airgas USA LLC (Lab)

Approved for Release:

Rod Marsala

ISO 17025:2005 A2LA accredited. Certificate Number 3082.06  
ISO 17034:2016 A2LA accredited. Certificate Number 3082.07