

RECEIVED

By Tracy Crews at 7:56 am, May 16, 2022



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY
ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

Table with 3 columns: ALCO SENSOR IV SN (097439), PRINTER SN (096.3580.990), DATE OF INSPECTION (05/09/2022)

Table with 2 columns: LOCATION OF INSTRUMENT (STREET AND CITY) (891 Technology Drive, Weldon Spring), TIME OF INSPECTION (9:50 am)

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- Checked items: DIGITAL READOUT (ALL ELEMENTS OPERATIONAL), TEMPERATURE OF ALCO SENSOR (10°C - 40°C), PRINTER WORKING PROPERLY, TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

- Checked items: SIMULATOR SOLUTION, STANDARD SUPPLIER (RepCo Marketing Inc., LOT # 20001, EXP. DATE 10/07/2022), SIMULATOR TEMPERATURE (34°C ± 0.2°C), SIMULATOR SN (MP2223), SIMULATOR EXP DATE (01/27/2023)

- Checked item: CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT). Sub-items: 0.100% STANDARD, 0.080% STANDARD, 0.040% STANDARD.

Table with 3 columns: TEST 1 (0.098), TEST 2 (0.097), TEST 3 (0.097)

Checked item: RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

Table with 8 columns showing test counts for various ranges: REFUSALS (0), (0-.04) (0), (.05-.09) (0), (.10-.14) (0), (.15-.19) (0), (OVER .19) (0)

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

INSPECTING OFFICER

Table with 2 columns: SIGNATURE (handwritten), PRINT NAME (D. Duffie); TYPE II PERMIT NUMBER/EXPIRATION DATE (210306/ 12/14/2023), TELEPHONE NUMBER ((636) 300-2800)

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
2875 James Boulevard
Poplar Bluff, MO 63901

CERTIFICATE OF ANALYSIS

MANUFACTURER AND SUPPLIER: RepCo Marketing Co.

LOT NUMBER: 20001

EXPIRATION DATE: October 7, 2022 at 11:59 p.m.

RepCo Marketing Co. certifies the following:

RepCo Marketing Co. prepared, tested and supplied Lot Number 20001 of Alcohol Certified Solution for simulators. This solution was manufactured and tested by RepCo Marketing Co., with confirmation by ISO 17025 and ISO 17034 accredited institution Alcohol Countermeasure Systems, using NIST standards. Random samples were analyzed by Alcohol Countermeasure Systems utilizing a gas chromatograph and found to contain .1227 gms/dl +/- .003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of .100 +/- 3% gms/210L breath when heated to 34 Degrees Celsius +/- 0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is October 8, 2020 The expiration date for this lot number is October 7, 2022 at 11:59 p.m.

This document is a true representation of the original Certificate of Analysis.



Alma Palmer, Operations Manager
RepCo Marketing Co.

AS IV Serial no: 097439
Version no: 532B

TEST RECORD 00246

Temp Date Time ^{9/} 210L

Air Blank:
05/09/22 09:51 .000
Calibration Check:
21 05/09/22 09:51 .098

Subject Name
Maint Test #1
Subject I.D.

Operator Name, I.D.
Duffie #1325
Location
Troop C HQ

AS IV Serial no: 097439
Version no: 532B

TEST RECORD 00247

Temp Date Time ^{9/} 210L

Air Blank:
05/09/22 09:53 .000
Calibration Check:
22 05/09/22 09:53 .097

Subject Name
Maint Test #1
Subject I.D.

Operator Name, I.D.
Duffie #1325
Location
Troop C HQ

AS IV Serial no: 097439
Version no: 532B

TEST RECORD 00248

Temp Date Time ^{9/} 210L

Air Blank:
05/09/22 09:55 .000
Calibration Check:
23 05/09/22 09:55 .097

Subject Name
Maint Test #3
Subject I.D.

Operator Name, I.D.
Duffie #1325
Location
Troop C HQ

AS IV Serial no: 097439
Version no: 532B

TEST RECORD 00249 ^{9/} 210L

Temp Date Time ^{9/} 210L
VOID: RFI
12 05/09/22 09:57

Subject Name
RFI Test
Subject I.D.

Operator Name, I.D.
Duffie #1325
Location
Troop C HQ



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM



PERMIT
TYPE II
DUSTIN J. DUFFIE

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER, INTOX DMT

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 12/14/2021

NUMBER 210306

EXPIRES 12/14/2023

Laura Q. Day

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

Donald A. Ramsey

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator DUFFIE, DUSTIN
Permit No 210306
Date Issued 12/14/2021 **Date Expires** 12/14/2023

