



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY

ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 097460	NAME OF AGENCY Battlefield Police Department	DATE OF INSPECTION 03/14/2022
LOCATION OF INSTRUMENT (STREET AND CITY) 5434 Tower Dr. Battlefield, MO		TIME OF INSPECTION 4:00 pm

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)

TEMPERATURE OF ALCO SENSOR (10°C - 40°C)

PRINTER WORKING PROPERLY

TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

SIMULATOR SOLUTION COMPRESSED ETHANOL-GAS MIXTURE

STANDARD SUPPLIER Guth Laboratories, Inc. LOT # 21380 EXP. DATE 09/13/2023

SIMULATOR TEMPERATURE (34°C ± 0.2°C) 34 SIM. SN SD 2251 SIM. NIST EXP DATE 10/11/2022

CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
- 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
- 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 ← .104

TEST 2 ← .103

TEST 3 ← .103

RFI DETECTOR OPERATING

**INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:
(DO NOT INCLUDE SELF-ADMINISTERED TESTS)**

REFUSALS	0	(0-.04)	0	(.05-.09)	0	(.10-.14)	1	(.15-.19)	0	(OVER .19)	0
----------	---	---------	---	-----------	---	-----------	---	-----------	---	------------	---

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

INSPECTING OFFICER

SIGNATURE

PRINT NAME
Tyler Moss

TYPE II PERMIT NUMBER/EXPIRATION DATE
Number: 200201 Expires: 07/08/2022

TELEPHONE NUMBER
(417) 890-9876

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office by mail, fax, or email.

AS IV Serial no: 097460
Version no: 532B

TEST RECORD 01130

Temp	Date	Time	s/	210L
	03/14/22	16:13	.000	
	22	03/14/22	16:13	.104

Air Blank:
03/14/22 16:13 .000
Subject Test: Man
22 03/14/22 16:13 .104

Subject Name

Test 1

Subject I.D.

Tyler moss

Operator Name, I.D.

200201 3/14/22

Location

Battlefield PD

AS IV Serial no: 097460
Version no: 532B

TEST RECORD 01131

Temp	Date	Time	s/	210L
	03/14/22	16:14	.000	
	22	03/14/22	16:14	.103

Air Blank:
03/14/22 16:14 .000
Subject Test: Man
22 03/14/22 16:14 .103

Subject Name

Test 2

Subject I.D.

Tyler moss

Operator Name, I.D.

200201 3/14/22

Location

Battlefield PD

AS IV Serial no: 097460
Version no: 532B

TEST RECORD 01132

Temp	Date	Time	s/	210L
	03/14/22	16:16	.000	
	23	03/14/22	16:16	.103

Air Blank:
03/14/22 16:16 .000
Subject Test: Man
23 03/14/22 16:16 .103

Subject Name

Test 3

Subject I.D.

Tyler moss

Operator Name, I.D.

200201 3/14/22

Location

Battlefield PD

AS IV Serial no: 097460
Version no: 532B

TEST RECORD 01133

Temp	Date	Time	s/	210L
	03/14/22	16:17		

VOID: RFI
12 03/14/22 16:17

Subject Name

RFI TEST

Subject I.D.

Tyler moss

Operator Name, I.D.

200201 3/14/22

Location

Battlefield PD



GUTH LABORATORIES, INC.

590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number **21380** of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on **September 15, 2021**, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain **0.1214%** (w/vol) ethyl alcohol. The expiration date for this lot number is **September 13, 2023** at 11:59 PM.

When used in a calibrated Simulator, operating at $34^{\circ}\text{C} \pm .2^{\circ}\text{C}$, this solution will give a breath alcohol analysis instrument reading of **0.100 g/210L \pm 3%**.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number FN03052002 whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.

THE STATE OF MISSOURI)
COUNTY OF GREENE)

AFFIDAVIT

Before me, the undersigned authority, personally appeared **Tyler A. Moss**, who, being by me duly sworn, deposed as follows:

My name is **Tyler A. Moss**, I am of sound mind, capable of making this affidavit, and personally acquainted with the facts herein stated:


I am the custodian of records for Type II maintenance reports for the **Battlefield Police Department**. Attached hereto is/are 4 pages of records that are kept by me in the regular course of business, and it was in the regular course of business of maintaining the ASIV w/ Printer instrument as an employee of the **Battlefield Police Department** with knowledge of the act, event, condition, opinion, or diagnosis recorded or to transmit information thereof to be included in such record; and the record was made at or near the time of the act, event, condition, opinion of diagnosis. The records attached hereto are the original or exact duplicate of the originals of the 03/14/2022 Alco-Sensor IV with printer report.



Affiant

In witness whereof, I have hereunto subscribed my name and affixed my official seal this 14th day of March, 2022.

MARILYN S. HORN
Notary Public – Notary Seal
STATE OF MISSOURI
Christian County
My Commission Expires May 13, 2022
Commission #14864371



Notary Public



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM



PERMIT
TYPE II
TYLER MOSS

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 7/8/2020

NUMBER 200201

EXPIRES 7/8/2022

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator MOSS, TYLER
Permit No 200201
Date Issued 7/8/2020 **Date Expires** 7/8/2022

