

**RECEIVED**

By Tracy Crews at 10:56 am, Oct 31, 2022

MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES  
STATE PUBLIC HEALTH LABORATORY**ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT**

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 099363	NAME OF AGENCY Missouri State Highway Patrol	DATE OF INSPECTION 10/27/2022
LOCATION OF INSTRUMENT (STREET AND CITY) Troop F Headquarters, 2920 N. Shamrock Drive, Jefferson City		TIME OF INSPECTION 3:55 pm

**CHECKLIST:** Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument. DIGITAL READOUT (ALL ELEMENTS OPERATIONAL) TEMPERATURE OF ALCO SENSOR (10°C - 40°C) PRINTER WORKING PROPERLY TIME AND DATE DISPLAYING PROPERLY**BREATH ALCOHOL ACCURACY STANDARDS** SIMULATOR SOLUTION COMPRESSED ETHANOL-GAS MIXTURE STANDARD SUPPLIER RepCo LOT # 20001 EXP. DATE 10/07/2022 SIMULATOR TEMPERATURE (34°C ± 0.2°C) 34.00 SIM. SN MP2123 SIM. NIST EXP DATE 12/16/2022 CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 → .103

TEST 2 → .103

TEST 3 → .102

 RFI DETECTOR OPERATING**INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:  
(DO NOT INCLUDE SELF-ADMINISTERED TESTS)**

REFUSALS	0	(0-.04)	0	(.05-.09)	0	(.10-.14)	0	(.15-.19)	0	(OVER .19)	0
----------	---	---------	---	-----------	---	-----------	---	-----------	---	------------	---

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

Replaced 9v battery

**INSPECTING OFFICER**

SIGNATURE

PRINT NAME

Marylyn Dickens

TYPE II PERMIT NUMBER/EXPIRATION DATE

210247 11/12/2023

TELEPHONE NUMBER

(573) 751-1000

**Return completed report to the:** Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office by mail, fax, or email.

AS IV Serial no: 099363  
Version no: 532B

TEST RECORD 00850

Temp Date Time 210L  
s/

Air Blank:  
10/27/22 15:56 .000  
Calibration Check:  
25 10/27/22 15:56 .103

Subject Name

Test 1

Subject I.D.

Operator Name, I.D.

Dickens 675 <sup>210</sup>247

Location

Troop F HQ

AS IV Serial no: 099363  
Version no: 532B

TEST RECORD 00851

Temp Date Time 210L  
s/

Air Blank:  
10/27/22 15:58 .000  
Calibration Check:  
26 10/27/22 15:58 .103

Subject Name

Test 2

Subject I.D.

Operator Name, I.D.

Dickens 675 <sup>210</sup>247

Location

Troop F HQ

AS IV Serial no: 099363  
Version no: 532B

TEST RECORD 00852

Temp Date Time 210L  
s/

Air Blank:  
10/27/22 16:01 .000  
Calibration Check:  
26 10/27/22 16:01 .102

Subject Name

Test 3

Subject I.D.

Operator Name, I.D.

Dickens 675 <sup>210</sup>247

Location

Troop F HQ

AS IV Serial no: 099363  
Version no: 532B

TEST RECORD 00853

Temp Date Time 210L  
s/

VOID: RFI  
12 10/27/22 16:02

Subject Name

RFI

Subject I.D.

Operator Name, I.D.

Dickens 675 <sup>210</sup>247

Location

Troop F HQ



RepCo Marketing Co  
3101-188 Stony Brook Drive  
Raleigh, NC 27604  
919-876-5480

## CERTIFICATE OF ANALYSIS

**MANUFACTURER AND SUPPLIER: RepCo Marketing Co.**  
**LOT NUMBER: 20001**  
**EXPIRATION DATE: October 7, 2022 at 11:59 p.m.**

RepCo Marketing Co. certifies the following:


RepCo Marketing Co. prepared, tested and supplied Lot Number 20001 of Alcohol Certified Solution for simulators. This solution was manufactured and tested by RepCo Marketing Co., with confirmation by ISO 17025 and ISO 17034 accredited institution Alcohol Countermeasure Systems, using NIST standards. Random samples were analyzed by Alcohol Countermeasure Systems utilizing a gas chromatograph and found to contain .1227 gms/dl +/- .003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of .100 +/- 3% gms/210L breath when heated to 34 Degrees Celsius +/- 0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is October 8, 2020 The expiration date for this lot number is October 7, 2022 at 11:59 p.m.

This document is a true representation of the original Certificate of Analysis.

  
Alma Palmer, Operations Manager  
RepCo Marketing Co.



STATE OF MISSOURI  
 DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 BREATH ALCOHOL PROGRAM

2

**PERMIT**  
**TYPE II**  
**MARYLYN A. DICKENS**

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

**ALCO-SENSOR IV WITH PRINTER, INTOX DMT**

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 11/12/2021

NUMBER 210247

EXPIRES 11/12/2023

MO 580-0771 (6-10)

*Laura G. Day*

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

*Donald A. Kamm*

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

LAB-4 (R6-10)

 STATE OF MISSOURI  
 DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 BREATH ALCOHOL PROGRAM

**INSTRUMENT OPERATOR CARD**

*The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.*

Operator DICKENS, MARYLYN  
 Permit No 210247  
 Date Issued 11/12/2021 Date Expires 11/12/2023

