

AS IV Serial no: 099365
Version no: 532B

TEST RECORD 01159

Temp	Date	Time	9/ 210L
------	------	------	------------

Air Blank:
02/26/22 12:46 .000
Calibration Check:
20 02/26/22 12:46 .099

Subject Name

Maint check 1
Subject I.D.

Steven McQueen
Operator Name, I.D.

220058
Location

Laurie PD

AS IV Serial no: 099365
Version no: 532B

TEST RECORD 01160

Temp	Date	Time	9/ 210L
------	------	------	------------

Air Blank:
02/26/22 12:48 .000
Calibration Check:
20 02/26/22 12:48 .099

Subject Name

Maint check 2
Subject I.D.

Operator Name, I.D.

Steven McQueen 220058
Location

Laurie PD

AS IV Serial no: 099365
Version no: 532B

TEST RECORD 01161

Temp	Date	Time	9/ 210L
------	------	------	------------

Air Blank:
02/26/22 12:50 .000
Calibration Check:
21 02/26/22 12:50 .100

Subject Name

Maint check 3
Subject I.D.

Operator Name, I.D.

Steven McQueen 220058
Location

Laurie PD

AS IV Serial no: 099365
Version no: 532B

TEST RECORD 01162

Temp	Date	Time	9/ 210L
------	------	------	------------

VOID: RFI
12 02/26/22 12:51

Subject Name

RFI check
Subject I.D.

Operator Name, I.D.

Steven McQueen 220058
Location

Laurie PD



GUTH LABORATORIES, INC.

590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number 21190 of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on June 9, 2021, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain 0.1206% (w/vol) ethyl alcohol. The expiration date for this lot number is June 8, 2023 at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of 0.100 g/210L +/- 3%.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number FN03052002 whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM



PERMIT
TYPE II
STEVEN McQUEEN

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 2/19/2022

NUMBER 220058

EXPIRES 2/19/2024

Laura E. Noy

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

Steven McQueen

, acting director
 DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

 STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator McQUEEN, STEVEN
 Permit No 220058
 Date Issued 2/19/2022 Date Expires 2/19/2024

