



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY
ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

RECEIVED
 By Tracy Crews at 9:27 am, Nov 03, 2022

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN <u>100286</u>	NAME OF AGENCY <u>Cuba PD</u>	DATE OF INSPECTION <u>11-1-22</u>
LOCATION OF INSTRUMENT (STREET AND CITY) <u>602 S. Franklin St Cuba</u>		TIME OF INSPECTION <u>1:07 pm</u>

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)

TEMPERATURE OF ALCO SENSOR (10°C - 40°C)

PRINTER WORKING PROPERLY

TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

SIMULATOR SOLUTION

COMPRESSED ETHANOL-GAS MIXTURE

STANDARD SUPPLIER Guth LOT # 21380 EXP. DATE 9-13-23

SIMULATOR TEMPERATURE (34°C ± 0.2°C) 34° SIM. SN MP3872 SIM. NIST EXP DATE 5-9-23

CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE

0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE

0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 • .102 TEST 2 • .102 TEST 3 • .101

RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS 0 (0-.04) 0 (.05-.09) 0 (.10-.14) 1 (.15-.19) 0 (OVER .19) 0

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

INSPECTING OFFICER

SIGNATURE

PRINT NAME
Michael Conluzzi

TYPE II PERMIT NUMBER/EXPIRATION DATE
220115 / 4-29-24

TELEPHONE NUMBER
(573) 885-7979

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office by mail, fax, or email.

AS IV Serial no: 180286
Version no: 532B

TEST RECORD 01137

Temp Date Time 210L

Air Blank: 11/01/22 13:07 .000
Calibration Check: 21 11/01/22 13:07 .102

Subject Name
Blank

Subject I.D.
Test 1

Operator Name, I.D.
Cerluni / 220115

Location
602 S. Franklin St

Cuba

AS IV Serial no: 180286
Version no: 532B

TEST RECORD 01138

Temp Date Time 210L

Air Blank: 11/01/22 13:09 .000
Calibration Check: 21 11/01/22 13:09 .102

Subject Name
Blank

Subject I.D.
Test 2

Operator Name, I.D.
Cerluni / 220115

Location
602 S. Franklin St

Cuba

AS IV Serial no: 180286
Version no: 532B

TEST RECORD 01139

Temp Date Time 210L

Air Blank: 11/01/22 13:11 .000
Calibration Check: 22 11/01/22 13:11 .101

Subject Name
Blank

Subject I.D.
Test 3

Operator Name, I.D.
Cerluni / 220115

Location
602 S. Franklin St

Cuba

AS IV Serial no: 180286
Version no: 532B

TEST RECORD 01140

Temp Date Time 210L

VOID: RFI
12 11/01/22 13:12

Subject Name
Blank

Subject I.D.
RFI

Operator Name, I.D.
Cerluni / 220115

Location
602 S. Franklin St

Cuba



GUTH LABORATORIES, INC.

590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number **21380** of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on **September 15, 2021**, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain **0.1214%** (w/vol) ethyl alcohol. The expiration date for this lot number is **September 13, 2023** at 11:59 PM.

When used in a calibrated Simulator, operating at $34^{\circ}\text{C} \pm .2^{\circ}\text{C}$, this solution will give a breath alcohol analysis instrument reading of **0.100 g/210L \pm 3%**.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number FN03052002 whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

2

PERMIT
TYPE II

MICHAEL S. CENTUNZI

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 4/29/2022

F. Anna G. Nay
DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

NUMBER 220115

David J. Nielsen

EXPIRES 4/29/2024

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES
LAB-4 (R6-10)

MO 680-0771 (6-10)

 STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator **CENTUNZI, MICHAEL**
Permit No **220115**
Date issued **4/29/2022** Date Expires **4/29/2024**

