

AS IV Serial no: 182457
Version no: 532B

TEST RECORD 01005
Temp Date Time 210L
%
Air Blank/

09/13/22 23:03 .000
Calibration Check/

19 29/13/22 23:03 .000
Subject Name
Test
Subject I.D.
1

Operator Name, I.D.
Jordan Payne 2271
Location

AS IV Serial no: 182457
Version no: 532B

TEST RECORD 01006
Temp Date Time 210L
%
Air Blank/

09/13/22 23:05 .000
Calibration Check/

19 29/13/22 23:05 .000
Subject Name
Test
Subject I.D.
2

Operator Name, I.D.
Jordan Payne 2271
Location

AS IV Serial no: 182457
Version no: 532B

TEST RECORD 01007
Temp Date Time 210L
%
Air Blank/

09/13/22 23:07 .000
Calibration Check/

20 29/13/22 23:07 .000
Subject Name
Test
Subject I.D.
3

Operator Name, I.D.
Jordan Payne 2271
Location

AS IV Serial no: 182457
Version no: 532B

TEST RECORD 01008
Temp Date Time 210L
%
Air Blank/

VOID: RFI
12 09/13/22 23:08

Subject Name
RFI
Subject I.D.

Test
Operator Name, I.D.
Jordan Payne 2271
Location

AS IV Serial no: 182457
Version no: 532B

TEST RECORD 01009
Temp Date Time 210L
%
Air Blank/

09/13/22 23:10 .000
Calibration Check/

20 29/13/22 23:10 .000
Subject Name
Self
Subject I.D.

Test
Operator Name, I.D.
Jordan Payne 2271
Location



Airgas USA LLC (LAB)
 3500 Bernard Street
 St. Louis, Mo. 63103
 Ph: (314) 533-3100
 Fax: (314) 533-7328

Certificate of Analysis

Customer Name
 Exclusive Supplier
 Intoximeters, Inc.
 2081 Craig Road
 St. Louis, Mo 63146

Test Date: 11-Apr-2022

Lot # AG209701 **Model** 108

| | | | |
|-------------------------------|-------------------------|---|---|
| Exp Date 7-Apr-2024 | Cyl. Type 108 | Component Ethanol Nitrogen | Certified Concentration 0.100 ± 2% BrAC (272 ppm) |
|-------------------------------|-------------------------|---|---|

Certification Traceable to N.I.S.T. RGM and to CRM Ethanol Standards:

| RGM Serial No. | Concentration | RGM Serial No. | Concentration |
|----------------|---------------|----------------|---------------|
| EB0010581 | 391.8 ppm | EB0010603 | 392.5 ppm |
| EB0010570 | 259.8 ppm | EB0010559 | 258.9 ppm |
| EB0010285 | 209.0 ppm | EB0010562 | 104.2 ppm |
| EB0010561 | 103.7 ppm | EB0010579 | 52.94 ppm |
| EB0010681 | 52.22 ppm | | |

| CRM Serial No. | Concentration | CRM Serial No. | Concentration |
|----------------|---------------|----------------|---------------|
| CC727481 | 800.0 ppm | CC727493 | 390.0 ppm |
| CC727496 | 253.0 ppm | CC727498 | 150.0 ppm |

Analytical Method: NDIR

Digitally signed by: Quality Control
 Reason: Dry gas standard certification of analysis
 Location: Airgas USA LLC (Lab)
 Date: 04.12.2022 15:30

Approved for Release: _____

Rod Marsala

ISO 17025:2017 A2LA accredited. Certificate Number 3082.06
 ISO 17034:2016 A2LA accredited. Certificate Number 3082.07



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREWERY ALCOHOL PROGRAM



PERMIT
 TYPE II

JORDAN PAYNE

is hereby authorized to install and maintain, inspect, calibrate, perform field service and repair, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER, INTOX DMT

for the determination of the alcoholic content of breath from a sample of expired air furnished under the provisions of sections 557.061, 557.062, 557.063 and 557.101 through 557.104.

DATE 2/21/2011
 NUMBER 242222
 NUMBER 242222
 MISSOURI

Jordan Payne
 BREWERY ALCOHOL PROGRAM

David M. Johnson
 DIRECTOR OF BREWERY ALCOHOL PROGRAMS

LS-4 (10-04)

